

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM073123\
 Data File : BM041181.D
 Acq On : 31 Jul 2023 10:00
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

Quant Time: Jul 31 16:30:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 31 16:29:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.798	152	167039	20.000	ng	0.00
21) Naphthalene-d8	10.604	136	618355	20.000	ng	0.00
39) Acenaphthene-d10	14.463	164	343912	20.000	ng	0.00
64) Phenanthrene-d10	17.221	188	695974	20.000	ng	0.00
76) Chrysene-d12	21.427	240	649582	20.000	ng	0.00
86) Perylene-d12	23.797	264	732286	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.363	112	880433	78.777	ng	0.00
7) Phenol-d6	6.975	99	1144894	79.028	ng	0.00
23) Nitrobenzene-d5	8.975	82	1038949	78.136	ng	0.00
42) 2,4,6-Tribromophenol	15.963	330	410304	84.347	ng	0.00
45) 2-Fluorobiphenyl	13.086	172	2168750	79.662	ng	0.00
79) Terphenyl-d14	19.862	244	3054962	85.032	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.263	88	178355	37.304	ng	98
3) Pyridine	3.669	79	499413	39.727	ng	98
4) n-Nitrosodimethylamine	3.587	42	212106	36.481	ng	93
6) Aniline	7.134	93	678329	40.009	ng	99
8) 2-Chlorophenol	7.363	128	436359	39.994	ng	99
9) Benzaldehyde	6.945	77	302071m	39.516	ng	
10) Phenol	7.004	94	555757	39.719	ng	99
11) bis(2-Chloroethyl)ether	7.234	93	447285	39.491	ng	98
12) 1,3-Dichlorobenzene	7.681	146	471286	39.682	ng	99
13) 1,4-Dichlorobenzene	7.834	146	479093	40.134	ng	99
14) 1,2-Dichlorobenzene	8.151	146	453600	39.630	ng	99
15) Benzyl Alcohol	8.051	79	370615	40.510	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.339	45	560171m	37.120	ng	
17) 2-Methylphenol	8.257	107	383751	40.144	ng	100
18) Hexachloroethane	8.869	117	175422	39.472	ng	94
19) n-Nitroso-di-n-propyla...	8.622	70	350539	39.837	ng	97
20) 3+4-Methylphenols	8.587	107	520958	40.759	ng	99
22) Acetophenone	8.634	105	648260	39.584	ng	# 99
24) Nitrobenzene	9.016	77	524713	39.022	ng	99
25) Isophorone	9.545	82	921015	40.805	ng	99
26) 2-Nitrophenol	9.728	139	229976	43.750	ng	96
27) 2,4-Dimethylphenol	9.792	122	383119	41.419	ng	99
28) bis(2-Chloroethoxy)met...	10.028	93	564723	39.942	ng	99
29) 2,4-Dichlorophenol	10.257	162	395308	42.379	ng	100
30) 1,2,4-Trichlorobenzene	10.469	180	426367	41.371	ng	99
31) Naphthalene	10.657	128	1328242	39.583	ng	99
32) Benzoic acid	9.945	122	284079	41.951	ng	99
33) 4-Chloroaniline	10.781	127	561314	41.858	ng	99
34) Hexachlorobutadiene	10.928	225	261907	42.254	ng	99
35) Caprolactam	11.598	113	114745	42.523	ng	89
36) 4-Chloro-3-methylphenol	11.916	107	409753	41.735	ng	99
37) 2-Methylnaphthalene	12.275	142	922478	40.777	ng	99
38) 1-Methylnaphthalene	12.498	142	858494	40.546	ng	100
40) 1,2,4,5-Tetrachloroben...	12.645	216	459123	41.114	ng	99
41) Hexachlorocyclopentadiene	12.610	237	217388	36.796	ng	100
43) 2,4,6-Trichlorophenol	12.892	196	307695	42.989	ng	99

08/01/2023
 Supervised By :mohammad
 Ahmed

08/01/2023

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM073123\
 Data File : BM041181.D
 Acq On : 31 Jul 2023 10:00
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

Quant Time: Jul 31 16:30:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 31 16:29:28 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	12.969	196	359779	43.429	ng	97	08/01/2023
46) 1,1'-Biphenyl	13.292	154	1109110	39.781	ng	99	Supervised By :mohammad ahmed
47) 2-Chloronaphthalene	13.339	162	844347	40.559	ng	98	
48) 2-Nitroaniline	13.557	65	276691	41.639	ng	99	
49) Acenaphthylene	14.186	152	1354909	40.323	ng	100	
50) Dimethylphthalate	13.933	163	1053416	40.720	ng	100	
51) 2,6-Dinitrotoluene	14.063	165	227186	42.384	ng	97	08/01/2023
52) Acenaphthene	14.533	154	801876	39.614	ng	100	
53) 3-Nitroaniline	14.392	138	245163	39.980	ng	94	
54) 2,4-Dinitrophenol	14.604	184	101518	31.835	ng	91	
55) Dibenzofuran	14.868	168	1312109	39.729	ng	99	
56) 4-Nitrophenol	14.704	139	186801	41.815	ng	96	
57) 2,4-Dinitrotoluene	14.851	165	300646	39.761	ng	94	
58) Fluorene	15.515	166	1035565	39.747	ng	100	
59) 2,3,4,6-Tetrachlorophenol	15.098	232	280043	41.745	ng	98	
60) Diethylphthalate	15.298	149	1032851	41.286	ng	99	
61) 4-Chlorophenyl-phenyle...	15.515	204	540135	41.371	ng	98	
62) 4-Nitroaniline	15.563	138	222636	39.726	ng	97	
63) Azobenzene	15.810	77	1075508	39.247	ng	97	
65) 4,6-Dinitro-2-methylph...	15.615	198	154833	36.724	ng	94	
66) n-Nitrosodiphenylamine	15.739	169	866075	40.012	ng	99	
67) 4-Bromophenyl-phenylether	16.410	248	319395	41.856	ng	96	
68) Hexachlorobenzene	16.521	284	350412	41.297	ng	97	
69) Atrazine	16.698	200	242059	42.583	ng	100	
70) Pentachlorophenol	16.874	266	256031	43.575	ng	100	
71) Phenanthrene	17.268	178	1528935	39.539	ng	100	
72) Anthracene	17.357	178	1560655	40.808	ng	100	
73) Carbazole	17.639	167	1404635	40.579	ng	99	
74) Di-n-butylphthalate	18.198	149	1723334	44.058	ng	99	
75) Fluoranthene	19.292	202	1804571	41.538	ng	99	
77) Benzidine	19.492	184	456719	51.585	ng	100	
78) Pyrene	19.656	202	1892974	41.671	ng	100	
80) Butylbenzylphthalate	20.562	149	768776	46.343	ng	97	
81) Benzo(a)anthracene	21.409	228	1846344	41.928	ng	99	
82) 3,3'-Dichlorobenzidine	21.350	252	620787	43.202	ng	99	
83) Chrysene	21.462	228	1736531	40.769	ng	99	
84) Bis(2-ethylhexyl)phtha...	21.339	149	1127487	43.074	ng	# 97	
85) Di-n-octyl phthalate	22.256	149	1894228	45.181	ng	97	
87) Indeno(1,2,3-cd)pyrene	26.244	276	2154932	40.026	ng	98	
88) Benzo(b)fluoranthene	23.074	252	1764153	40.553	ng	99	
89) Benzo(k)fluoranthene	23.121	252	1823848	40.986	ng	99	
90) Benzo(a)pyrene	23.691	252	1723258	41.227	ng	99	
91) Dibenzo(a,h)anthracene	26.256	278	1812687	40.366	ng	98	
92) Benzo(g,h,i)perylene	26.997	276	1727569	39.317	ng	98	

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.798 min Scan# 71
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 167035
Ion Ratio Lower Upper
152 100
150 154.2 125.5 188.3
115 60.3 48.0 72.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#2
1,4-Dioxane
Concen: 37.304 ng
RT: 3.263 min Scan# 13
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion: 88 Resp: 178355
Ion Ratio Lower Upper
88 100
58 70.9 58.2 87.4
43 29.5 24.4 36.6





#3
Pyridine
Concen: 39.727 ng
RT: 3.669 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 49941
Ion Ratio Lower Upper
79 100
52 63.2 52.1 78.1
51 33.2 27.5 41.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#4
n-Nitrosodimethylamine
Concen: 36.481 ng
RT: 3.587 min Scan# 68
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Tgt Ion: 42 Resp: 212106
Ion Ratio Lower Upper
42 100
74 133.8 100.8 151.2
44 5.8 5.0 7.4







#5
2-Fluorophenol
Concen: 78.777 ng
RT: 5.363 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 88043
Ion Ratio Lower Upper
112 100
64 59.5 48.5 72.7
63 32.1 26.5 39.7



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#6
Aniline
Concen: 40.009 ng
RT: 7.134 min Scan# 671
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion: 93 Resp: 678329
Ion Ratio Lower Upper
93 100
66 38.7 32.0 48.0
65 20.7 16.6 24.8





#7
Phenol-d6
Concen: 79.028 ng
RT: 6.975 min Scan# 645
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 99 Resp: 1144894
Ion Ratio Lower Upper
99 100
42 18.5 15.4 23.0
71 35.6 28.2 42.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#8
2-Chlorophenol
Concen: 39.994 ng
RT: 7.363 min Scan# 710
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Tgt Ion:128 Resp: 436359
Ion Ratio Lower Upper
128 100
130 32.4 12.3 52.3
64 47.7 29.3 69.3







#9
Benzaldehyde
Concen: 39.516 ng m
RT: 6.945 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 30207
Ion Ratio Lower Upper
77 100
105 94.2 72.8 112.8
106 91.7 70.8 110.8



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#10
Phenol
Concen: 39.719 ng
RT: 7.004 min Scan# 649
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion: 94 Resp: 555757
Ion Ratio Lower Upper
94 100
65 29.0 9.5 49.5
66 37.7 18.0 58.0





#11
bis(2-Chloroethyl)ether
Concen: 39.491 ng
RT: 7.234 min Scan# 611
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :

BNA_M



ClientSampleId :

SSTDCCC040

Tgt Ion: 93 Resp: 447285
Ion Ratio Lower Upper
93 100
63 77.2 58.8 98.8
95 32.4 11.8 51.8



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#12
1,3-Dichlorobenzene
Concen: 39.682 ng
RT: 7.681 min Scan# 764
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:146 Resp: 471286
Ion Ratio Lower Upper
146 100
148 64.4 51.9 77.9
75 35.1 28.3 42.5







#13
1,4-Dichlorobenzene
Concen: 40.134 ng
RT: 7.834 min Scan# 791
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 47909
Ion Ratio Lower Upper
146 100
148 64.0 51.4 77.0
111 39.6 32.4 48.6



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#14
1,2-Dichlorobenzene
Concen: 39.630 ng
RT: 8.151 min Scan# 844
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:146 Resp: 453600
Ion Ratio Lower Upper
146 100
148 64.3 50.9 76.3
111 42.2 33.8 50.6







#15
Benzyl Alcohol
Concen: 40.510 ng
RT: 8.051 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 37061
Ion Ratio Lower Upper
79 100
108 75.1 59.8 89.8
77 66.5 53.8 80.6



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.120 ng m
RT: 8.339 min Scan# 876
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion: 45 Resp: 560171
Ion Ratio Lower Upper
45 100
77 16.1 0.0 34.5
79 12.8 0.0 32.1






#17
2-Methylphenol
Concen: 40.144 ng
RT: 8.257 min Scan# 862
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :



SSTDCCC040



Tgt Ion:107 Resp: 38375
Ion Ratio Lower Upper
107 100
108 112.2 89.9 134.9
77 50.3 39.8 59.8
79 46.9 37.7 56.5


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#18
Hexachloroethane
Concen: 39.472 ng
RT: 8.869 min Scan# 966
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:117 Resp: 175422
Ion Ratio Lower Upper
117 100
119 97.9 74.2 111.4
201 100.1 74.5 111.7





#19
n-Nitroso-di-n-propylamine
Concen: 39.837 ng
RT: 8.622 min Scan# 918
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 70 Resp: 350539
Ion Ratio Lower Upper
70 100
42 54.8 45.8 68.8
101 9.4 7.4 11.0
130 20.0 15.2 22.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#20
3+4-Methylphenols
Concen: 40.759 ng
RT: 8.587 min Scan# 918
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion:107 Resp: 520958
Ion Ratio Lower Upper
107 100
108 88.3 67.2 107.2
77 36.2 15.4 55.4
79 30.1 9.2 49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.604 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:136 Resp: 61835
Ion Ratio Lower Upper
136 100
137 10.7 8.8 13.2
54 8.1 6.9 10.3
68 6.2 5.4 8.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#22
Acetophenone
Concen: 39.584 ng
RT: 8.634 min Scan# 926
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion:105 Resp: 648260
Ion Ratio Lower Upper
105 100
71 2.9 1.2 1.8#
51 27.1 22.0 33.0
120 22.2 18.0 27.0





#23
Nitrobenzene-d5
Concen: 78.136 ng
RT: 8.975 min Scan# 901
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 1038949
Ion Ratio Lower Upper
82 100
128 41.3 31.8 47.6
54 48.6 40.1 60.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#24
Nitrobenzene
Concen: 39.022 ng
RT: 9.016 min Scan# 991
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion: 77 Resp: 524713
Ion Ratio Lower Upper
77 100
123 42.1 34.2 51.2
65 13.1 10.7 16.1





#25
Isophorone
Concen: 40.805 ng
RT: 9.545 min Scan# 1082
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 921015
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 16.5 12.8 19.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#26
2-Nitrophenol
Concen: 43.750 ng
RT: 9.728 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:139 Resp: 229976
Ion Ratio Lower Upper
139 100
109 31.8 25.4 38.2
65 54.4 47.2 70.8





#27
2,4-Dimethylphenol
Concen: 41.419 ng
RT: 9.792 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:122 Resp: 383119
Ion Ratio Lower Upper
122 100
107 117.8 94.3 141.5
121 58.9 45.8 68.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#28
bis(2-Chloroethoxy)methane
Concen: 39.942 ng
RT: 10.028 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion: 93 Resp: 564723
Ion Ratio Lower Upper
93 100
95 32.2 25.9 38.9
123 10.6 9.1 13.7





#29
2,4-Dichlorophenol
Concen: 42.379 ng
RT: 10.257 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:162 Resp: 395308
Ion Ratio Lower Upper
162 100
164 64.2 44.3 84.3
98 35.7 15.1 55.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#30
1,2,4-Trichlorobenzene
Concen: 41.371 ng
RT: 10.469 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion:180 Resp: 426367
Ion Ratio Lower Upper
180 100
182 96.3 75.9 113.9
145 28.7 23.6 35.4





#31
Naphthalene
Concen: 39.583 ng
RT: 10.657 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 132824
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.4
127 13.1 10.5 15.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#32
Benzoic acid
Concen: 41.951 ng
RT: 9.945 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:122 Resp: 284079
Ion Ratio Lower Upper
122 100
105 125.6 104.0 144.0
77 99.0 79.9 119.9





#33
4-Chloroaniline
Concen: 41.858 ng
RT: 10.781 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:127 Resp: 561314
Ion Ratio Lower Upper
127 100
129 31.7 25.9 38.9
65 33.0 26.9 40.3
92 19.7 15.8 23.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#34
Hexachlorobutadiene
Concen: 42.254 ng
RT: 10.928 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:225 Resp: 261907
Ion Ratio Lower Upper
225 100
223 61.9 49.0 73.4
227 63.9 50.8 76.2





#35
Caprolactam
Concen: 42.523 ng
RT: 11.598 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 114749
Ion Ratio Lower Upper
113 100
55 163.9 162.5 202.5
56 124.5 114.1 154.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#36
4-Chloro-3-methylphenol
Concen: 41.735 ng
RT: 11.916 min Scan# 1484
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:107 Resp: 409753
Ion Ratio Lower Upper
107 100
144 25.5 20.1 30.1
142 78.4 62.2 93.4





#37
2-Methylnaphthalene
Concen: 40.777 ng
RT: 12.275 min Scan# 1545
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 922478
Ion Ratio Lower Upper
142 100
141 89.1 70.7 106.1
115 36.4 29.3 43.9


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#38
1-Methylnaphthalene
Concen: 40.546 ng
RT: 12.498 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:142 Resp: 858494
Ion Ratio Lower Upper
142 100
141 91.8 73.4 110.2
116 4.0 3.2 4.8





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.463 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:164 Resp: 343912
Ion Ratio Lower Upper
164 100
162 100.8 80.4 120.6
160 45.1 36.2 54.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.114 ng
RT: 12.645 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:216 Resp: 459123
Ion Ratio Lower Upper
216 100
214 78.6 63.2 94.8
179 20.1 16.2 24.2
108 19.5 17.0 25.4





#41
Hexachlorocyclopentadiene
Concen: 36.796 ng
RT: 12.610 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:237 Resp: 21738
Ion Ratio Lower Upper
237 100
235 63.4 43.6 83.6
272 12.8 0.0 33.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#42
2,4,6-Tribromophenol
Concen: 84.347 ng
RT: 15.963 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:330 Resp: 410304
Ion Ratio Lower Upper
330 100
332 98.2 77.7 116.5
141 34.0 28.3 42.5





#43
2,4,6-Trichlorophenol
Concen: 42.989 ng
RT: 12.892 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:196 Resp: 30769
Ion Ratio Lower Upper
196 100
198 95.8 76.2 114.4
200 31.4 24.6 36.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#44
2,4,5-Trichlorophenol
Concen: 43.429 ng
RT: 12.969 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:196 Resp: 359779
Ion Ratio Lower Upper
196 100
198 96.8 78.2 117.4
97 48.9 42.7 64.1
132 24.9 21.6 32.4





#45
2-Fluorobiphenyl
Concen: 79.662 ng
RT: 13.086 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:172 Resp: 2168750
Ion Ratio Lower Upper
172 100
171 35.3 28.1 42.1
170 23.6 18.9 28.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#46
1,1'-Biphenyl
Concen: 39.781 ng
RT: 13.292 min Scan# 1718
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:154 Resp: 1109110
Ion Ratio Lower Upper
154 100
153 40.0 20.5 60.5
76 14.0 0.0 34.1





#47
2-Chloronaphthalene
Concen: 40.559 ng
RT: 13.339 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:162 Resp: 84434
Ion Ratio Lower Upper
162 100
127 37.2 30.6 45.8
164 32.8 25.6 38.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#48
2-Nitroaniline
Concen: 41.639 ng
RT: 13.557 min Scan# 1763
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion: 65 Resp: 276691
Ion Ratio Lower Upper
65 100
92 66.0 53.0 79.4
138 99.8 79.1 118.7





#49
Acenaphthylene
Concen: 40.323 ng
RT: 14.186 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 1354909
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1
153 12.9 10.5 15.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#50
Dimethylphthalate
Concen: 40.720 ng
RT: 13.933 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:163 Resp: 1053416
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.3 8.3 12.5





#51
2,6-Dinitrotoluene
Concen: 42.384 ng
RT: 14.063 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 227180
Ion Ratio Lower Upper
165 100
63 61.5 50.9 76.3
89 60.6 50.6 76.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#52
Acenaphthene
Concen: 39.614 ng
RT: 14.533 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:154 Resp: 801876
Ion Ratio Lower Upper
154 100
153 113.2 90.2 135.4
152 54.5 43.4 65.0





#53
3-Nitroaniline
Concen: 39.980 ng
RT: 14.392 min Scan# 1905
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 24516
Ion Ratio Lower Upper
138 100
108 9.4 7.6 11.4
92 116.7 98.9 148.3


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#54
2,4-Dinitrophenol
Concen: 31.835 ng
RT: 14.604 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:184 Resp: 101518
Ion Ratio Lower Upper
184 100
63 71.4 66.5 99.7
154 62.8 52.8 79.2





#55
Dibenzenofuran
Concen: 39.729 ng
RT: 14.868 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 1312109
Ion Ratio Lower Upper
168 100
139 36.5 29.8 44.6
169 12.9 10.5 15.7


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#56
4-Nitrophenol
Concen: 41.815 ng
RT: 14.704 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:139 Resp: 186801
Ion Ratio Lower Upper
139 100
109 78.9 61.2 101.2
65 105.1 89.5 129.5





#57
2,4-Dinitrotoluene
Concen: 39.761 ng
RT: 14.851 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 300640
Ion Ratio Lower Upper
165 100
63 56.1 48.3 72.5
89 84.1 71.8 107.8
182 2.7 2.0 3.0


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#58
Fluorene
Concen: 39.747 ng
RT: 15.515 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:166 Resp: 1035565
Ion Ratio Lower Upper
166 100
165 97.7 78.2 117.4
167 13.4 10.6 16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 41.745 ng
RT: 15.098 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:232 Resp: 280041
Ion Ratio Lower Upper
232 100
131 44.9 36.9 55.3
130 2.4 2.0 3.0
166 29.3 24.2 36.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#60
Diethylphthalate
Concen: 41.286 ng
RT: 15.298 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:149 Resp: 1032851
Ion Ratio Lower Upper
149 100
177 22.1 17.5 26.3
150 12.3 10.1 15.1






#61
4-Chlorophenyl-phenylether
Concen: 41.371 ng
RT: 15.515 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :



SSTDCCC040



Tgt Ion:204 Resp: 54013
Ion Ratio Lower Upper
204 100
206 33.0 26.5 39.7
141 53.8 44.8 67.2


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#62
4-Nitroaniline
Concen: 39.726 ng
RT: 15.563 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion:138 Resp: 222636
Ion Ratio Lower Upper
138 100
92 51.6 35.1 75.1
108 86.7 67.8 107.8





#63
Azobenzene
Concen: 39.247 ng
RT: 15.810 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 77 Resp: 1075508
Ion Ratio Lower Upper
77 100
182 23.3 1.3 41.3
105 19.1 0.0 38.5
51 27.6 9.1 49.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.221 min Scan# 2386
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:188 Resp: 695974
Ion Ratio Lower Upper
188 100
94 9.3 8.2 12.2
80 10.2 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 36.724 ng
RT: 15.615 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:198 Resp: 15483
Ion Ratio Lower Upper
198 100
51 45.9 31.7 71.7
105 45.6 27.8 67.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#66
n-Nitrosodiphenylamine
Concen: 40.012 ng
RT: 15.739 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:169 Resp: 866075
Ion Ratio Lower Upper
169 100
168 67.0 54.6 81.8
167 36.9 29.9 44.9






#67
4-Bromophenyl-phenylether
Concen: 41.856 ng
RT: 16.410 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :



SSTDCCC040



Tgt Ion:248 Resp: 31939
Ion Ratio Lower Upper
248 100
250 99.0 78.0 117.0
141 72.8 53.8 80.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#68
Hexachlorobenzene
Concen: 41.297 ng
RT: 16.521 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Tgt Ion:284 Resp: 350412
Ion Ratio Lower Upper
284 100
142 32.2 28.0 42.0
249 30.1 23.7 35.5







#69
Atrazine
Concen: 42.583 ng
RT: 16.698 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 242059
Ion Ratio Lower Upper
200 100
173 25.8 5.6 45.6
215 48.5 28.5 68.5



Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#70
Pentachlorophenol
Concen: 43.575 ng
RT: 16.874 min Scan# 2327
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00


Tgt Ion:266 Resp: 256031
Ion Ratio Lower Upper
266 100
268 64.0 51.0 76.6
264 61.6 49.4 74.2





#71
Phenanthrene
Concen: 39.539 ng
RT: 17.268 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 1528935
Ion Ratio Lower Upper
178 100
176 19.5 15.6 23.4
179 15.3 12.1 18.1


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#72
Anthracene
Concen: 40.808 ng
RT: 17.357 min Scan# 2409
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00




Tgt Ion:178 Resp: 1560655
Ion Ratio Lower Upper
178 100
176 19.2 15.2 22.8
179 15.7 12.4 18.6





#73
 Carbazole
 Concen: 40.579 ng
 RT: 17.639 min Scan# 2458
 Delta R.T. 0.000 min
 Lab File: BM041181.D
 Acq: 31 Jul 2023 10:00



Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040



Tgt Ion:167 Resp: 1404635
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.3 25.9
 139 12.7 10.3 15.5


Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel




08/01/2023
 Supervised By :mohammad
 ahmed



08/01/2023




#74
 Di-n-butylphthalate
 Concen: 44.058 ng
 RT: 18.198 min Scan# 2552
 Delta R.T. 0.000 min
 Lab File: BM041181.D
 Acq: 31 Jul 2023 10:00




Tgt Ion:149 Resp: 1723334
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 5.3 4.5 6.7





#75
Fluoranthene
Concen: 41.538 ng
RT: 19.292 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 180457
Ion Ratio Lower Upper
202 100
101 10.4 0.0 30.8
203 17.5 0.0 37.6


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.427 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:240 Resp: 649582
Ion Ratio Lower Upper
240 100
120 9.7 8.3 12.5
236 25.6 20.6 30.8





#77

Benzidine

Concen: 51.585 ng

RT: 19.492 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:184 Resp: 456719

Ion Ratio Lower Upper

184 100

185 14.1 11.4 17.0



183 11.3 9.2 13.8

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



08/01/2023

Supervised By :mohammad
ahmed

08/01/2023



#78

Pyrene

Concen: 41.671 ng

RT: 19.656 min Scan# 2800

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00




Tgt Ion:202 Resp: 1892974


Ion Ratio Lower Upper

202 100

200 21.4 16.9 25.3

203 17.7 14.1 21.1





#79

Terphenyl-d14

Concen: 85.032 ng

RT: 19.862 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM041181.D


Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:244 Resp: 305496

Ion Ratio Lower Upper

244 100

212 6.8 5.4 8.2



122 8.9 7.8 11.6

Manual Integrations

APPROVED

Reviewed By :Yogesh


Patel



08/01/2023

Supervised By :mohammad
ahmed

08/01/2023



#80

Butylbenzylphthalate

Concen: 46.343 ng

RT: 20.562 min Scan# 2954

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00




Tgt Ion:149 Resp: 768776


Ion Ratio Lower Upper

149 100

91 72.6 60.8 91.2


206 21.0 16.4 24.6





#81
Benzo(a)anthracene
Concen: 41.928 ng
RT: 21.409 min Scan# 3099
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 1846344
Ion Ratio Lower Upper
228 100
226 27.0 22.4 33.6
229 20.0 15.8 23.8


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




08/01/2023
Supervised By :mohammad
ahmed



08/01/2023




#82
3,3'-Dichlorobenzidine
Concen: 43.202 ng
RT: 21.350 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Tgt Ion:252 Resp: 620787
Ion Ratio Lower Upper
252 100
254 65.7 51.8 77.6
126 10.4 8.6 13.0





#83

Chrysene

Concen: 40.769 ng

RT: 21.462 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 173653

Ion Ratio Lower Upper

228 100

226 29.8 24.5 36.7



229 20.0 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



08/01/2023

Supervised By :mohammad
ahmed

08/01/2023



#84

Bis(2-ethylhexyl)phthalate

Concen: 43.074 ng

RT: 21.339 min Scan# 3086

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00




Tgt Ion:149 Resp: 1127487


Ion Ratio Lower Upper

149 100

167 27.8 21.0 31.6


279 4.9 3.1 4.7#





#85
Di-n-octyl phthalate
Concen: 45.181 ng
RT: 22.256 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 1894228
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 9.2 8.3 12.5


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.797 min Scan# 3504
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:264 Resp: 732286
Ion Ratio Lower Upper
264 100
260 23.4 18.5 27.7
265 22.0 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 40.026 ng

RT: 26.244 min Scan# 3922

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 215493

Ion Ratio Lower Upper

276 100

138 23.0 20.0 30.0



277 25.2 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



08/01/2023

Supervised By :mohammad
ahmed

08/01/2023



#88

Benzo(b)fluoranthene

Concen: 40.553 ng

RT: 23.074 min Scan# 3381

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00




Tgt Ion:252 Resp: 1764153


Ion Ratio Lower Upper

252 100

253 22.4 17.7 26.5


125 9.4 8.2 12.4





#89
Benzo(k)fluoranthene
Concen: 40.986 ng
RT: 23.121 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00



Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:252 Resp: 182384
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.0
125 9.4 7.6 11.4


Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel






08/01/2023
Supervised By :mohammad
ahmed


08/01/2023



#90
Benzo(a)pyrene
Concen: 41.227 ng
RT: 23.691 min Scan# 3486
Delta R.T. 0.000 min
Lab File: BM041181.D
Acq: 31 Jul 2023 10:00

Tgt Ion:252 Resp: 1723258
Ion Ratio Lower Upper
252 100
253 21.8 17.3 25.9
125 10.2 9.0 13.4





#91

Dibenzo(a,h)anthracene

Concen: 40.366 ng

RT: 26.256 min Scan# 3924

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:278 Resp: 181268

Ion Ratio Lower Upper

278 100

139 14.8 12.7 19.1



279 23.5 19.1 28.7

Manual Integrations

APPROVED

Reviewed By :Yogesh



Patel



08/01/2023

Supervised By :mohammad
ahmed

08/01/2023



#92

Benzo(g,h,i)perylene

Concen: 39.317 ng

RT: 26.997 min Scan# 4048

Delta R.T. 0.000 min

Lab File: BM041181.D

Acq: 31 Jul 2023 10:00

Tgt Ion:276 Resp: 1727569

Ion Ratio Lower Upper

276 100

277 24.1 19.1 28.7

138 19.6 17.1 25.7

