

#21

Benzo(b)fluoranthene

Concen: 0.078 ng/ul

RT: 22.83 min Scan# 6493

Delta R.T. 0.01 min

Lab File: BM021775.D

Acq: 03 Aug 2019 04:44

Instrument :

BNA_M

ClientSampleId :


Tot Ion:252 Resp: 629

Ion Ratio Lower Upper

252 100

253 66.3 0.0 62.8#




125 74.9 0.0 62.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#22

Benzo(k)fluoranthene

Concen: 0.069 ng/ul

RT: 22.88 min Scan# 6512

Delta R.T. 0.01 min

Lab File: BM021775.D

Acq: 03 Aug 2019 04:44


Tgt Ion:252 Resp: 587

Ion Ratio Lower Upper

252 100

253 64.2 25.8 38.8#

125 72.8 27.0 40.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E

