

#3

Naphthalene

Concen: 0.090 ng/ul

RT: 10.56 min Scan# 3006

Delta R.T. 0.05 min

Lab File: BM021774.D

Acq: 03 Aug 2019 04:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:128 Resp: 855

Ion Ratio Lower Upper

128 100

129 38.1 11.5 17.3#




127 37.1 13.1 19.7#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#5

2-Methylnaphthalene

Concen: 0.093 ng/ul

RT: 12.20 min Scan# 3372

Delta R.T. 0.06 min

Lab File: BM021774.D


Acq: 03 Aug 2019 04:08

Tgt Ion:142 Resp: 509

Ion Ratio Lower Upper

142 100

141 94.7 72.0 108.0



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

