

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
 Data File : BM031387.D
 Acq On : 02 Aug 2021 22:15
 Operator : CG/JU
 Sample : M3123-09DL 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0070DL

Manual Integrations
 APPROVED

mohammad
 8/3/2021 11:50:53 AM

Quant Time: Aug 03 02:06:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 11:29:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.601	152	1786	0.40	ng/ul	0.00
4) Naphthalene-d8	10.361	136	6790	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.231	164	3958	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.982	188	7194	0.40	ng/ul	0.00
17) Chrysene-d12	21.171	240	5370	0.40	ng/ul	0.00
23) Perylene-d12	23.331	264	5092	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	392	0.20	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.962	152	180	0.02	ng/ul	0.00
18) Fluoranthene-d10	19.008	212	342	0.02	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.415	128	1232	0.06	ng/ul#	86
7) 2-Methylnaphthalene	12.029	142	5736	0.42	ng/ul	100
8) 1-Methylnaphthalene	12.254	142	3053	0.22	ng/ul	100
12) Fluorene	15.285	166	442	0.03	ng/ul#	57
15) Phenanthrene	17.020	178	9145	0.40	ng/ul	99
19) Fluoranthene	19.047	202	1640	0.07	ng/ul#	82
20) Pyrene	19.405	202	1512	0.07	ng/ul#	80
21) Benzo(a)anthracene	21.156	228	492	0.02	ng/ul#	1
22) Chrysene	21.207	228	4024	0.18	ng/ul#	89
24) Benzo(b)fluoranthene	22.691	252	1436m	0.06	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
Data File : BM031387.D
Acq On : 02 Aug 2021 22:15
Operator : CG/JU
Sample : M3123-09DL 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0070DL

Manual Integrations
APPROVED

mohammad
8/3/2021 11:50:53 AM

Quant Time: Aug 03 02:06:58 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 11:29:41 2021
Response via : Initial Calibration

