

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
 Data File : BM031415.D
 Acq On : 03 Aug 2021 17:16
 Operator : CG/JU
 Sample : M3123-17DL 5X
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0078DL

Manual Integrations
 APPROVED

mohammad
 8/4/2021 3:47:52 PM

Quant Time: Aug 03 18:00:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 13:47:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.601	152	1878	0.400	ng/ul	0.00
4) Naphthalene-d8	10.361	136	7213	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.228	164	4066	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.978	188	7784	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.169	240	5970	0.400	ng/ul	0.00
23) Perylene-d12	23.331	264	5814	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.200	96	606	0.291	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.953	152	238	0.020	ng/ul	0.00
18) Fluoranthene-d10	19.009	212	465	0.023	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.406	128	1402	0.065	ng/ul#	89
7) 2-Methylnaphthalene	12.029	142	1184	0.081	ng/ul	98
8) 1-Methylnaphthalene	12.250	142	719	0.050	ng/ul	100
10) Acenaphthylene	13.950	152	1235	0.071	ng/ul#	90
12) Fluorene	15.285	166	434	0.026	ng/ul#	83
15) Phenanthrene	17.016	178	10712	0.433	ng/ul	99
16) Anthracene	17.110	178	1766	0.076	ng/ul#	93
19) Fluoranthene	19.039	202	28967	1.136	ng/ul	97
20) Pyrene	19.401	202	23599	0.913	ng/ul	96
21) Benzo(a)anthracene	21.154	228	11264	0.467	ng/ul	96
22) Chrysene	21.205	228	14896	0.597	ng/ul	98
24) Benzo(b)fluoranthene	22.689	252	20576m	0.785	ng/ul	
25) Benzo(k)fluoranthene	22.728	252	7265m	0.268	ng/ul	
26) Benzo(a)pyrene	23.237	252	10912	0.476	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.468	276	9567	0.323	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.483	278	2338	0.098	ng/ul#	68
29) Benzo(g,h,i)perylene	26.125	276	1236	0.049	ng/ul#	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
Data File : BM031415.D
Acq On : 03 Aug 2021 17:16
Operator : CG/JU
Sample : M3123-17DL 5X
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0078DL

Manual Integrations
APPROVED

mohammad
8/4/2021 3:47:52 PM

Quant Time: Aug 03 18:00:23 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 13:47:07 2021
Response via : Initial Calibration

