

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
 Data File : BM031449.D
 Acq On : 04 Aug 2021 16:06
 Operator : CG/JU
 Sample : M3141-21
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGES2

Quant Time: Aug 04 17:11:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 04 12:17:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.594	152	3394	0.400	ng/ul	0.00
4) Naphthalene-d8	10.352	136	13297	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.224	164	8321	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.971	188	16247	0.400	ng/ul	0.00
17) Chrysene-d12	21.166	240	12588	0.400	ng/ul	0.00
23) Perylene-d12	23.326	264	12539	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.189	96	12346	3.275	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.948	152	4791	0.214	ng/ul	0.00
18) Fluoranthene-d10	19.005	212	9570	0.228	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
Data File : BM031449.D
Acq On : 04 Aug 2021 16:06
Operator : CG/JU
Sample : M3141-21
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGES2

Quant Time: Aug 04 17:11:49 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 04 12:17:22 2021
Response via : Initial Calibration

