

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
 Data File : BM031459.D
 Acq On : 04 Aug 2021 22:15
 Operator : CG/JU
 Sample : M3141-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGEP2

Manual Integrations
 APPROVED

mohammad
 8/5/2021 5:04:52 PM

Quant Time: Aug 05 01:02:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 04 12:17:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.594	152	2219	0.400	ng/ul	0.00
4) Naphthalene-d8	10.352	136	8422	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.224	164	5193	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.971	188	10469	0.400	ng/ul	0.00
17) Chrysene-d12	21.164	240	8679	0.400	ng/ul	0.00
23) Perylene-d12	23.321	264	8870	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.189	96	6774	2.749	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.948	152	2903	0.205	ng/ul	0.00
18) Fluoranthene-d10	19.005	212	10236	0.353	ng/ul	0.00
Target Compounds						
					Qvalue	
19) Fluoranthene	19.035	202	1108	0.030	ng/ul#	72
20) Pyrene	19.397	202	887	0.024	ng/ul#	61
22) Chrysene	21.200	228	737	0.020	ng/ul#	84
24) Benzo(b)fluoranthene	22.684	252	919m	0.023	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
Data File : BM031459.D
Acq On : 04 Aug 2021 22:15
Operator : CG/JU
Sample : M3141-03
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGEP2

Manual Integrations
APPROVED

mohammad
8/5/2021 5:04:52 PM

Quant Time: Aug 05 01:02:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 04 12:17:22 2021
Response via : Initial Calibration

