

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
 Data File : BM031470.D
 Acq On : 05 Aug 2021 05:35
 Operator : CG/JU
 Sample : M3141-08
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGEQ5

Quant Time: Aug 05 06:18:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 05 01:04:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.594	152	1483	0.400	ng/ul	0.00
4) Naphthalene-d8	10.352	136	5672	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.224	164	3623	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.971	188	6908	0.400	ng/ul	0.00
17) Chrysene-d12	21.164	240	5569	0.400	ng/ul	0.00
23) Perylene-d12	23.321	264	5967	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.189	96	8860	5.379	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.948	152	3119	0.326	ng/ul	0.00
18) Fluoranthene-d10	19.005	212	6042	0.325	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
Data File : BM031470.D
Acq On : 05 Aug 2021 05:35
Operator : CG/JU
Sample : M3141-08
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGEQ5

Quant Time: Aug 05 06:18:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 05 01:04:34 2021
Response via : Initial Calibration

