

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
 Data File : BM031370.D
 Acq On : 02 Aug 2021 10:46
 Operator : CG/JU
 Sample : PB137960BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK960

Quant Time: Aug 02 11:34:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 11:29:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.601	152	1083	0.400	ng/ul	0.00
4) Naphthalene-d8	10.361	136	4209	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.231	164	2677	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.982	188	4876	0.400	ng/ul	0.00
17) Chrysene-d12	21.176	240	3345	0.400	ng/ul	# 0.00
23) Perylene-d12	23.336	264	3579	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.189	96	8486	7.055	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.962	152	2829	0.399	ng/ul	0.00
18) Fluoranthene-d10	19.012	212	5262	0.471	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080221\
Data File : BM031370.D
Acq On : 02 Aug 2021 10:46
Operator : CG/JU
Sample : PB137960BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK960

Quant Time: Aug 02 11:34:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 11:29:41 2021
Response via : Initial Calibration

