




Data File : BM021849.D
Acq On : 05 Aug 2019 19:17
Operator : HP/JU
Sample : K3823-01RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Aug 06 01:32:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 06 01:29:26 2019
Response via : Initial Calibration



TIC: BM021849.D

(10) Phenanthrene-d10 (I)
17.136min (+0.087) 0.40ng/ul
response 277075

Ion	Exp%	Act%
188.00	100	100
94.00	10.40	9.27
80.00	11.40	8.82#
0.00	0.00	0.00