

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM080319\
 Data File : BM021801.D
 Acq On : 04 Aug 2019 08:06
 Operator : HP/JU
 Sample : K3863-05
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A52G1

Manual Integrations
 APPROVED

mohammad
 8/5/2019 10:14:58 PM

Quant Time: Aug 04 09:31:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 04 04:24:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.64	152	1078	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.41	136	4223	0.40	ng/ul	-0.03
6) Acenaphthene-d10	14.29	164	2394	0.40	ng/ul	-0.02
10) Phenanthrene-d10	17.04	188	5788m	0.40	ng/ul	-0.03
16) Chrysene-d12	21.25	240	5601m	0.40	ng/ul	0.00
20) Perylene-d12	23.46	264	5357m	0.40	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.04	152	1487	0.21	ng/ul	-0.02
14) Fluoranthene-d10	19.08	212	4299m	0.22	ng/ul	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM080319\
Data File : BM021801.D
Acq On : 04 Aug 2019 08:06
Operator : HP/JU
Sample : K3863-05
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A52G1

Manual Integrations
APPROVED

mohammad
8/5/2019 10:14:58 PM

Quant Time: Aug 04 09:31:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 04 04:24:24 2019
Response via : Initial Calibration

