

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM080323\  
 Data File : BM041270.D  
 Acq On : 03 Aug 2023 20:15  
 Operator : MA/JU  
 Sample : 03756-16  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 RPTP-12-CS-12

Quant Time: Aug 04 01:23:25 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM072823.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 02 19:20:51 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min)    |
|-------------------------------|--------|------|----------|--------|-------|-------------|
| -----                         |        |      |          |        |       |             |
| Internal Standards            |        |      |          |        |       |             |
| 1) 1,4-Dichlorobenzene-d4     | 7.798  | 152  | 123300   | 20.000 | ng    | 0.00        |
| 21) Naphthalene-d8            | 10.604 | 136  | 573239   | 20.000 | ng    | 0.00        |
| 39) Acenaphthene-d10          | 14.462 | 164  | 415100   | 20.000 | ng    | 0.00        |
| 64) Phenanthrene-d10          | 17.221 | 188  | 1011290  | 20.000 | ng    | 0.00        |
| 76) Chrysene-d12              | 21.421 | 240  | 870559   | 20.000 | ng    | -0.01       |
| 86) Perylene-d12              | 23.797 | 264  | 822852   | 20.000 | ng    | 0.00        |
|                               |        |      |          |        |       |             |
| System Monitoring Compounds   |        |      |          |        |       |             |
| 5) 2-Fluorophenol             | 5.369  | 112  | 673335   | 81.619 | ng    | 0.00        |
| 7) Phenol-d6                  | 6.981  | 99   | 934338   | 87.373 | ng    | 0.00        |
| 23) Nitrobenzene-d5           | 8.969  | 82   | 566345   | 45.945 | ng    | 0.00        |
| 42) 2,4,6-Tribromophenol      | 15.962 | 330  | 525293   | 89.467 | ng    | 0.00        |
| 45) 2-Fluorobiphenyl          | 13.080 | 172  | 1491143  | 45.379 | ng    | 0.00        |
| 79) Terphenyl-d14             | 19.862 | 244  | 2960410  | 61.484 | ng    | 0.00        |
|                               |        |      |          |        |       |             |
| Target Compounds              |        |      |          |        |       |             |
| 84) Bis(2-ethylhexyl)phtha... | 21.327 | 149  | 6731     | 2.720  | ng    | Qvalue # 98 |
| -----                         |        |      |          |        |       |             |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM080323\  
Data File : BM041270.D  
Acq On : 03 Aug 2023 20:15  
Operator : MA/JU  
Sample : 03756-16  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
RPTP-12-CS-12

Quant Time: Aug 04 01:23:25 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM072823.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 02 19:20:51 2023  
Response via : Initial Calibration

