

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080818\
 Data File : BM016246.D
 Acq On : 08 Aug 2018 12:47
 Operator : SJ/JU
 Sample : J4327-01MS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampled :

Quant Time: Aug 08 16:24:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 16:23:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	41995	20.00	ng	0.00
21) Naphthalene-d8	10.95	136	155742	20.00	ng	0.00
38) Acenaphthene-d10	14.77	164	77314	20.00	ng	0.00
63) Phenanthrene-d10	17.50	188	155315	20.00	ng	0.00
75) Chrysene-d12	21.63	240	159226	20.00	ng	0.00
86) Perylene-d12	23.97	264	162455	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.66	112	238724	94.43	ng	0.00
7) Phenol-d6	7.30	99	287161	86.98	ng	0.00
23) Nitrobenzene-d5	9.32	82	214715	64.12	ng	0.00
41) 2,4,6-Tribromophenol	16.26	330	84973	86.65	ng	0.00
44) 2-Fluorobiphenyl	13.39	172	388241	61.10	ng	0.00
78) Terphenyl-d14	20.09	244	500709	65.83	ng	0.00

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.45	88	27887	23.543	ng	# 100
3) Pyridine	3.90	79	38757	13.582	ng	# 82
4) n-Nitrosodimethylamine	3.80	42	74583	33.511	ng	# 33
6) Aniline	7.46	93	77424	20.367	ng	# 88
8) 2-Chlorophenol	7.70	128	82883	27.585	ng	96
9) Benzaldehyde	7.28	77	42959	18.520	ng	91
10) Phenol	7.33	94	94419	28.600	ng	97
11) bis(2-Chloroethyl)ether	7.56	93	70837	27.161	ng	87
12) 1,3-Dichlorobenzene	8.02	146	93580	26.608	ng	# 87
13) 1,4-Dichlorobenzene	8.17	146	93510	26.216	ng	# 94
14) 1,2-Dichlorobenzene	8.49	146	92754	26.576	ng	# 92
15) Benzyl Alcohol	8.39	79	74407	28.304	ng	88
16) 2,2'-oxybis(1-Chloropropan	8.65	45	101113	25.332	ng	98
17) 2-Methylphenol	8.59	107	64772	28.705	ng	# 84
18) Hexachloroethane	9.21	117	42677	27.679	ng	94
19) n-Nitroso-di-n-propylamine	8.95	70	62416	25.865	ng	# 71
20) 3+4-Methylphenols	8.92	107	84881	26.145	ng	84
22) Acetophenone	8.98	105	122866	31.331	ng	# 83
24) Nitrobenzene	9.37	77	101619	30.150	ng	97
25) Isophorone	9.88	82	158190	34.538	ng	# 92
26) 2-Nitrophenol	10.07	139	42622	30.293	ng	# 49
27) 2,4-Dimethylphenol	10.13	122	69530	35.470	ng	86
28) bis(2-Chloroethoxy)methane	10.36	93	87853	29.511	ng	98
29) 2,4-Dichlorophenol	10.61	162	72120	30.962	ng	93
30) 1,2,4-Trichlorobenzene	10.82	180	82147	29.026	ng	# 96
31) Naphthalene	11.00	128	242368	30.749	ng	99
32) Benzoic acid	10.26	122	23017	15.403	ng	# 79
33) 4-Chloroaniline	11.13	127	73757	22.930	ng	# 89
34) Hexachlorobutadiene	11.26	225	56873	28.970	ng	98
35) Caprolactam	11.93	113	17637	27.338	ng	93
36) 4-Chloro-3-methylphenol	12.24	107	75792	29.583	ng	# 82
37) 2-Methylnaphthalene	12.60	142	171693	32.517	ng	# 87
39) 1,2,4,5-Tetrachlorobenzene	12.97	216	83172	31.599	ng	# 100
40) Hexachlorocyclopentadiene	12.93	237	75300	60.365	ng	96

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080818\
 Data File : BM016246.D
 Acq On : 08 Aug 2018 12:47
 Operator : SJ/JU
 Sample : J4327-01MS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 08 16:24:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 16:23:37 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.22	196	53561	32.524	ng	97
43) 2,4,5-Trichlorophenol	13.29	196	54410	32.031	ng #	78
45) 1,1'-Biphenyl	13.60	154	208809	29.927	ng	99
46) 2-Chloronaphthalene	13.66	162	162386	29.923	ng #	89
47) 2-Nitroaniline	13.87	65	54094	30.150	ng #	79
48) Acenaphthylene	14.50	152	259084	32.612	ng	99
49) Dimethylphthalate	14.23	163	237155	35.050	ng	99
50) 2,6-Dinitrotoluene	14.37	165	40872	30.022	ng #	51
51) Acenaphthene	14.83	154	146397	29.962	ng	93
52) 3-Nitroaniline	14.70	138	36278	24.671	ng #	80
53) 2,4-Dinitrophenol	14.93	184	30021	53.546	ng #	66
54) Dibenzofuran	15.17	168	237725	29.525	ng #	79
55) 4-Nitrophenol	15.02	139	59542	56.485	ng #	89
56) 2,4-Dinitrotoluene	15.16	165	57082	29.324	ng #	73
57) Fluorene	15.82	166	184907	30.904	ng	98
58) 2,3,4,6-Tetrachlorophenol	15.40	232	46148	31.607	ng #	100
59) Diethylphthalate	15.57	149	214923	29.470	ng	94
60) 4-Chlorophenyl-phenylether	15.80	204	92620	27.807	ng #	78
61) 4-Nitroaniline	15.86	138	37726	26.741	ng #	1
62) Azobenzene	16.09	77	237223	30.886	ng #	73
64) 4,6-Dinitro-2-methylphenol	15.93	198	28396	30.121	ng	86
65) n-Nitrosodiphenylamine	16.02	169	155637	30.432	ng	95
66) 4-Bromophenyl-phenylether	16.69	248	52916	30.088	ng #	69
67) Hexachlorobenzene	16.81	284	56419	29.131	ng #	61
68) Atrazine	16.96	200	56765	30.969	ng	98
69) Pentachlorophenol	17.16	266	57944	48.315	ng	98
70) Phenanthrene	17.55	178	267805	30.798	ng	97
71) Anthracene	17.64	178	270465	32.004	ng	97
72) Carbazole	17.92	167	247310	28.248	ng #	95
73) Di-n-butylphthalate	18.44	149	343554	31.491	ng #	97
74) Fluoranthene	19.54	202	298623	30.787	ng	98
76) Benzidine	19.73	184	38413	8.633	ng	98
77) Pyrene	19.90	202	305249	31.990	ng	99
79) Butylbenzylphthalate	20.76	149	159280	32.859	ng #	80
80) Benzo(a)anthracene	21.62	228	314844	32.104	ng	98
81) 3,3'-Dichlorobenzidine	21.55	252	93235	23.603	ng	98
82) Chrysene	21.67	228	293870	31.238	ng	99
83) Bis(2-ethylhexyl)phthalate	21.51	149	230710	32.841	ng	94
84) Di-n-octyl phthalate	22.41	149	392490	33.242	ng #	88
85) Indeno(1,2,3-cd)pyrene	26.36	276	370416	33.182	ng #	96
87) Benzo(b)fluoranthene	23.27	252	319131	33.365	ng #	95
88) Benzo(k)fluoranthene	23.31	252	294760	30.408	ng #	97
89) Benzo(a)pyrene	23.87	252	301023	32.737	ng	99
90) Dibenzo(a,h)anthracene	26.36	278	311428	32.689	ng #	95
91) Benzo(g,h,i)perylene	27.10	276	297058	32.968	ng #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080818\
 Data File : BM016246.D
 Acq On : 08 Aug 2018 12:47
 Operator : SJ/JU
 Sample : J4327-01MS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 Client Sampled :

Quant Time: Aug 08 16:24:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 16:23:37 2018
 Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.13 min Scan# 824
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
152	41995		
152	100		
150	150.6	119.5	179.3
115	68.1	43.0	64.4#



#2
 1,4-Dioxane
 Concen: 23.543 ng
 RT: 3.45 min Scan# 28
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
88	27887		
88	100		
58	80.1	0.0	0.0#
43	54.0	0.0	0.0#








#3
 Pyridine
 Concen: 13.582 ng
 RT: 3.90 min Scan# 104
 Delta R.T. 0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
79	100		
52	68.5	51.1	76.7
51	55.4	26.1	39.1#



#4
 n-Nitrosodimethylamine
 Concen: 33.511 ng
 RT: 3.80 min Scan# 88
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
42	100		
74	60.6	119.3	178.9#
44	4.2	5.4	8.2#








#5
 2-Fluorophenol
 Concen: 33.511 ng
 RT: 5.66 min Scan# 404
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
112	100		
64	62.8	38.6	57.8#
63	45.2	19.3	28.9#



#6
 Aniline
 Concen: 20.367 ng
 RT: 7.46 min Scan# 710
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
93	100		
66	44.8	30.8	46.2
65	26.3	16.0	24.0#








#7
 Phenol-d6
 Concen: 20.367 ng
 RT: 7.30 min Scan# 683
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
99	100		
42	38.6	11.0	16.4#
71	54.5	23.4	35.2#



#8
 2-Chlorophenol
 Concen: 27.585 ng
 RT: 7.70 min Scan# 750
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
128	100		
130	32.7	12.0	52.0
64	49.0	24.9	64.9









#9
Benzaldehyde
Concen: 18.520 ng
RT: 7.28 min Scan# 679
Delta R.T. -0.00 min
Lab File: BM016246.D
Acq: 08 Aug 2018 12:47




Instrument :
BNA_M
ClientSampled :


Tgt Ion	Resp	Lower	Upper
77	100		
105	90.1	78.8	118.8
106	84.0	73.7	113.7



#10
Phenol
Concen: 28.600 ng
RT: 7.33 min Scan# 688
Delta R.T. 0.01 min
Lab File: BM016246.D
Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
94	100		
65	40.5	18.8	58.8
66	58.1	39.9	79.9








#11
 bis(2-Chloroethyl)ether
 Concen: 27.161 ng
 RT: 7.56 min Scan# 726
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
93	70837		
93	100		
63	84.8	49.5	89.5
95	32.9	13.5	53.5



#12
 1,3-Dichlorobenzene
 Concen: 26.608 ng
 RT: 8.02 min Scan# 805
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
146	93580		
146	100		
148	66.0	50.9	76.3
75	46.6	21.4	32.2#









#13
 1,4-Dichlorobenzene
 Concen: 26.216 ng
 RT: 8.17 min Scan# 830
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
146	93510		
146	100		
148	63.6	51.9	77.9
111	44.3	29.2	43.8#



#14
 1,2-Dichlorobenzene
 Concen: 26.576 ng
 RT: 8.49 min Scan# 884
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
146	92754		
146	100		
148	60.7	52.3	78.5
111	45.9	30.6	45.8#









#15
Benzyl Alcohol
Concen: 28.304 ng
RT: 8.39 min Scan# 868
Delta R.T. -0.00 min
Lab File: BM016246.D
Acq: 08 Aug 2018 12:47




Instrument :
BNA_M
ClientSampled :


Tgt Ion	Resp	Lower	Upper
79	100		
108	67.0	65.0	97.4
77	72.6	53.0	79.6



#16
2,2'-oxybis(1-Chloropropane)
Concen: 25.332 ng
RT: 8.65 min Scan# 912
Delta R.T. -0.01 min
Lab File: BM016246.D
Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
45	100		
77	16.7	0.0	35.2
79	12.0	0.0	32.0








#17
 2-Methylphenol
 Concen: 28.705 ng
 RT: 8.59 min Scan# 902
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Instrument :
 BNA_M
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
107	100		
108	114.4	89.8	134.8
77	65.0	32.6	48.8#
79	59.4	32.8	49.2#




Abundance
 Ion 107.00 (106.70 to 107.70): E
 Ion 108.00 (107.70 to 108.70): E
 Ion 77.00 (76.70 to 77.70): BM
 Ion 79.00 (78.70 to 79.70): BM









#18
 Hexachloroethane
 Concen: 27.679 ng
 RT: 9.21 min Scan# 1007
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
117	100		
119	96.2	79.2	118.8
201	95.7	69.8	104.6



Abundance
 Ion 117.00 (116.70 to 117.70): E
 Ion 119.00 (118.70 to 119.70): E
 Ion 201.00 (200.70 to 201.70): E









#19
 n-Nitroso-di-n-propylamine
 Concen: 25.865 ng
 RT: 8.95 min Scan# 963
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47


Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
70	100		
42	85.5	44.5	66.7#
101	9.5	7.4	11.2
130	22.0	15.3	22.9



#20
 3+4-Methylphenols
 Concen: 26.145 ng
 RT: 8.92 min Scan# 958
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
107	100		
108	84.9	73.2	113.2
77	48.1	10.7	50.7
79	40.8	9.6	49.6









#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.95 min Scan# 1303
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Instrument :
 BNA_M
 ClientSampled :

Tgt Ion:136 Resp: 155742




Ion	Ratio	Lower	Upper
136	100		
137	10.4	8.7	13.1
54	10.5	4.6	6.8#
68	4.4	3.7	5.5




#22
 Acetophenone
 Concen: 31.331 ng
 RT: 8.98 min Scan# 968
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion:105 Resp: 122866

Ion	Ratio	Lower	Upper
105	100		
71	0.0	0.6	1.0#
51	41.5	21.6	32.4#
120	20.9	16.1	24.1









#23
 Nitrobenzene-d5
 Concen: 31.331 ng
 RT: 9.32 min Scan# 1026
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :




Tgt Ion	Resp	Lower	Upper
82	100		
128	36.5	32.8	49.2
54	64.4	40.6	60.8



#24
 Nitrobenzene
 Concen: 30.150 ng
 RT: 9.37 min Scan# 1034
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
77	100		
123	39.4	32.6	49.0
65	16.3	10.9	16.3









#25
 Isophorone
 Concen: 34.538 ng
 RT: 9.88 min Scan# 1121
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47


Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
82	158190		
82	100		
95	11.0	5.8	8.6#
138	17.4	16.3	24.5



#26
 2-Nitrophenol
 Concen: 30.293 ng
 RT: 10.07 min Scan# 1154
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
139	42622		
139	100		
109	54.9	20.1	30.1#
65	67.7	31.5	47.3#









#27
 2,4-Dimethylphenol
 Concen: 35.470 ng
 RT: 10.13 min Scan# 1163
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
122	69530		
107	125.2	84.7	127.1
121	61.4	46.6	69.8



#28
 bis(2-Chloroethoxy)methane
 Concen: 29.511 ng
 RT: 10.36 min Scan# 1202
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
93	87853		
95	30.8	25.5	38.3
123	11.9	8.6	13.0








#29
 2,4-Dichlorophenol
 Concen: 30.962 ng
 RT: 10.61 min Scan# 1245
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
162	100		
164	68.2	42.8	82.8
98	38.1	14.5	54.5



#30
 1,2,4-Trichlorobenzene
 Concen: 29.026 ng
 RT: 10.82 min Scan# 1280
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
180	100		
182	97.1	77.6	116.4
145	37.3	23.4	35.2#









#31
 Naphthalene
 Concen: 30.749 ng
 RT: 11.00 min Scan# 1312
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
128	100		
129	10.6	8.9	13.3
127	13.4	10.4	15.6



#32
 Benzoic acid
 Concen: 15.403 ng
 RT: 10.26 min Scan# 1186
 Delta R.T. -0.03 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
122	100		
105	136.1	99.9	139.9
77	121.2	73.7	113.7#









#33
 4-Chloroaniline
 Concen: 22.930 ng
 RT: 11.13 min Scan# 1333
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
127	100		
129	31.4	25.3	37.9
65	36.7	17.6	26.4
92	19.0	13.1	19.7



#34
 Hexachlorobutadiene
 Concen: 28.970 ng
 RT: 11.26 min Scan# 1355
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
225	100		
223	62.6	47.9	71.9
227	63.5	50.1	75.1









#35
 Caprolactam
 Concen: 27.338 ng
 RT: 11.93 min Scan# 1469
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
113	100		
55	163.0	145.9	185.9
56	137.2	101.0	141.0



#36
 4-Chloro-3-methylphenol
 Concen: 29.583 ng
 RT: 12.24 min Scan# 1523
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
107	100		
144	22.7	23.3	34.9#
142	71.3	72.6	109.0#









#37
 2-Methylnaphthalene
 Concen: 32.517 ng
 RT: 12.60 min Scan# 1584
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
142	171693		
141	94.1	71.9	107.9
115	52.4	25.4	38.0



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.77 min Scan# 1953
 Delta R.T. 0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
164	77314		
162	111.5	82.4	123.6
160	51.5	35.8	53.6









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 31.599 ng
 RT: 12.97 min Scan# 1646
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Instrument :
 BNA_M
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
216	100		
214	78.7	0.0	0.0#
179	25.2	0.0	0.0#
108	25.9	0.0	0.0#




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 60.365 ng
 RT: 12.93 min Scan# 1639
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
237	100		
235	65.0	42.0	82.0
272	12.3	0.0	33.4



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E








#41
 2,4,6-Tribromophenol
 Concen: 60.365 ng
 RT: 16.26 min Scan# 2206
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
330	100		
332	96.2	0.0	0.0#
141	73.3	0.0	0.0#



#42
 2,4,6-Trichlorophenol
 Concen: 32.524 ng
 RT: 13.22 min Scan# 1688
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
196	100		
198	93.3	76.3	114.5
200	28.2	25.6	38.4








#43
 2,4,5-Trichlorophenol
 Concen: 32.031 ng
 RT: 13.29 min Scan# 1701
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
196	100		
198	91.0	79.4	119.0
97	71.5	26.8	40.2#
132	31.1	18.6	28.0#



#44
 2-Fluorobiphenyl
 Concen: 32.031 ng
 RT: 13.39 min Scan# 1718
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
172	100		
171	36.3	28.5	42.7
170	24.2	18.8	28.2









#45
 1,1'-Biphenyl
 Concen: 29.927 ng
 RT: 13.60 min Scan# 1754
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
154	100		
153	41.3	20.7	60.7
76	11.3	0.0	30.9



#46
 2-Chloronaphthalene
 Concen: 29.923 ng
 RT: 13.66 min Scan# 1763
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
162	100		
127	43.9	26.9	40.3#
164	31.7	26.8	40.2









#47
 2-Nitroaniline
 Concen: 30.150 ng
 RT: 13.87 min Scan# 1800
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
65	54094		
92	62.2	59.1	88.7
138	88.6	93.9	140.9#



#48
 Acenaphthylene
 Concen: 32.612 ng
 RT: 14.50 min Scan# 1906
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
152	259084		
151	20.2	15.9	23.9
153	14.3	10.6	16.0








#49
 Dimethylphthalate
 Concen: 35.050 ng
 RT: 14.23 min Scan# 1860
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Instrument :
 BNA_M
 ClientSampled :

Tgt Ion:163 Resp: 237155





Ion	Ratio	Lower	Upper
163	100		
194	3.5	2.7	4.1
164	10.6	8.2	12.2




#50
 2,6-Dinitrotoluene
 Concen: 30.022 ng
 RT: 14.37 min Scan# 1884
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion:165 Resp: 40872

Ion	Ratio	Lower	Upper
165	100		
63	94.5	44.1	66.1#
89	86.8	44.3	66.5#









#51
 Acenaphthene
 Concen: 29.962 ng
 RT: 14.83 min Scan# 1963
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
154	100		
153	119.7	90.0	135.0
152	57.8	42.6	63.8



#52
 3-Nitroaniline
 Concen: 24.671 ng
 RT: 14.70 min Scan# 1941
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
138	100		
108	8.9	7.3	10.9
92	116.6	76.6	114.8#









#53
 2,4-Dinitrophenol
 Concen: 53.546 ng
 RT: 14.93 min Scan# 1979
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
184	100		
63	106.9	69.2	103.8#
154	106.6	53.3	79.9#



#54
 Dibenzofuran
 Concen: 29.525 ng
 RT: 15.17 min Scan# 2020
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
168	100		
139	50.7	27.3	40.9#
169	13.0	10.6	16.0








#55
 4-Nitrophenol
 Concen: 56.485 ng
 RT: 15.02 min Scan# 1994
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
139	100		
109	146.5	130.0	170.0
65	126.1	130.0	170.0#



#56
 2,4-Dinitrotoluene
 Concen: 29.324 ng
 RT: 15.16 min Scan# 2019
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
165	100		
63	99.9	52.4	78.6#
89	105.6	72.4	108.6
182	2.7	2.0	3.0








#57
 Fluorene
 Concen: 30.904 ng
 RT: 15.82 min Scan# 2130
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
166	100		
165	101.5	79.0	118.4
167	12.9	10.3	15.5



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 31.607 ng
 RT: 15.40 min Scan# 2059
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
232	100		
131	68.0	0.0	0.0#
130	4.6	0.0	0.0#
166	39.5	0.0	0.0#









#59
 Diethylphthalate
 Concen: 29.470 ng
 RT: 15.57 min Scan# 2089
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Instrument :
 BNA_M
 ClientSampled :

Tgt Ion:149 Resp: 214923




Ion	Ratio	Lower	Upper
149	100		
177	19.3	18.3	27.5
150	13.2	9.8	14.8




#60
 4-Chlorophenyl-phenylether
 Concen: 27.807 ng
 RT: 15.80 min Scan# 2127
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion:204 Resp: 92620

Ion	Ratio	Lower	Upper
204	100		
206	33.2	26.6	39.8
141	82.3	45.2	67.8#








#61
 4-Nitroaniline
 Concen: 26.741 ng
 RT: 15.86 min Scan# 2138
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
138	100		
92	64.6	16.7	56.7#
108	162.7	37.6	77.6#



#62
 Azobenzene
 Concen: 30.886 ng
 RT: 16.09 min Scan# 2177
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
77	100		
182	18.9	6.5	46.5
105	14.9	3.8	43.8
51	49.3	5.5	45.5#








#63
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.50 min Scan# 2417
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
188	100		
94	7.3	8.7	13.1#
80	8.9	9.3	13.9#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 30.121 ng
 RT: 15.93 min Scan# 2149
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
198	100		
51	67.7	28.8	68.8
105	49.7	30.5	70.5








#65
 n-Nitrosodiphenylamine
 Concen: 30.432 ng
 RT: 16.02 min Scan# 2165
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47





Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
169	100		
168	71.0	53.9	80.9
167	39.3	28.9	43.3



#66
 4-Bromophenyl-phenylether
 Concen: 30.088 ng
 RT: 16.69 min Scan# 2279
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
248	100		
250	102.0	77.4	116.0
141	111.9	45.2	67.8#









#67
 Hexachlorobenzene
 Concen: 29.131 ng
 RT: 16.81 min Scan# 2299
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
284	100		
142	59.4	20.4	30.6#
249	38.2	23.8	35.6#



#68
 Atrazine
 Concen: 30.969 ng
 RT: 16.96 min Scan# 2325
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
200	100		
173	24.7	4.8	44.8
215	45.2	26.9	66.9









#69
 Pentachlorophenol
 Concen: 48.315 ng
 RT: 17.16 min Scan# 2359
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
266	100		
268	62.3	52.0	78.0
264	61.8	49.0	73.4



#70
 Phenanthrene
 Concen: 30.798 ng
 RT: 17.55 min Scan# 2425
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
178	100		
176	20.8	15.2	22.8
179	15.6	12.1	18.1









#71
 Anthracene
 Concen: 32.004 ng
 RT: 17.64 min Scan# 2440
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
178	100		
176	19.7	14.6	22.0
179	14.7	12.6	19.0



#72
 Carbazole
 Concen: 28.248 ng
 RT: 17.92 min Scan# 2487
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
167	100		
166	22.7	17.2	25.8
139	16.9	10.8	16.2#









#73
 Di-n-butylphthalate
 Concen: 31.491 ng
 RT: 18.44 min Scan# 2576
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	7.8	3.7	5.5#



#74
 Fluoranthene
 Concen: 30.787 ng
 RT: 19.54 min Scan# 2763
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
202	100		
101	7.4	0.0	28.7
203	17.6	0.0	37.2









#75
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.63 min Scan# 3119
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
240	159226		
120	8.1	8.2	12.2#
236	27.2	20.0	30.0



#76
 Benzidine
 Concen: 8.633 ng
 RT: 19.73 min Scan# 2795
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
184	38413		
185	13.3	11.3	16.9
183	11.5	8.9	13.3









#77
 Pyrene
 Concen: 31.990 ng
 RT: 19.90 min Scan# 2825
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
202	100		
200	22.1	16.9	25.3
203	17.3	14.1	21.1



#78
 Terphenyl-d14
 Concen: 31.990 ng
 RT: 20.09 min Scan# 2856
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
244	100		
212	9.8	4.9	7.3#
122	9.9	6.2	9.4#









#79
 Butylbenzylphthalate
 Concen: 32.859 ng
 RT: 20.76 min Scan# 2971
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	100		
91	81.6	50.1	75.1#
206	17.6	16.2	24.2



#80
 Benzo(a)anthracene
 Concen: 32.104 ng
 RT: 21.62 min Scan# 3117
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
228	100		
226	28.6	21.7	32.5
229	19.5	16.0	24.0









#81
 3,3'-Dichlorobenzidine
 Concen: 23.603 ng
 RT: 21.55 min Scan# 3105
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
252	100		
254	66.3	51.9	77.9
126	13.5	9.8	14.8



#82
 Chrysene
 Concen: 31.238 ng
 RT: 21.67 min Scan# 3126
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
228	100		
226	30.4	24.1	36.1
229	18.8	15.5	23.3









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 32.841 ng
 RT: 21.51 min Scan# 3098
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	100		
167	24.5	22.4	33.6
279	3.4	3.4	5.0



#84
 Di-n-octyl phthalate
 Concen: 33.242 ng
 RT: 22.41 min Scan# 3251
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	14.1	7.3	10.9#









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 33.182 ng
 RT: 26.36 min Scan# 3923
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
276	100		
138	19.2	12.0	18.0#
277	25.4	20.0	30.0



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.97 min Scan# 3517
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
264	100		
260	23.2	18.6	27.8
265	21.9	17.3	25.9









#87
 Benzo(b)fluoranthene
 Concen: 33.365 ng
 RT: 23.27 min Scan# 3397
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
252	100		
253	22.7	18.0	27.0
125	7.4	9.9	14.9#



#88
 Benzo(k)fluoranthene
 Concen: 30.408 ng
 RT: 23.31 min Scan# 3404
 Delta R.T. -0.01 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
252	100		
253	20.9	17.2	25.8
125	7.9	8.1	12.1#









#89
 Benzo(a)pyrene
 Concen: 32.737 ng
 RT: 23.87 min Scan# 3500
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47




Instrument :
 BNA_M
 ClientSampled :




Tgt Ion	Resp	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	8.6	7.4	11.2



#90
 Dibenzo(a,h)anthracene
 Concen: 32.689 ng
 RT: 26.36 min Scan# 3923
 Delta R.T. 0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Tgt Ion	Resp	Lower	Upper
278	100		
139	13.4	13.4	20.0#
279	24.8	19.0	28.4





#91
 Benzo(a,h,i)perylene
 Concen: 32.968 ng
 RT: 27.10 min Scan# 4048
 Delta R.T. -0.00 min
 Lab File: BM016246.D
 Acq: 08 Aug 2018 12:47

Instrument :
 BNA_M
 ClientSampled :

Tot Ion: 276 Resp: 297058

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.5	27.7
138	17.4	19.0	28.6#

