

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080818\
 Data File : BM016249.D
 Acq On : 08 Aug 2018 14:37
 Operator : SJ/JU
 Sample : J4313-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C-7

Manual Integrations
 APPROVED

Sohil
 8/9/2018 2:23:47 PM

Quant Time: Aug 08 17:11:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 16:23:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	39961	20.00	ng	0.00
21) Naphthalene-d8	10.95	136	140480	20.00	ng	0.00
38) Acenaphthene-d10	14.77	164	67458	20.00	ng	0.00
63) Phenanthrene-d10	17.50	188	138060	20.00	ng	0.00
75) Chrysene-d12	21.63	240	149280	20.00	ng	0.00
86) Perylene-d12	23.97	264	164252	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.66	112	242494	100.80	ng	0.00
7) Phenol-d6	7.30	99	287992	91.67	ng	0.00
23) Nitrobenzene-d5	9.32	82	215958	71.50	ng	0.00
41) 2,4,6-Tribromophenol	16.26	330	80352	93.91	ng	0.00
44) 2-Fluorobiphenyl	13.39	172	397784	71.75	ng	0.00
78) Terphenyl-d14	20.09	244	527376	73.95	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.23	163	29460	4.990	ng	Qvalue # 99
70) Phenanthrene	17.55	178	52995	6.856	ng	96
74) Fluoranthene	19.54	202	91854	10.653	ng	98
77) Pyrene	19.90	202	87275	9.756	ng	99
80) Benzo(a)anthracene	21.62	228	56652	6.162	ng	96
82) Chrysene	21.67	228	48999	5.556	ng	96
83) Bis(2-ethylhexyl)phthalate	21.51	149	18340	2.785	ng	# 90
85) Indeno(1,2,3-cd)pyrene	26.36	276	36635	3.500	ng	# 95
87) Benzo(b)fluoranthene	23.27	252	71269	7.370	ng	95
88) Benzo(k)fluoranthene	23.31	252	19999m	2.041	ng	
89) Benzo(a)pyrene	23.87	252	48804	5.249	ng	# 97
91) Benzo(a,h,i)perylene	27.10	276	34545	3.792	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080818\
Data File : BM016249.D
Acq On : 08 Aug 2018 14:37
Operator : SJ/JU
Sample : J4313-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C-7

Manual Integrations
APPROVED

Sohil
8/9/2018 2:23:47 PM

Quant Time: Aug 08 17:11:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM072318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 16:23:37 2018
Response via : Initial Calibration

