

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080823\
 Data File : BM041322.D
 Acq On : 08 Aug 2023 09:33
 Operator : MA/JU
 Sample : PB154641BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB154641BL

Quant Time: Aug 09 01:11:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM072823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 17:25:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	128779	20.000	ng	0.00
21) Naphthalene-d8	10.586	136	545590	20.000	ng	0.00
39) Acenaphthene-d10	14.451	164	353455	20.000	ng	0.00
64) Phenanthrene-d10	17.209	188	798955	20.000	ng	0.00
76) Chrysene-d12	21.421	240	695484	20.000	ng	0.00
86) Perylene-d12	23.791	264	698381	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	1098503	127.490	ng	0.00
7) Phenol-d6	6.963	99	1396920	125.073	ng	0.00
23) Nitrobenzene-d5	8.957	82	882904	75.256	ng	0.00
42) 2,4,6-Tribromophenol	15.951	330	626979	125.410	ng	0.00
45) 2-Fluorobiphenyl	13.063	172	2061656	73.684	ng	0.00
79) Terphenyl-d14	19.850	244	3549266	92.270	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080823\
Data File : BM041322.D
Acq On : 08 Aug 2023 09:33
Operator : MA/JU
Sample : PB154641BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB154641BL

Quant Time: Aug 09 01:11:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM072823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 17:25:06 2023
Response via : Initial Calibration

