

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080919\
 Data File : BM021965.D
 Acq On : 09 Aug 2019 11:45
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.446

Quant Time: Aug 10 05:22:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 09 01:29:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	126	0.00
3	Naphthalene	1.130	1.028	9.0	114	0.00
4 SURR	2-Methylnaphthalene-d10	0.660	0.622	5.8	120	0.00
5	2-Methylnaphthalene	0.753	0.698	7.3	117	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
7	Acenaphthylene	1.866	1.634	12.4	102	0.00
8 C	Acenaphthene	1.526	1.448	5.1	112	0.00
9	Fluorene	1.695	1.540	9.1	106	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
11	Pentachlorophenol	0.074	0.063	14.9	102	0.00
12	Phenanthrene	1.164	1.099	5.6	102	0.00
13	Anthracene	0.993	0.969	2.4	106	0.00
14 SURR	Fluoranthene-d10	1.349	1.206	10.6	95	0.00
15 C	Fluoranthene	1.559	1.350	13.4	94	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
17	Pyrene	1.644	1.466	10.8	91	0.00
18	Benzo(a)anthracene	1.431	1.303	8.9	94	0.00
19	Chrysene	1.840	1.576	14.3	89	0.00
20 I	Perylene-d12	1.000	1.000	0.0	104	0.00
21	Benzo(b)fluoranthene	1.349	1.235	8.5	93	0.00
22	Benzo(k)fluoranthene	1.526	1.369	10.3	92	0.00
23 C	Benzo(a)pyrene	1.372	1.235	10.0	94	0.00
24	Indeno(1,2,3-cd)pyrene	1.638	1.492	8.9	94	0.00
25	Dibenzo(a,h)anthracene	1.324	1.205	9.0	94	0.00
26	Benzo(g,h,i)perylene	1.475	1.303	11.7	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0