

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM080919\
 Data File : BM021990.D
 Acq On : 10 Aug 2019 03:53
 Operator : HP/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.448

Quant Time: Aug 10 05:24:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 09 01:29:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
3	Naphthalene	1.130	0.990	12.4	91	0.00
4 SURR	2-Methylnaphthalene-d10	0.660	0.596	9.7	95	0.00
5	2-Methylnaphthalene	0.753	0.670	11.0	92	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
7	Acenaphthylene	1.866	1.588	14.9	82	0.00
8 C	Acenaphthene	1.526	1.395	8.6	90	0.00
9	Fluorene	1.695	1.462	13.7	84	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
11	Pentachlorophenol	0.074	0.063	14.9	85	0.00
12	Phenanthrene	1.164	1.033	11.3	79	0.00
13	Anthracene	0.993	0.895	9.9	81	0.00
14 SURR	Fluoranthene-d10	1.349	1.135	15.9	74	0.00
15 C	Fluoranthene	1.559	1.252	19.7	72	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
17	Pyrene	1.644	1.442	12.3	71	0.00
18	Benzo(a)anthracene	1.431	1.230	14.0	70	0.00
19	Chrysene	1.840	1.547	15.9	69	0.00
20 I	Perylene-d12	1.000	1.000	0.0	83	0.00
21	Benzo(b)fluoranthene	1.349	1.178	12.7	70	0.00
22	Benzo(k)fluoranthene	1.526	1.328	13.0	71	0.00
23 C	Benzo(a)pyrene	1.372	1.167	14.9	70	0.00
24	Indeno(1,2,3-cd)pyrene	1.638	1.422	13.2	71	0.00
25	Dibenzo(a,h)anthracene	1.324	1.145	13.5	71	0.00
26	Benzo(g,h,i)perylene	1.475	1.263	14.4	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0