

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM082019\
 Data File : BM022201.D
 Acq On : 19 Aug 2019 22:08
 Operator : HP/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 20 02:21:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 20 02:18:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	37332	20.00	ng	0.00
21) Naphthalene-d8	10.35	136	156341	20.00	ng	0.00
39) Acenaphthene-d10	14.23	164	99102	20.00	ng	0.00
64) Phenanthrene-d10	16.99	188	201347	20.00	ng	0.00
76) Chrysene-d12	21.20	240	228325	20.00	ng	0.00
87) Perylene-d12	23.39	264	232517	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.18	112	176383	84.19	ng	0.00
7) Phenol-d6	6.76	99	232126	83.20	ng	0.00
23) Nitrobenzene-d5	8.75	82	264167	82.18	ng	0.00
42) 2,4,6-Tribromophenol	15.74	330	129648	80.83	ng	0.00
45) 2-Fluorobiphenyl	12.84	172	630453	79.46	ng	0.00
79) Terphenyl-d14	19.63	244	1196597	76.36	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.17	88	37821	43.088	ng	# 92
3) Pyridine	3.56	79	84521	38.009	ng	# 85
4) n-Nitrosodimethylamine	3.49	42	56201	42.185	ng	# 61
6) Aniline	6.92	93	148568	39.775	ng	96
8) 2-Chlorophenol	7.14	128	104902	42.055	ng	93
9) Benzaldehyde	6.74	77	75531	42.723	ng	89
10) Phenol	6.79	94	123493	41.812	ng	85
11) bis(2-Chloroethyl)ether	7.02	93	94238	40.334	ng	94
12) 1,3-Dichlorobenzene	7.45	146	122327	40.236	ng	# 91
13) 1,4-Dichlorobenzene	7.60	146	126698	41.048	ng	97
14) 1,2-Dichlorobenzene	7.92	146	121479	40.411	ng	97
15) Benzyl Alcohol	7.83	79	98135	40.016	ng	91
16) 2,2'-oxybis(1-Chloropropan	8.09	45	128866	39.808	ng	99
17) 2-Methylphenol	8.02	107	85445	40.560	ng	# 91
18) Hexachloroethane	8.62	117	53993	40.292	ng	92
19) n-Nitroso-di-n-propylamine	8.39	70	87632	39.878	ng	# 86
20) 3+4-Methylphenols	8.36	107	115172	40.482	ng	94
22) Acetophenone	8.40	105	160886	40.114	ng	# 90
24) Nitrobenzene	8.79	77	136605	41.508	ng	95
25) Isophorone	9.30	82	231670	40.611	ng	# 97
26) 2-Nitrophenol	9.48	139	58936	42.277	ng	# 75
27) 2,4-Dimethylphenol	9.54	122	84245	40.092	ng	89
28) bis(2-Chloroethoxy)methane	9.78	93	138018	40.896	ng	100
29) 2,4-Dichlorophenol	10.01	162	103659	40.880	ng	98
30) 1,2,4-Trichlorobenzene	10.21	180	125992	40.339	ng	98
31) Naphthalene	10.40	128	332723	40.976	ng	100
32) Benzoic acid	9.73	122	66324	36.975	ng	93
33) 4-Chloroaniline	10.53	127	135545	39.887	ng	95
34) Hexachlorobutadiene	10.65	225	109903	41.009	ng	100
35) Caprolactam	11.37	113	26545	38.945	ng	92
36) 4-Chloro-3-methylphenol	11.66	107	111982	41.781	ng	95
37) 2-Methylnaphthalene	12.02	142	240444	40.512	ng	# 93
38) 1-Methylnaphthalene	12.24	142	234100	41.294	ng	# 99
40) 1,2,4,5-Tetrachlorobenzene	12.39	216	158091	40.754	ng	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM082019\
 Data File : BM022201.D
 Acq On : 19 Aug 2019 22:08
 Operator : HP/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 20 02:21:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 20 02:18:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.35	237	72485	38.189	ng	99
43) 2,4,6-Trichlorophenol	12.65	196	91983	39.556	ng	98
44) 2,4,5-Trichlorophenol	12.73	196	94813	40.168	ng #	94
46) 1,1'-Biphenyl	13.05	154	316373	40.593	ng	98
47) 2-Chloronaphthalene	13.09	162	254960	39.749	ng	96
48) 2-Nitroaniline	13.33	65	82699	40.790	ng	85
49) Acenaphthylene	13.95	152	403474	41.010	ng	100
50) Dimethylphthalate	13.70	163	338019	40.569	ng	100
51) 2,6-Dinitrotoluene	13.84	165	69911	42.488	ng #	68
52) Acenaphthene	14.29	154	233946	40.775	ng	99
53) 3-Nitroaniline	14.17	138	67710	41.321	ng	98
54) 2,4-Dinitrophenol	14.40	184	36301	41.340	ng #	67
55) Dibenzofuran	14.63	168	387786	40.674	ng #	91
56) 4-Nitrophenol	14.50	139	46815	42.867	ng #	54
57) 2,4-Dinitrotoluene	14.64	165	98360	41.640	ng #	52
58) Fluorene	15.29	166	322729	40.154	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.87	232	95027	40.380	ng #	93
60) Diethylphthalate	15.07	149	358242	40.773	ng	97
61) 4-Chlorophenyl-phenylether	15.28	204	198418	39.171	ng	93
62) 4-Nitroaniline	15.35	138	63488	41.861	ng #	49
63) Azobenzene	15.58	77	348974	41.218	ng	84
65) 4,6-Dinitro-2-methylphenol	15.41	198	52119	41.149	ng	89
66) n-Nitrosodiphenylamine	15.51	169	262531	40.664	ng	99
67) 4-Bromophenyl-phenylether	16.18	248	120205	40.288	ng #	91
68) Hexachlorobenzene	16.29	284	136033	40.131	ng #	88
69) Atrazine	16.47	200	69838	32.667	ng	96
70) Pentachlorophenol	16.64	266	66260	40.178	ng	98
71) Phenanthrene	17.03	178	480231	41.129	ng	98
72) Anthracene	17.13	178	476298	41.093	ng	99
73) Carbazole	17.41	167	408949	42.283	ng #	97
74) Di-n-butylphthalate	17.96	149	611127	41.646	ng #	98
75) Fluoranthene	19.06	202	606488	43.073	ng	98
77) Benzidine	19.27	184	188245	36.758	ng	100
78) Pyrene	19.43	202	619966	38.576	ng	100
80) Butylbenzylphthalate	20.33	149	280775	39.060	ng #	86
81) Benzo(a)anthracene	21.18	228	654206	40.742	ng	99
82) 3,3'-Dichlorobenzidine	21.13	252	227250	40.591	ng	99
83) Chrysene	21.23	228	623606	40.231	ng	99
84) Bis(2-ethylhexyl)phthalate	21.10	149	418382	38.701	ng	94
85) Di-n-octyl phthalate	21.97	149	698720	39.932	ng #	94
86) Indeno(1,2,3-cd)pyrene	25.60	276	737026	42.448	ng	97
88) Benzo(b)fluoranthene	22.74	252	615062	39.616	ng	99
89) Benzo(k)fluoranthene	22.78	252	610249	40.210	ng	99
90) Benzo(a)pyrene	23.30	252	566560	40.070	ng #	98
91) Dibenzo(a,h)anthracene	25.60	278	616995	41.068	ng #	97
92) Benzo(g,h,i)perylene	26.27	276	545001	41.493	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.57 min Scan# 779
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampleId :

Tot Ion:152 Resp: 37332
Ion Ratio Lower Upper
152 100
150 152.8 123.1 184.7
115 62.2 44.2 66.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 43.088 ng
RT: 3.17 min Scan# 31
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion: 88 Resp: 37821
Ion Ratio Lower Upper
88 100
58 68.3 51.9 77.9
43 35.5 21.4 32.2#



Abundance Ion 88.00 (87.70 to 88.70): BM0
Ion 58.00 (57.70 to 58.70): BM0
Ion 43.00 (42.70 to 43.70): BM0





#3

Pvridine

Concen: 38.009 ng

RT: 3.56 min Scan# 98

Delta R.T. -0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :





Tot Ion: 79 Resp: 84521

Ion Ratio Lower Upper

79 100

52 69.6 48.8 73.2

51 44.4 25.7 38.5#



#4

n-Nitrosodimethylamine

Concen: 42.185 ng

RT: 3.49 min Scan# 86

Delta R.T. 0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08




Tgt Ion: 42 Resp: 56201


Ion Ratio Lower Upper

42 100

74 91.4 113.8 170.6#

44 5.1 8.2 12.2#





#5

2-Fluorophenol

Concen: 84.193 ng

RT: 5.18 min Scan# 373

Delta R.T. -0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampled :


Tot Ion: 112 Resp: 176383

Ion Ratio Lower Upper

112 100

64 59.9 45.2 67.8




63 38.3 23.5 35.3#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM

Ion 63.00 (62.70 to 63.70): BM



#6

Aniline

Concen: 39.775 ng

RT: 6.92 min Scan# 669

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion: 93 Resp: 148568

Ion Ratio Lower Upper

93 100

66 41.7 31.3 46.9



65 21.9 15.9 23.9




Abundance Ion 93.00 (92.70 to 93.70): BM

Ion 66.00 (65.70 to 66.70): BM

Ion 65.00 (64.70 to 65.70): BM





#7

Phenol-d6

Concen: 83.204 ng

RT: 6.76 min Scan# 642

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion: 99 Resp: 232126

Ion Ratio Lower Upper

99 100

42 26.7 13.5 20.3#




71 47.9 29.3 43.9#



Abundance Ion 99.00 (98.70 to 99.70): BM022201.D (-625) (-)

Ion 42.00 (41.70 to 42.70): BM022201.D (-625) (-)

Ion 71.00 (70.70 to 71.70): BM022201.D (-625) (-)



#8

2-Chlorophenol

Concen: 42.055 ng

RT: 7.14 min Scan# 706

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion: 128 Resp: 104902

Ion Ratio Lower Upper

128 100

130 30.7 12.2 52.2



64 49.2 23.2 63.2




Abundance Ion 128.00 (127.70 to 128.70): BM022201.D (-689) (-)

Ion 130.00 (129.70 to 130.70): BM022201.D (-689) (-)

Ion 64.00 (63.70 to 64.70): BM022201.D (-689) (-)





#9

Benzaldehyde

Concen: 42.723 ng

RT: 6.74 min Scan# 638

Delta R.T. -0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampled :


Tot Ion: 77 Resp: 75531

Ion Ratio Lower Upper

77 100

105 83.1 77.8 117.8




106 85.3 71.8 111.8



Abundance Ion 77.00 (76.70 to 77.70): BM022201.D (-622) (-)

Ion 105.00 (104.70 to 105.70): BM022201.D (-622) (-)

Ion 106.00 (105.70 to 106.70): BM022201.D (-622) (-)



#10

Phenol

Concen: 41.812 ng

RT: 6.79 min Scan# 647

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion: 94 Resp: 123493

Ion Ratio Lower Upper

94 100

65 34.0 8.7 48.7



66 50.6 19.0 59.0




Abundance Ion 94.00 (93.70 to 94.70): BM022201.D (-630) (-)

Ion 65.00 (64.70 to 65.70): BM022201.D (-630) (-)

Ion 66.00 (65.70 to 66.70): BM022201.D (-630) (-)









#11
bis(2-Chloroethyl)ether
Concen: 40.334 ng
RT: 7.02 min Scan# 686
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08




Instrument :
BNA_M
ClientSampleId :


Tot Ion: 93 Resp: 94238
Ion Ratio Lower Upper
93 100
63 78.6 53.4 93.4
95 30.5 13.0 53.0



#12
1,3-Dichlorobenzene
Concen: 40.236 ng
RT: 7.45 min Scan# 759
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion: 146 Resp: 122327
Ion Ratio Lower Upper
146 100
148 62.9 52.2 78.2
75 39.3 22.2 33.4#









#13
 1,4-Dichlorobenzene
 Concen: 41.048 ng
 RT: 7.60 min Scan# 785
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampleId :




Tot Ion	Ratio	Lower	Upper
146	100		
148	62.9	51.9	77.9
111	41.4	31.4	47.0



#14
 1,2-Dichlorobenzene
 Concen: 40.411 ng
 RT: 7.92 min Scan# 838
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.6	51.2	76.8
111	44.0	32.8	49.2





#15

Benzyl Alcohol

Concen: 40.016 ng

RT: 7.83 min Scan# 824

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08





Tot Ion: 79 Resp: 98135

Ion Ratio Lower Upper

79 100

108 67.6 58.4 87.6

77 71.2 50.1 75.1



#16

2,2'-oxybis(1-Chloropropane)

Concen: 39.808 ng

RT: 8.09 min Scan# 867

Delta R.T. -0.02 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion: 45 Resp: 128866


Ion Ratio Lower Upper

45 100

77 17.2 0.0 37.5

79 12.5 0.0 33.5









#17
2-Methylphenol
Concen: 40.560 ng
RT: 8.02 min Scan# 856
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08




Instrument :
BNA_M
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
107	100		
108	111.4	87.5	131.3
77	59.3	37.8	56.8#
79	59.3	37.5	56.3#



#18
Hexachloroethane
Concen: 40.292 ng
RT: 8.62 min Scan# 957
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
117	100		
119	95.7	77.5	116.3
201	113.6	102.3	153.5





#19

n-Nitroso-di-n-propylamine

Concen: 39.878 ng

RT: 8.39 min Scan# 918

Delta R.T. 0.00 min

Lab File: BM022201.D


Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :

Tot Ion:	70	Resp:	87632
Ion	Ratio	Lower	Upper
70	100		
42	66.7	41.9	62.9#
101	10.2	8.1	12.1
130	18.7	17.5	26.3






Abundance

Ion 70.00 (69.70 to 70.70): BM0

Ion 42.00 (41.70 to 42.70): BM0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20

3+4-Methylphenols

Concen: 40.482 ng


RT: 8.36 min Scan# 913

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Tgt Ion:	107	Resp:	115172
Ion	Ratio	Lower	Upper
107	100		
108	87.2	67.9	107.9
77	42.9	14.6	54.6
79	37.3	10.8	50.8





Abundance


Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.35 min Scan# 1251
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampled :

Tot Ion:	136	Resp:	156341
Ion	Ratio	Lower	Upper
136	100		
137	10.4	8.6	13.0
54	9.0	5.3	7.9#
68	5.1	4.4	6.6






Abundance

Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BM

Ion 68.00 (67.70 to 68.70): BM



#22
Acetophenone
Concen: 40.114 ng
RT: 8.40 min Scan# 921
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion:	105	Resp:	160886
Ion	Ratio	Lower	Upper
105	100		
71	1.8	1.2	1.8#
51	32.5	19.1	28.7#
120	21.1	17.9	26.9





Abundance


Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BM

Ion 51.00 (50.70 to 51.70): BM

Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 82.181 ng

RT: 8.75 min Scan# 979

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampled :


Tot Ion: 82 Resp: 264167

Ion Ratio Lower Upper




82 100

128 37.3 35.9 53.9

54 53.3 36.6 55.0



Abundance Ion 82.00 (81.70 to 82.70): BM022201.D (-962) (-)



#24

Nitrobenzene

Concen: 41.508 ng

RT: 8.79 min Scan# 986

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion: 77 Resp: 136605

Ion Ratio Lower Upper



77 100


123 41.7 36.3 54.5

65 14.9 10.6 16.0



Abundance Ion 77.00 (76.70 to 77.70): BM022201.D (-969) (-)





#25

Isophorone

Concen: 40.611 ng

RT: 9.30 min Scan# 1074

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion: 82 Resp: 231670

Ion Ratio Lower Upper

82 100

95 9.1 6.1 9.1#




138 17.9 15.0 22.6



Abundance Ion 82.00 (81.70 to 82.70): BM022201.D (-1066) (-)

Ion 95.00 (94.70 to 95.70): BM022201.D (-1066) (-)

Ion 138.00 (137.70 to 138.70): BM022201.D (-1066) (-)



#26

2-Nitrophenol

Concen: 42.277 ng

RT: 9.48 min Scan# 1104

Delta R.T. -0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion: 139 Resp: 58936

Ion Ratio Lower Upper

139 100

109 48.0 25.2 37.8#



65 60.8 37.4 56.0#




Abundance Ion 139.00 (138.70 to 139.70): BM022201.D (-1097) (-)

Ion 109.00 (108.70 to 109.70): BM022201.D (-1097) (-)

Ion 65.00 (64.70 to 65.70): BM022201.D (-1097) (-)









#27
 2,4-Dimethylphenol
 Concen: 40.092 ng
 RT: 9.54 min Scan# 1114
 Delta R.T. -0.01 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampled :


Tot Ion:122 Resp: 84245
 Ion Ratio Lower Upper
 122 100
 107 136.8 95.3 142.9
 121 60.9 47.5 71.3



#28
 bis(2-Chloroethoxy)methane
 Concen: 40.896 ng
 RT: 9.78 min Scan# 1155
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion: 93 Resp: 138018
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.2 39.2
 123 11.5 9.0 13.4









#29
 2,4-Dichlorophenol
 Concen: 40.880 ng
 RT: 10.01 min Scan# 1194
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
162	100		
164	64.8	43.7	83.7
98	36.0	13.4	53.4



#30
 1,2,4-Trichlorobenzene
 Concen: 40.339 ng
 RT: 10.21 min Scan# 1228
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
180	100		
182	98.4	78.0	117.0
145	31.8	23.4	35.0









#31
Naphthalene
Concen: 40.976 ng
RT: 10.40 min Scan# 1260
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampleId :

Tot Ion:128 Resp: 332723
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.4 10.8 16.2




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 36.975 ng
RT: 9.73 min Scan# 1147
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion:122 Resp: 66324
Ion Ratio Lower Upper
122 100
105 127.5 108.9 148.9
77 110.5 76.7 116.7



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BM









#33
4-Chloroaniline
Concen: 39.887 ng
RT: 10.53 min Scan# 1283
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampleId :

Tot Ion:	127	Resp:	135545
Ion	Ratio	Lower	Upper
127	100		
129	33.1	25.1	37.7
65	35.3	23.8	35.8
92	20.5	15.7	23.5




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BM
Ion 92.00 (91.70 to 92.70): BM






#34
Hexachlorobutadiene
Concen: 41.009 ng
RT: 10.65 min Scan# 1303
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion:	225	Resp:	109903
Ion	Ratio	Lower	Upper
225	100		
223	61.9	49.6	74.4
227	63.4	50.4	75.6



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 38.945 ng
 RT: 11.37 min Scan# 1425
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Instrument :
 BNA_M
 ClientSampled :

Tot Ion:113 Resp: 26545
 Ion Ratio Lower Upper
 113 100
 55 156.5 129.8 169.8
 56 126.6 94.3 134.3




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BM
 Ion 56.00 (55.70 to 56.70): BM






#36
 4-Chloro-3-methylphenol
 Concen: 41.781 ng
 RT: 11.66 min Scan# 1475
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:107 Resp: 111982
 Ion Ratio Lower Upper
 107 100
 144 24.1 21.3 31.9
 142 76.7 64.9 97.3



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#37
2-Methylnaphthalene
Concen: 40.512 ng
RT: 12.02 min Scan# 1535
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
142	100		
141	91.2	72.2	108.2
115	48.0	27.9	41.9#






Abundance

Ion 142.00 (141.70 to 142.70): E


Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38
1-Methylnaphthalene
Concen: 41.294 ng
RT: 12.24 min Scan# 1573
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
142	100		
141	93.3	73.9	110.9
116	5.4	3.0	4.6#






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E









#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.23 min Scan# 1911
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Instrument :
 BNA_M
 ClientSampleId :

Tot Ion:164 Resp: 99102
 Ion Ratio Lower Upper
 164 100
 162 105.4 82.1 123.1
 160 47.3 37.1 55.7






Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E




#40
 1,2,4,5-Tetrachlorobenzene
 Concen: 40.754 ng
 RT: 12.39 min Scan# 1599
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:216 Resp: 158091
 Ion Ratio Lower Upper
 216 100
 214 78.1 63.0 94.6
 179 20.7 16.4 24.6
 108 16.9 12.5 18.7

Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 38.189 ng

RT: 12.35 min Scan# 1591

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:237 Resp: 72485

Ion Ratio Lower Upper

237 100

235 61.6 41.7 81.7




272 11.1 0.0 34.1



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 80.825 ng

RT: 15.74 min Scan# 2168

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:330 Resp: 129648

Ion Ratio Lower Upper

330 100

332 94.6 77.1 115.7



141 44.6 21.8 32.8#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 39.556 ng
 RT: 12.65 min Scan# 1643
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampled :


Tot Ion:196 Resp: 91983
 Ion Ratio Lower Upper
 196 100
 198 97.3 76.7 115.1
 200 32.6 24.2 36.4



#44
 2,4,5-Trichlorophenol
 Concen: 40.168 ng
 RT: 12.73 min Scan# 1656
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:196 Resp: 94813
 Ion Ratio Lower Upper
 196 100
 198 96.1 78.2 117.4
 97 54.4 34.1 51.1#
 132 24.4 19.2 28.8









#45
2-Fluorobiphenyl
Concen: 79.457 ng
RT: 12.84 min Scan# 1675
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08




Instrument :
BNA_M
ClientSampled :


Tot Ion:172 Resp: 630453
Ion Ratio Lower Upper
172 100
171 35.8 28.6 43.0
170 24.0 18.7 28.1



#46
1,1'-Biphenyl
Concen: 40.593 ng
RT: 13.05 min Scan# 1711
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion:154 Resp: 316373
Ion Ratio Lower Upper
154 100
153 39.8 21.3 61.3
76 11.2 0.0 30.9









#47
 2-Chloronaphthalene
 Concen: 39.749 ng
 RT: 13.09 min Scan# 1718
 Delta R.T. -0.01 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
162	100		
127	41.5	29.4	44.2
164	33.0	26.5	39.7



#48
 2-Nitroaniline
 Concen: 40.790 ng
 RT: 13.33 min Scan# 1759
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
65	100		
92	64.2	56.3	84.5
138	95.1	92.6	139.0





#49

Acenaphthylene

Concen: 41.010 ng

RT: 13.95 min Scan# 1864

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:152 Resp: 403474

Ion Ratio Lower Upper

152 100

151 20.2 16.1 24.1




153 13.5 10.6 16.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 40.569 ng

RT: 13.70 min Scan# 1822

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:163 Resp: 338019

Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.2



164 10.1 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 42.488 ng
 RT: 13.84 min Scan# 1845
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampleId :


Tot Ion:165 Resp: 69911
 Ion Ratio Lower Upper
 165 100
 63 69.4 37.8 56.6#
 89 71.3 39.7 59.5#



#52
 Acenaphthene
 Concen: 40.775 ng
 RT: 14.29 min Scan# 1922
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:154 Resp: 233946
 Ion Ratio Lower Upper
 154 100
 153 115.4 91.5 137.3
 152 54.7 42.3 63.5









#53
 3-Nitroaniline
 Concen: 41.321 ng
 RT: 14.17 min Scan# 1902
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampleId :


Tot Ion:138 Resp: 67710
 Ion Ratio Lower Upper
 138 100
 108 10.7 9.0 13.4
 92 117.5 92.2 138.2



#54
 2,4-Dinitrophenol
 Concen: 41.340 ng
 RT: 14.40 min Scan# 1940
 Delta R.T. -0.01 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:184 Resp: 36301
 Ion Ratio Lower Upper
 184 100
 63 84.2 45.5 68.3#
 154 86.6 50.6 76.0#





#55

Dibenzofuran

Concen: 40.674 ng

RT: 14.63 min Scan# 1980

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:168 Resp: 387786

Ion Ratio Lower Upper

168 100

139 42.9 28.4 42.6#




169 13.4 10.6 16.0



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 42.867 ng

RT: 14.50 min Scan# 1957

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:139 Resp: 46815

Ion Ratio Lower Upper

139 100

109 127.5 53.3 93.3#



65 118.8 70.3 110.3#




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BM









#57
2,4-Dinitrotoluene
Concen: 41.640 ng
RT: 14.64 min Scan# 1981
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08




Instrument :
BNA_M
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
165	100		
63	74.1	29.6	44.4#
89	94.3	49.5	74.3#
182	2.6	2.4	3.6



#58
Fluorene
Concen: 40.154 ng
RT: 15.29 min Scan# 2091
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.1	78.6	117.8
167	13.4	10.9	16.3






#59
 2,3,4,6-Tetrachlorophenol
 Concen: 40.380 ng
 RT: 14.87 min Scan# 2020
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Instrument :
 BNA_M
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
232	100		
131	45.0	30.6	46.0
130	2.7	1.7	2.5
166	28.2	23.4	35.0






Abundance

Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#60
 Diethylphthalate
 Concen: 40.773 ng
 RT: 15.07 min Scan# 2055
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
149	100		
177	21.5	19.1	28.7
150	12.5	10.1	15.1






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#61
 4-Chlorophenyl-phenylether
 Concen: 39.171 ng
 RT: 15.28 min Scan# 2090
 Delta R.T. -0.01 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Instrument :
 BNA_M
 ClientSampled :

Tot Ion:204 Resp: 198418
 Ion Ratio Lower Upper
 204 100
 206 33.2 26.1 39.1
 141 58.5 40.7 61.1




Abundance Ion 204.00 (203.70 to 204.70): E
 Ion 206.00 (205.70 to 206.70): E
 Ion 141.00 (140.70 to 141.70): E






#62
 4-Nitroaniline
 Concen: 41.861 ng
 RT: 15.35 min Scan# 2102
 Delta R.T. -0.01 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:138 Resp: 63488
 Ion Ratio Lower Upper
 138 100
 92 58.4 29.1 69.1
 108 143.7 59.2 99.2#



Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BM0
 Ion 108.00 (107.70 to 108.70): E






#63
Azobenzene
Concen: 41.218 ng
RT: 15.58 min Scan# 2141
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
77	100		
182	23.1	12.6	52.6
105	16.4	1.1	41.1
51	37.2	6.9	46.9






Abundance Ion 77.00 (76.70 to 77.70): BM022201.D (-2124) (-)

Ion 182.00 (181.70 to 182.70): BM022201.D (-2124) (-)


Ion 105.00 (104.70 to 105.70): BM022201.D (-2124) (-)

Ion 51.00 (50.70 to 51.70): BM022201.D (-2124) (-)



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.99 min Scan# 2381
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08



Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.4	5.8	8.8
80	7.5	6.9	10.3




Abundance Ion 188.00 (187.70 to 188.70): BM022201.D (-2364) (-)

Ion 94.00 (93.70 to 94.70): BM022201.D (-2364) (-)

Ion 80.00 (79.70 to 80.70): BM022201.D (-2364) (-)





#65

4,6-Dinitro-2-methylphenol

Concen: 41.149 ng

RT: 15.41 min Scan# 2112

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampled :


Tot Ion:198 Resp: 52119

Ion Ratio Lower Upper

198 100

51 46.6 15.3 55.3




105 38.4 16.8 56.8



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BM0

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 40.664 ng

RT: 15.51 min Scan# 2129

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:169 Resp: 262531

Ion Ratio Lower Upper

169 100

168 68.9 54.5 81.7



167 37.5 29.0 43.6




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 40.288 ng
 RT: 16.18 min Scan# 2243
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Instrument :
 BNA_M
 ClientSampleId :

Tot Ion:248 Resp: 120205
 Ion Ratio Lower Upper
 248 100
 250 98.2 77.5 116.3
 141 76.0 47.8 71.6#




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#68
 Hexachlorobenzene
 Concen: 40.131 ng
 RT: 16.29 min Scan# 2261
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:284 Resp: 136033
 Ion Ratio Lower Upper
 284 100
 142 34.5 21.2 31.8#
 249 34.0 23.4 35.0



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 32.667 ng

RT: 16.47 min Scan# 2293

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:200 Resp: 69838

Ion Ratio Lower Upper

200 100

173 21.4 4.3 44.3




215 46.0 28.2 68.2



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 40.178 ng

RT: 16.64 min Scan# 2322

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:266 Resp: 66260

Ion Ratio Lower Upper

266 100

268 62.4 51.0 76.4



264 64.3 49.9 74.9




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E





#71

Phenanthrene

Concen: 41.129 ng

RT: 17.03 min Scan# 2388

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:178 Resp: 480231

Ion Ratio Lower Upper

178 100

176 20.4 15.2 22.8




179 15.4 12.2 18.4



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 41.093 ng

RT: 17.13 min Scan# 2404

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:178 Resp: 476298

Ion Ratio Lower Upper

178 100

176 19.3 14.7 22.1



179 15.2 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#73

Carbazole

Concen: 42.283 ng

RT: 17.41 min Scan# 2452

Delta R.T. -0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:167 Resp: 408949

Ion Ratio Lower Upper




167 100

166 22.3 17.4 26.2

139 15.0 9.7 14.5#



Abundance Ion 167.00 (166.70 to 167.70): E
Ion 166.00 (165.70 to 166.70): E
Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 41.646 ng

RT: 17.96 min Scan# 2546

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:149 Resp: 611127

Ion Ratio Lower Upper



149 100


150 9.0 7.4 11.0

104 6.6 3.8 5.8#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E





#75

Fluoranthene

Concen: 43.073 ng

RT: 19.06 min Scan# 2733

Delta R.T. 0.00 min

Lab File: BM022201.D





Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :

Tot Ion:	202	Resp:	606488
Ion	Ratio	Lower	Upper
202	100		
101	6.8	0.0	28.2
203	17.2	0.0	37.6



#76

Chrysene-d12

Concen: 20.000 ng




RT: 21.20 min Scan# 3096


Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Tgt Ion:	240	Resp:	228325
Ion	Ratio	Lower	Upper
240	100		
120	6.8	6.9	10.3#
236	26.4	20.4	30.6





#77

Benzidine

Concen: 36.758 ng

RT: 19.27 min Scan# 2768

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:184 Resp: 188245

Ion Ratio Lower Upper

184 100

185 14.2 11.4 17.2

183 12.2 9.6 14.4






Abundance

Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 38.576 ng

RT: 19.43 min Scan# 2795

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:202 Resp: 619966

Ion Ratio Lower Upper

202 100

200 21.4 17.0 25.4

203 17.3 14.0 21.0






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 76.362 ng

RT: 19.63 min Scan# 2830

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:244 Resp: 1196597

Ion Ratio Lower Upper

244 100

212 8.2 5.3 7.9#




122 6.3 6.6 9.8#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 39.060 ng

RT: 20.33 min Scan# 2949

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:149 Resp: 280775

Ion Ratio Lower Upper

149 100

91 76.8 50.9 76.3#



206 21.2 19.7 29.5




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BM0

Ion 206.00 (205.70 to 206.70): E









#81
Benzo(a)anthracene
Concen: 40.742 ng
RT: 21.18 min Scan# 3093
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08




Instrument :
BNA_M
ClientSampleId :


Tot	Ion:228	Resp:	654206
Ion	Ratio	Lower	Upper
228	100		
226	28.1	21.6	32.4
229	19.6	15.5	23.3



#82
3,3'-Dichlorobenzidine
Concen: 40.591 ng
RT: 21.13 min Scan# 3084
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion	252	Resp: Lower	227250 Upper
	Ratio		
252	100		
254	64.1	51.7	77.5
126	7.9	6.6	9.8









#83
 Chrysene
 Concen: 40.231 ng
 RT: 21.23 min Scan# 3102
 Delta R.T. -0.01 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08




Instrument :
 BNA_M
 ClientSampleId :


Tot Ion:228	Resp:	623606
Ion	Ratio	Lower Upper
228	100	
226	31.0	24.1 36.1
229	20.1	15.6 23.4



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 38.701 ng
 RT: 21.10 min Scan# 3080
 Delta R.T. 0.00 min
 Lab File: BM022201.D
 Acq: 19 Aug 2019 22:08

Tgt Ion:149	Resp:	418382
Ion	Ratio	Lower Upper
149	100	
167	24.8	23.0 34.4
279	5.0	4.6 7.0





#85

Di-n-octyl phthalate

Concen: 39.932 ng

RT: 21.97 min Scan# 3228

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampleId :


Tot Ion:149 Resp: 698720

Ion Ratio Lower Upper

149 100

167 1.5 1.3 1.9




43 10.3 6.2 9.2#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BM



#86

Indeno(1,2,3-cd)pyrene

Concen: 42.448 ng

RT: 25.60 min Scan# 3844

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:276 Resp: 737026

Ion Ratio Lower Upper

276 100

138 15.9 15.4 23.0



277 25.4 20.4 30.6




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Pervlene-d12
Concen: 20.000 ng
RT: 23.39 min Scan# 3469
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08




Instrument :
BNA_M
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.0	28.4
265	21.8	17.7	26.5



#88
Benzo(b)fluoranthene
Concen: 39.616 ng
RT: 22.74 min Scan# 3358
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.0
125	6.4	5.9	8.9





#89

Benzo(k)fluoranthene

Concen: 40.210 ng

RT: 22.78 min Scan# 3365

Delta R.T. -0.01 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08

Instrument :

BNA_M

ClientSampled :


Tot Ion:252 Resp: 610249

Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3




125 6.2 5.9 8.9



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Benzo(a)pyrene

Concen: 40.070 ng

RT: 23.30 min Scan# 3454

Delta R.T. 0.00 min

Lab File: BM022201.D

Acq: 19 Aug 2019 22:08


Tgt Ion:252 Resp: 566560

Ion Ratio Lower Upper

252 100

253 21.4 17.5 26.3



125 6.5 6.6 10.0#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#91
Dibenzo(a,h)anthracene
Concen: 41.068 ng
RT: 25.60 min Scan# 3845
Delta R.T. 0.00 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Instrument :
BNA_M
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
278	100		
139	9.3	9.8	14.8#
279	23.1	18.7	28.1




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92
Benzo(g,h,i)perylene
Concen: 41.493 ng
RT: 26.27 min Scan# 3958
Delta R.T. -0.01 min
Lab File: BM022201.D
Acq: 19 Aug 2019 22:08

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.0	18.8	28.2
138	13.7	13.1	19.7



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

