

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM082019\  
 Data File : BM022208.D  
 Acq On : 20 Aug 2019 12:32  
 Operator : HP/JU  
 Sample : K4330-02  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 0631-SITE-WATER-MUNICIPAL

Quant Time: Aug 20 23:44:43 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Aug 20 02:18:05 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)     |
|-----------------------------|-------|------|----------|--------|-------|--------------|
| 1) 1,4-Dichlorobenzene-d4   | 7.57  | 152  | 35278    | 20.00  | ng    | 0.00         |
| 21) Naphthalene-d8          | 10.35 | 136  | 134953   | 20.00  | ng    | 0.00         |
| 39) Acenaphthene-d10        | 14.23 | 164  | 90909    | 20.00  | ng    | 0.00         |
| 64) Phenanthrene-d10        | 16.99 | 188  | 195633   | 20.00  | ng    | 0.00         |
| 76) Chrysene-d12            | 21.20 | 240  | 178816   | 20.00  | ng    | 0.00         |
| 87) Perylene-d12            | 23.40 | 264  | 184194   | 20.00  | ng    | 0.00         |
| System Monitoring Compounds |       |      |          |        |       |              |
| 5) 2-Fluorophenol           | 5.19  | 112  | 136714   | 69.06  | ng    | 0.00         |
| 7) Phenol-d6                | 6.76  | 99   | 133506   | 50.64  | ng    | 0.00         |
| 23) Nitrobenzene-d5         | 8.74  | 82   | 279565   | 100.76 | ng    | 0.00         |
| 42) 2,4,6-Tribromophenol    | 15.73 | 330  | 237132   | 161.16 | ng    | 0.00         |
| 45) 2-Fluorobiphenyl        | 12.83 | 172  | 735782   | 101.09 | ng    | 0.00         |
| 79) Terphenyl-d14           | 19.63 | 244  | 1499559  | 122.19 | ng    | 0.00         |
| Target Compounds            |       |      |          |        |       |              |
| 50) Dimethylphthalate       | 13.70 | 163  | 60669    | 7.938  | ng    | Qvalue<br>99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM082019\  
Data File : BM022208.D  
Acq On : 20 Aug 2019 12:32  
Operator : HP/JU  
Sample : K4330-02  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
0631-SITE-WATER-MUNICIPAL

Quant Time: Aug 20 23:44:43 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Aug 20 02:18:05 2019  
Response via : Initial Calibration

