


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $5.28 \times 10^{-1} \times \text{Amt} - 2.76 \times 10^{-1}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM082019.M
Calibration Table Last Updated: Tue Aug 20 02:18:05 2019

2,4-Dinitrophenol

Response Ratio




Resp Ratio = $2.02 \times 10^{-1} \times \text{Amt} - 5.10 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM082019.M
Calibration Table Last Updated: Tue Aug 20 02:18:05 2019

4,6-Dinitro-2-methylphenol

Response Ratio



Resp Ratio = $1.47 \times 10^{-1} \times \text{Amt} - 4.37 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM082019.M
Calibration Table Last Updated: Tue Aug 20 02:18:05 2019