

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM082120\
 Data File : BM027182.D
 Acq On : 21 Aug 2020 22:41
 Operator : JU/CG
 Sample : L3729-17
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COMP-4-(3-4)

Manual Integrations
 APPROVED

Sohil
 8/24/2020 1:48:40 PM

Quant Time: Sep 01 11:32:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 21 15:18:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	121494	20.00	ng	0.00
21) Naphthalene-d8	10.38	136	457506	20.00	ng	0.00
39) Acenaphthene-d10	14.25	164	295197	20.00	ng	0.00
64) Phenanthrene-d10	16.99	188	636893	20.00	ng	0.00
76) Chrysene-d12	21.19	240	665037	20.00	ng	0.00
86) Perylene-d12	23.38	264	729592	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	533452	85.60	ng	0.00
7) Phenol-d6	6.80	99	733388	85.71	ng	0.00
23) Nitrobenzene-d5	8.76	82	551818	49.76	ng	0.00
42) 2,4,6-Tribromophenol	15.74	330	341560	71.45	ng	0.00
45) 2-Fluorobiphenyl	12.87	172	1139107	50.69	ng	0.00
79) Terphenyl-d14	19.64	244	1753414	49.25	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.71	163	133156	5.556	ng	99
75) Fluoranthene	19.06	202	168477	3.809	ng	97
78) Pyrene	19.42	202	223750	5.529	ng	99
81) Benzo(a)anthracene	21.18	228	92894m	2.156	ng	
83) Chrysene	21.23	228	93067	2.224	ng	99
88) Benzo(b)fluoranthene	22.73	252	164141m	3.374	ng	
90) Benzo(a)pyrene	23.29	252	134006	3.017	ng	# 98
92) Benzo(g,h,i)perylene	26.23	276	181282	3.959	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM082120\
Data File : BM027182.D
Acq On : 21 Aug 2020 22:41
Operator : JU/CG
Sample : L3729-17
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COMP-4-(3-4)

Manual Integrations
APPROVED

Sohil
8/24/2020 1:48:40 PM

Quant Time: Sep 01 11:32:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 21 15:18:25 2020
Response via : Initial Calibration

