


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM082619\
 Data File : BM022323.D
 Acq On : 26 Aug 2019 10:26
 Operator : HP/JU
 Sample : DFTPP08
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP08

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM082219MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Aug 27 17:04:34 2019



AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2443

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.1	33498	PASS
68	69	0.00	2	1.9	602	PASS
69	198	0.00	100	37.3	31120	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	42.4	35413	PASS
197	198	0.00	2	0.8	671	PASS
198	198	100	100	100.0	83541	PASS
199	198	5	9	7.3	6082	PASS
275	198	10	60	34.2	28592	PASS
365	198	1	100	8.4	6999	PASS
441	198	0.01	100	11.6	9690	PASS
442	198	50	100	71.0	59288	PASS
443	442	15	24	18.9	11187	PASS

m/z	Abundance
38.00	517.0
39.00	2173.0
50.10	10055.0
51.00	31976.0
52.00	1445.0
55.90	899.0
57.00	1740.0
60.90	461.0
62.00	435.0
63.00	1274.0
65.10	352.0
68.00	524.0
69.00	27664.0
73.00	560.0
74.00	4294.0
75.00	5595.0
76.00	2528.0
77.00	36712.0
78.10	2440.0
79.00	3243.0
80.00	1890.0
81.00	3206.0
81.90	896.0
82.90	446.0
85.00	385.0
86.00	996.0
87.00	379.0
91.10	580.0
92.00	657.0
93.00	6071.0
94.10	371.0
98.00	5310.0
99.00	2702.0
100.00	311.0
101.00	1438.0
103.00	602.0
103.90	1206.0
105.00	932.0
107.00	11855.0
108.00	1707.0
110.00	16848.0
111.00	2998.0
112.00	312.0
116.00	591.0
117.00	20584.0
118.00	1477.0
121.90	1003.0
122.90	1357.0
123.90	766.0
125.00	529.0
127.00	29904.0
128.00	2542.0
129.00	15557.0
130.00	1345.0
133.90	425.0
135.00	1565.0
136.00	633.0
137.00	807.0
139.90	342.0
141.00	2375.0
142.00	677.0
142.90	512.0
145.90	537.0
147.00	1144.0
148.00	4494.0
149.00	940.0
153.00	797.0
153.90	456.0
155.00	1413.0
156.00	2031.0
157.90	555.0
158.90	356.0
159.90	1004.0
161.00	1068.0
161.90	338.0
165.00	1251.0
165.90	938.0
167.00	6969.0
168.00	4097.0
169.00	334.0
171.80	551.0
173.00	816.0
174.00	898.0
175.00	1972.0
175.90	516.0
177.00	1045.0
178.00	398.0
178.90	4200.0
180.00	2864.0
180.90	1040.0
183.90	424.0
184.90	1856.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

186.00	14129.0
187.00	4058.0
188.10	497.0
188.90	1051.0
189.90	306.0
190.90	708.0
192.00	1514.0
193.00	1759.0
194.00	364.0
195.90	1997.0
197.00	537.0
197.90	66320.0
199.00	4085.0
199.90	395.0
203.00	868.0
204.00	4168.0
205.00	6523.0
206.00	23648.0
207.00	4209.0
208.00	1187.0
209.00	525.0
210.10	485.0
211.00	1250.0
214.90	400.0
216.00	698.0
216.90	8382.0
218.00	906.0
220.90	1900.0
222.00	990.0
223.00	1730.0
224.00	15181.0
225.00	3851.0
226.00	397.0
227.00	8155.0
227.90	1064.0
229.00	1662.0
231.10	410.0
234.00	506.0
234.90	475.0
235.90	323.0
237.00	723.0
238.80	311.0
241.00	435.0
242.00	822.0
243.00	956.0
244.00	10197.0
245.00	1430.0
246.00	2964.0
246.90	703.0
248.80	367.0
254.10	381.0
255.00	55456.0
256.00	8583.0
256.90	761.0
257.90	4937.0
258.90	826.0
265.00	1870.0
273.00	1654.0
274.00	4225.0
275.00	21248.0
276.00	2768.0
277.00	2707.0
277.90	589.0
282.90	404.0
285.00	388.0
292.90	650.0
296.00	9926.0
296.90	1449.0
303.00	812.0
313.90	398.0
314.90	1182.0
315.90	504.0
323.00	1922.0
323.90	363.0
326.90	602.0
333.90	1488.0
335.00	337.0
345.90	501.0
352.00	674.0
353.10	402.0
354.00	584.0
365.00	4420.0
366.00	681.0
371.90	921.0
403.00	532.0
421.00	439.0
421.90	477.0
423.00	2039.0
423.90	584.0
441.00	5362.0
442.00	32456.0
443.00	6299.0
444.00	702.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
207.00	709.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	460.0
39.00	1900.0
49.00	336.0
50.00	8871.0
51.00	27384.0
52.00	1430.0
56.00	947.0
57.00	1662.0
60.90	351.0
61.90	550.0
63.00	1284.0
65.00	431.0
68.00	485.0
69.00	26384.0
73.00	663.0
74.00	3822.0
75.00	5114.0
76.00	1942.0
77.00	35232.0
78.00	2733.0
79.00	2516.0
80.00	1914.0
80.90	2317.0
81.90	706.0
83.00	389.0
84.90	412.0
86.00	961.0
86.90	392.0
91.00	475.0
92.00	826.0
93.00	6350.0
94.00	315.0
98.00	5309.0
99.00	2800.0
100.90	1428.0
103.00	604.0
104.00	978.0
105.00	1162.0
106.00	378.0
107.00	11489.0
108.00	1735.0
109.00	333.0
110.00	16038.0
110.90	2254.0
111.90	512.0
116.00	792.0
117.00	19680.0
118.00	1271.0
122.00	874.0
123.00	1441.0
123.90	676.0
124.90	566.0
127.00	31864.0
128.00	2401.0
129.00	15731.0
130.00	1341.0
131.00	307.0
133.90	531.0
135.00	1570.0
135.90	767.0
137.00	783.0
141.00	2661.0
141.90	751.0
142.90	518.0
146.00	617.0
146.90	1430.0
148.00	4960.0
149.00	842.0
151.00	375.0
152.90	786.0
154.00	570.0
155.00	1597.0
156.00	1815.0
156.90	440.0
157.90	575.0
159.00	446.0
160.00	1006.0
161.00	1179.0
164.90	1215.0
165.90	702.0
167.00	8216.0
168.00	4014.0
169.00	549.0
171.90	580.0
173.00	775.0
173.90	1240.0
175.00	2296.0
176.00	526.0
177.00	1251.0
177.90	460.0
179.00	4256.0
180.00	3028.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

180.90	1370.0
181.90	309.0
183.80	407.0
185.00	2367.0
186.00	15604.0
187.00	4604.0
187.90	507.0
188.90	1514.0
190.90	557.0
192.00	1882.0
193.00	2227.0
194.10	388.0
196.00	2258.0
196.80	608.0
197.90	79584.0
198.90	5939.0
199.90	551.0
203.00	1238.0
204.00	5455.0
205.00	8474.0
206.00	30960.0
207.00	4943.0
208.00	1481.0
208.90	570.0
210.00	734.0
211.00	1433.0
214.90	646.0
215.90	937.0
216.90	10521.0
218.00	1463.0
220.90	2670.0
221.90	1414.0
223.00	2585.0
224.00	19256.0
225.00	4639.0
226.00	618.0
227.00	10949.0
228.00	1513.0
228.90	1713.0
231.00	557.0
233.90	685.0
235.00	698.0
235.90	593.0
237.00	787.0
239.00	385.0
240.00	380.0
241.00	560.0
242.00	1142.0
243.00	1344.0
244.00	14502.0
245.00	1787.0
246.00	4021.0
247.00	670.0
249.00	460.0
252.90	341.0
254.10	556.0
255.00	79232.0
256.00	11127.0
257.00	866.0
258.00	6976.0
258.90	810.0
265.00	2903.0
266.00	314.0
273.00	2623.0
274.00	6161.0
275.00	29224.0
276.00	4000.0
277.00	4536.0
277.90	706.0
281.00	362.0
283.00	600.0
283.90	393.0
285.00	548.0
292.90	828.0
296.00	16173.0
297.00	2353.0
303.00	1461.0
303.90	356.0
313.90	685.0
314.90	1754.0
315.90	704.0
321.00	529.0
323.00	3252.0
324.00	688.0
326.90	934.0
328.10	343.0
331.90	302.0
333.00	411.0
334.00	2605.0
335.00	630.0
340.90	537.0
345.90	996.0
352.00	1256.0
353.00	630.0
354.00	1144.0
365.00	7901.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

365.90	1127.0
372.00	1658.0
372.90	347.0
382.90	431.0
402.00	682.0
403.00	943.0
421.00	710.0
422.00	700.0
423.00	4435.0
424.00	952.0
441.00	12099.0
442.00	73376.0
443.00	13671.0
444.10	1335.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	627.0
39.10	2952.0
48.90	428.0
50.00	12557.0
51.10	41136.0
52.00	2069.0
56.00	1030.0
57.00	2320.0
61.00	446.0
62.00	603.0
63.00	1682.0
65.00	749.0
68.00	798.0
69.00	39312.0
73.00	758.0
74.10	6427.0
75.00	7573.0
76.00	3073.0
77.00	49856.0
78.00	3523.0
79.00	4012.0
80.00	2750.0
81.00	3852.0
82.00	917.0
83.00	627.0
84.90	469.0
86.00	1576.0
87.00	475.0
91.00	727.0
92.00	1276.0
93.00	8463.0
94.00	513.0
95.90	378.0
98.00	7412.0
99.00	4264.0
99.90	365.0
100.90	1990.0
102.90	670.0
104.00	1391.0
105.00	1542.0
106.10	402.0
107.00	18024.0
108.00	2518.0
109.00	541.0
110.00	21592.0
111.00	3702.0
111.90	544.0
116.10	1062.0
117.00	27872.0
118.00	2128.0
122.00	1475.0
122.90	1867.0
123.90	960.0
125.00	684.0
127.00	44472.0
128.00	3782.0
129.00	20872.0
130.00	1799.0
131.00	460.0
133.90	593.0
134.90	1838.0
136.00	883.0
137.00	1169.0
139.90	342.0
141.00	3903.0
142.00	1145.0
143.10	778.0
146.00	878.0
147.00	2234.0
148.00	7358.0
149.00	1190.0
151.00	506.0
152.90	1099.0
153.90	898.0
155.00	2112.0
156.00	2893.0
157.00	523.0
158.00	490.0
159.00	464.0
159.90	1190.0
160.90	1860.0
161.90	474.0
165.00	1474.0
165.90	1235.0
167.00	10700.0
168.00	6034.0
169.00	793.0
169.90	322.0
171.00	376.0
171.90	778.0
172.90	1057.0
174.00	1748.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

175.00	3471.0
176.00	660.0
177.00	1376.0
177.90	598.0
178.90	7297.0
180.00	4724.0
181.00	1840.0
184.00	490.0
185.00	3248.0
186.00	20944.0
187.00	6966.0
188.10	651.0
189.00	1577.0
189.90	419.0
190.90	933.0
192.00	2278.0
193.00	2727.0
194.00	360.0
196.00	3048.0
196.90	868.0
197.90	104720.0
198.90	8224.0
199.90	843.0
201.40	437.0
202.90	1252.0
204.00	7015.0
205.00	11179.0
206.00	40128.0
207.00	6199.0
208.00	1706.0
209.10	614.0
209.90	845.0
211.00	1712.0
214.90	643.0
216.00	1168.0
216.90	14903.0
217.90	1801.0
221.00	3743.0
221.90	1485.0
223.00	2833.0
224.00	24040.0
225.00	6283.0
226.00	680.0
227.00	13366.0
228.00	2069.0
229.00	2425.0
230.00	376.0
231.00	879.0
233.90	785.0
235.00	889.0
236.00	694.0
236.90	957.0
239.00	517.0
239.90	493.0
240.90	771.0
242.00	1652.0
243.00	1325.0
244.00	18288.0
245.00	2323.0
246.00	4970.0
246.90	796.0
248.90	485.0
252.90	491.0
254.00	646.0
255.00	95920.0
256.00	14722.0
257.00	1196.0
258.00	7714.0
259.00	1282.0
265.00	3640.0
266.00	460.0
272.90	3063.0
274.00	7325.0
275.00	35304.0
276.00	4938.0
277.00	4828.0
277.90	809.0
282.90	517.0
285.00	692.0
292.90	1005.0
295.10	403.0
296.00	17688.0
297.00	2541.0
303.00	1921.0
303.90	420.0
314.00	638.0
315.00	1888.0
315.90	952.0
321.00	615.0
322.00	311.0
323.00	3595.0
324.00	923.0
326.90	1069.0
328.00	393.0
332.90	305.0
334.00	2895.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08

335.00	777.0
340.90	567.0
346.00	1032.0
352.00	1005.0
353.00	681.0
354.00	955.0
365.00	8676.0
365.90	1220.0
372.00	1598.0
372.90	408.0
383.00	411.0
402.00	528.0
403.00	799.0
403.90	328.0
421.00	701.0
421.90	570.0
423.00	4503.0
424.00	932.0
441.10	11609.0
442.00	72032.0
443.10	13592.0
444.00	1370.0

Instrument :
BNA_M
ClientSampleId :
DFTPP08