


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM082720\
 Data File : BM027267.D
 Acq On : 27 Aug 2020 09:12
 Operator : JU/CG
 Sample : DFTPP69
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP69

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM082520MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Aug 27 16:42:10 2020



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.2	67592	PASS
68	69	0.00	2	1.8	1987	PASS
69	198	0.00	100	47.1	112979	PASS
70	69	0.00	2	0.6	699	PASS
127	198	10	80	43.9	105432	PASS
197	198	0.00	2	0.4	983	PASS
198	198	100	100	100.0	239893	PASS
199	198	5	9	6.6	15863	PASS
275	198	10	60	30.4	72946	PASS
365	198	1	100	4.9	11742	PASS
441	198	0.01	100	12.6	30202	PASS
442	198	50	100	83.9	201194	PASS
443	442	15	24	19.4	38938	PASS

Scan 2453 (17.415 min): BM027267.D
DFTPP69

m/z	Abundance
40.00	444.0
44.00	1041.0
68.80	170.0
73.10	358.0
77.00	210.0
91.00	284.0
132.90	223.0
190.80	242.0
193.00	281.0
206.90	1653.0
207.90	437.0
208.90	159.0
281.00	489.0
341.00	171.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

m/z	Abundance
37.00	460.0
38.10	1010.0
39.10	5482.0
40.10	692.0
41.00	190.0
44.00	1015.0
44.90	232.0
49.10	623.0
50.10	19616.0
51.10	69424.0
52.10	4282.0
52.90	267.0
55.10	620.0
56.00	2448.0
57.00	5326.0
57.90	391.0
60.00	159.0
61.00	1308.0
62.00	1416.0
63.00	4243.0
64.00	571.0
65.10	2050.0
67.00	235.0
68.00	2009.0
69.00	112072.0
70.00	693.0
73.10	1822.0
74.00	11861.0
75.00	17968.0
76.10	5977.0
77.10	113056.0
78.10	6899.0
79.00	8620.0
80.00	6287.0
81.00	8624.0
82.00	2102.0
83.00	2037.0
83.90	358.0
85.00	1618.0
86.00	3143.0
86.90	941.0
87.90	446.0
88.90	229.0
91.00	1709.0
92.00	2813.0
93.00	15637.0
93.90	945.0
96.00	642.0
97.10	294.0
98.00	11107.0
99.00	9225.0
100.00	734.0
101.00	4000.0
102.00	209.0
103.00	1225.0
104.00	2752.0
105.00	2774.0
106.00	1126.0
107.00	34016.0
108.00	4843.0
109.00	944.0
110.00	53192.0
111.00	8624.0
112.00	1260.0
112.90	320.0
115.00	315.0
116.00	2001.0
117.00	33912.0
118.00	2566.0
118.90	264.0
120.00	351.0
120.90	193.0
122.00	1924.0
123.00	3408.0
124.00	1606.0
125.10	1504.0
126.10	587.0
127.00	96872.0
128.00	8114.0
129.00	43224.0
130.00	2861.0
131.00	762.0
131.80	315.0
133.00	170.0
133.90	1392.0
135.00	3409.0
136.00	1397.0
137.00	1976.0
137.80	309.0
139.00	322.0
140.00	766.0
141.00	6036.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

142.00	1827.0
143.00	1255.0
144.00	323.0
144.90	211.0
146.10	1368.0
147.00	3498.0
148.00	7614.0
149.00	2424.0
149.90	272.0
150.90	778.0
151.70	505.0
152.90	1662.0
154.00	1282.0
155.00	3040.0
156.00	4595.0
157.10	1212.0
158.00	1288.0
158.90	1136.0
160.00	1939.0
161.00	2874.0
161.90	814.0
163.00	255.0
164.00	529.0
165.00	2428.0
166.00	1770.0
167.00	12325.0
168.00	8573.0
169.00	1650.0
170.00	414.0
171.00	573.0
171.90	1150.0
173.00	1666.0
174.00	2476.0
175.00	4685.0
175.90	1281.0
176.90	2686.0
177.90	903.0
179.00	9753.0
180.00	7249.0
181.00	2935.0
182.10	488.0
182.90	152.0
183.90	778.0
185.00	4910.0
186.00	34032.0
187.00	9786.0
187.90	1089.0
189.00	2468.0
189.90	400.0
191.00	1217.0
192.00	3346.0
193.10	3707.0
194.00	786.0
195.00	236.0
196.00	4950.0
196.70	1152.0
197.10	976.0
198.00	208320.0
199.00	14275.0
200.00	1602.0
201.40	794.0
202.20	349.0
202.90	1857.0
204.00	9261.0
205.00	15457.0
206.00	60344.0
207.00	10223.0
208.00	2767.0
209.00	844.0
209.90	1186.0
211.00	2487.0
211.80	570.0
213.00	266.0
215.00	712.0
216.00	2077.0
217.00	17400.0
217.90	2538.0
219.00	182.0
221.00	6920.0
221.80	2117.0
223.00	3815.0
224.00	34952.0
225.00	8185.0
226.00	1161.0
227.00	16920.0
228.00	1797.0
229.00	3545.0
230.00	347.0
231.00	1042.0
231.90	204.0
232.90	231.0
234.00	991.0
234.90	1140.0
235.90	867.0
237.00	1976.0
238.00	180.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

239.00	601.0
239.90	825.0
241.00	1112.0
242.10	2000.0
243.00	1815.0
244.00	24816.0
245.00	3267.0
246.00	5757.0
247.00	1234.0
247.80	276.0
249.00	988.0
249.90	264.0
250.90	184.0
252.00	324.0
253.00	813.0
254.10	941.0
255.00	136512.0
256.00	21640.0
257.00	1468.0
258.00	9149.0
259.00	1408.0
260.00	308.0
261.00	187.0
263.90	226.0
265.00	4125.0
266.10	730.0
269.90	236.0
270.90	421.0
272.00	513.0
273.10	4127.0
274.00	10094.0
275.00	60368.0
276.00	8412.0
277.00	5685.0
277.90	1221.0
278.90	237.0
281.00	547.0
281.90	257.0
283.00	685.0
283.90	728.0
285.00	1191.0
288.80	168.0
292.10	232.0
292.90	993.0
294.00	432.0
295.00	441.0
296.00	18768.0
297.00	2927.0
298.00	277.0
301.00	275.0
302.00	291.0
303.00	1768.0
304.00	519.0
305.10	156.0
308.00	320.0
308.90	257.0
310.00	288.0
313.00	162.0
314.00	771.0
315.00	2403.0
316.00	1320.0
317.00	206.0
321.00	490.0
321.90	243.0
323.00	5640.0
324.00	1033.0
327.00	1027.0
327.90	540.0
331.90	551.0
332.90	529.0
334.00	3329.0
335.00	1151.0
336.10	259.0
341.10	782.0
341.90	227.0
345.90	1307.0
346.90	254.0
352.00	1363.0
353.10	1209.0
354.10	1880.0
355.00	495.0
365.00	8158.0
366.00	1352.0
371.00	605.0
372.10	2793.0
373.10	791.0
376.90	171.0
383.00	679.0
384.20	346.0
390.10	166.0
391.10	243.0
400.80	276.0
401.90	854.0
403.00	1478.0
404.10	661.0
421.00	924.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

422.00	1259.0
423.10	7560.0
424.10	2858.0
425.10	677.0
441.10	18456.0
442.10	131392.0
443.10	26312.0
444.00	2748.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

m/z	Abundance
36.90	329.0
38.00	1117.0
39.10	5839.0
40.00	640.0
41.00	212.0
44.00	764.0
49.00	733.0
50.00	22768.0
51.00	82320.0
52.00	4054.0
53.00	226.0
55.00	804.0
56.00	2858.0
57.00	6532.0
57.90	299.0
61.00	1502.0
62.00	1439.0
63.00	4665.0
64.00	668.0
65.00	1878.0
65.90	199.0
67.00	179.0
68.00	2273.0
69.00	139456.0
70.10	858.0
73.00	1915.0
74.00	14238.0
75.00	21504.0
76.10	7166.0
77.00	135232.0
78.10	8820.0
79.00	10787.0
80.00	8338.0
81.00	11889.0
82.00	2815.0
83.00	2163.0
84.00	204.0
85.00	1617.0
86.00	4224.0
86.90	1407.0
88.10	521.0
89.00	405.0
91.00	2652.0
92.00	3154.0
93.00	18896.0
93.90	1353.0
94.90	237.0
95.20	227.0
96.00	749.0
96.80	324.0
97.20	395.0
98.00	14597.0
99.00	10042.0
100.00	1055.0
101.00	5624.0
102.10	373.0
103.00	1639.0
104.00	3932.0
105.00	4106.0
106.20	1015.0
107.00	42600.0
108.00	5828.0
109.10	1215.0
110.00	68360.0
111.00	11908.0
112.00	1597.0
112.90	320.0
115.10	394.0
115.90	2722.0
117.00	43696.0
118.00	2998.0
119.10	371.0
120.10	546.0
122.10	3020.0
123.00	4153.0
124.00	1759.0
125.00	1717.0
127.00	130528.0
128.00	9640.0
129.00	54856.0
130.00	4927.0
131.00	1039.0
132.00	559.0
133.00	294.0
134.00	1755.0
135.00	4365.0
136.10	1871.0
137.00	2972.0
137.80	541.0
139.00	314.0
139.90	880.0
141.00	7613.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

142.00	2349.0
143.00	2091.0
144.10	453.0
145.00	574.0
146.00	1443.0
147.00	4714.0
148.00	10197.0
149.00	2294.0
150.10	563.0
151.00	1108.0
151.90	433.0
153.00	2694.0
154.00	1904.0
155.00	4496.0
156.00	6723.0
157.00	1019.0
158.00	1708.0
159.00	1344.0
160.00	2311.0
161.00	3324.0
162.00	1094.0
163.00	464.0
163.90	375.0
165.00	3467.0
166.00	2277.0
167.00	16832.0
168.00	10783.0
169.00	1509.0
170.00	665.0
170.90	769.0
172.00	1735.0
173.00	2234.0
174.00	3272.0
175.10	6421.0
176.00	1532.0
177.00	4364.0
178.00	1516.0
179.00	12535.0
180.00	9086.0
181.00	3405.0
182.10	652.0
182.80	413.0
184.00	1154.0
185.00	5179.0
186.00	46056.0
187.00	12695.0
188.00	1402.0
189.00	3341.0
190.00	572.0
191.00	1743.0
192.00	4685.0
193.00	4908.0
194.00	1168.0
194.90	555.0
196.00	7250.0
196.80	1798.0
198.00	292032.0
199.00	18656.0
199.90	1518.0
201.60	1304.0
203.00	3115.0
204.00	13358.0
205.00	21176.0
206.00	84192.0
207.00	12795.0
208.00	3757.0
208.90	1148.0
210.00	1323.0
211.00	3320.0
211.90	585.0
212.90	312.0
215.00	1190.0
216.10	2013.0
217.00	25296.0
218.00	3011.0
219.10	276.0
221.00	9648.0
221.90	3236.0
223.00	4545.0
224.00	49728.0
225.00	12804.0
226.00	1839.0
227.00	24576.0
228.00	3863.0
229.00	4603.0
229.90	550.0
231.00	1689.0
232.00	400.0
233.00	435.0
234.00	1468.0
235.00	1726.0
236.10	971.0
237.00	1998.0
238.00	487.0
239.00	870.0
240.00	836.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

240.90	1143.0
242.10	2266.0
243.00	2713.0
244.00	35600.0
245.10	5035.0
246.00	9158.0
247.00	1419.0
248.00	263.0
249.00	1333.0
249.90	380.0
251.00	403.0
252.00	321.0
253.00	903.0
254.10	1235.0
255.00	203392.0
256.00	30808.0
257.00	2618.0
258.00	12791.0
259.00	2198.0
260.00	342.0
260.90	246.0
263.00	165.0
264.00	494.0
265.00	5522.0
265.90	575.0
267.00	255.0
269.90	195.0
271.00	347.0
272.10	603.0
273.00	6044.0
274.00	15449.0
275.00	86592.0
276.00	11526.0
277.00	8948.0
278.00	1866.0
279.00	476.0
281.10	557.0
282.20	226.0
283.00	1061.0
284.10	606.0
285.00	1484.0
286.10	459.0
288.90	331.0
289.80	316.0
291.00	325.0
292.00	325.0
293.00	1883.0
294.00	409.0
295.00	559.0
296.00	28960.0
297.00	4465.0
298.20	194.0
301.00	384.0
302.10	515.0
303.00	3304.0
304.00	750.0
305.00	150.0
307.90	451.0
308.90	271.0
309.90	322.0
313.00	236.0
314.00	1111.0
315.00	3368.0
316.00	2065.0
316.90	221.0
321.00	821.0
322.00	471.0
323.00	8051.0
324.00	1564.0
324.90	253.0
327.00	2068.0
327.90	702.0
332.00	574.0
333.00	995.0
334.10	5161.0
335.10	1450.0
336.10	157.0
341.00	943.0
342.00	221.0
346.00	2085.0
347.00	422.0
350.90	187.0
352.00	2658.0
353.00	1799.0
354.00	2940.0
355.00	792.0
365.00	14341.0
366.00	2359.0
366.90	253.0
369.80	199.0
371.00	578.0
372.10	4837.0
373.00	1275.0
377.00	165.0
383.10	1035.0
384.00	425.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

390.00	609.0
391.00	454.0
392.00	289.0
401.10	351.0
401.90	1947.0
403.10	2372.0
403.90	917.0
421.10	1936.0
422.10	2283.0
423.00	13774.0
424.10	3437.0
425.00	429.0
440.20	157.0
441.10	36760.0
442.10	232384.0
443.10	46064.0
444.10	4147.0
445.00	436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

m/z	Abundance
36.80	171.0
38.10	727.0
39.00	4180.0
39.90	671.0
41.00	230.0
44.00	1244.0
44.90	203.0
49.10	734.0
50.00	15080.0
51.10	51032.0
52.10	2730.0
54.90	429.0
56.00	2209.0
57.00	4298.0
57.90	255.0
60.90	1166.0
62.00	1039.0
63.00	3307.0
64.00	477.0
65.00	1596.0
67.20	341.0
68.10	1679.0
69.00	87920.0
69.90	546.0
73.10	1316.0
74.00	8980.0
75.00	14582.0
76.00	4707.0
77.10	88864.0
78.00	6194.0
79.00	6319.0
80.00	5265.0
81.00	7231.0
81.90	1436.0
83.00	1699.0
84.00	529.0
85.00	1392.0
86.00	2214.0
87.00	1044.0
87.90	432.0
88.90	191.0
91.10	1498.0
92.00	1855.0
93.00	13178.0
94.10	704.0
95.90	760.0
96.90	172.0
98.00	10158.0
99.00	7014.0
100.10	624.0
101.00	3352.0
101.90	332.0
102.90	1286.0
104.00	2487.0
104.90	2130.0
106.10	926.0
107.00	28304.0
108.00	4114.0
108.90	638.0
110.00	46088.0
111.00	7464.0
112.00	910.0
113.00	341.0
115.00	167.0
116.10	1813.0
117.00	28360.0
118.00	1895.0
119.00	257.0
120.10	307.0
121.10	184.0
122.00	2082.0
123.00	2951.0
124.00	1511.0
125.00	1382.0
126.10	532.0
127.00	88896.0
128.00	6583.0
129.00	37360.0
130.00	3595.0
131.00	646.0
131.80	301.0
133.00	289.0
134.00	1350.0
135.00	3423.0
136.00	1428.0
137.00	1473.0
137.80	412.0
139.00	283.0
140.00	445.0
141.00	5760.0
142.10	1668.0
143.00	1365.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

143.90	266.0
144.90	243.0
146.10	1144.0
147.00	2995.0
148.00	7813.0
149.00	1401.0
149.90	267.0
151.10	886.0
153.00	1980.0
153.90	1378.0
155.00	3381.0
156.00	4579.0
157.00	998.0
157.90	1292.0
159.00	854.0
160.00	2433.0
161.00	2557.0
161.90	541.0
162.90	412.0
164.00	234.0
165.00	2338.0
166.00	1686.0
167.00	12987.0
168.00	6766.0
169.00	1186.0
170.10	452.0
171.10	291.0
172.00	1147.0
173.00	1709.0
174.00	3283.0
175.00	4729.0
175.90	1063.0
177.10	2421.0
178.00	769.0
179.00	9881.0
180.00	6603.0
181.00	2902.0
182.10	443.0
183.00	338.0
184.00	767.0
185.00	4381.0
186.00	33872.0
187.00	10350.0
188.00	1188.0
189.00	2838.0
190.00	344.0
191.00	969.0
192.00	3224.0
193.00	4018.0
193.90	766.0
195.00	273.0
196.00	5688.0
197.10	1597.0
198.00	219328.0
198.90	14660.0
200.00	1370.0
201.00	540.0
201.50	874.0
202.00	294.0
202.90	2133.0
204.00	10358.0
205.00	15272.0
206.00	65232.0
207.10	10157.0
208.00	2514.0
209.00	1382.0
210.00	1459.0
211.00	2357.0
211.80	569.0
212.90	156.0
214.90	892.0
216.00	1537.0
217.00	20216.0
218.00	2390.0
219.00	433.0
220.20	479.0
221.00	7832.0
222.00	2296.0
223.00	4012.0
224.00	39104.0
225.00	10119.0
226.10	956.0
227.00	17976.0
228.00	2566.0
229.00	4120.0
230.00	421.0
231.00	1357.0
231.80	256.0
232.90	356.0
234.00	1664.0
234.90	1084.0
236.10	714.0
236.90	1206.0
237.80	153.0
239.00	858.0
240.00	746.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

240.80	989.0
242.00	2306.0
243.00	2114.0
244.00	28656.0
245.10	3988.0
246.00	6825.0
247.00	1383.0
247.90	353.0
249.00	1071.0
249.90	202.0
251.00	427.0
251.90	276.0
252.90	699.0
253.90	1003.0
255.00	165696.0
256.00	24704.0
257.10	1597.0
258.00	12443.0
259.00	1644.0
259.90	365.0
260.90	309.0
263.00	180.0
264.00	351.0
265.00	5240.0
266.00	709.0
267.10	200.0
271.00	444.0
272.00	527.0
273.00	5045.0
274.00	13223.0
275.00	71880.0
276.00	9379.0
277.00	6664.0
278.00	1685.0
278.90	269.0
281.00	745.0
282.00	356.0
283.00	922.0
284.00	621.0
285.00	1318.0
285.80	183.0
286.10	162.0
288.90	433.0
289.90	269.0
291.00	230.0
292.10	378.0
293.00	1455.0
293.90	405.0
295.00	563.0
296.00	26712.0
297.00	3211.0
297.90	291.0
300.90	359.0
302.10	438.0
303.00	3134.0
304.00	955.0
308.00	265.0
309.00	164.0
310.00	382.0
313.00	191.0
314.10	970.0
315.00	2866.0
315.90	1377.0
317.10	365.0
321.00	995.0
322.00	430.0
323.00	7397.0
324.00	1513.0
327.00	1221.0
327.90	995.0
332.00	762.0
333.00	703.0
334.00	5581.0
335.00	1630.0
336.10	191.0
341.00	821.0
342.00	239.0
345.90	1416.0
346.90	479.0
352.00	2656.0
353.00	1418.0
354.00	2417.0
355.00	594.0
365.00	12729.0
366.00	1687.0
366.90	178.0
369.90	161.0
371.00	730.0
372.10	4302.0
373.10	1076.0
383.00	1127.0
383.90	271.0
390.00	653.0
391.00	419.0
392.00	227.0
400.90	229.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69

402.00	2115.0
403.00	2430.0
404.10	870.0
405.00	203.0
421.00	2033.0
422.00	1944.0
423.10	12977.0
424.10	3800.0
425.00	427.0
441.10	35392.0
442.10	239808.0
443.10	44440.0
444.00	4173.0
445.20	227.0

Instrument :
BNA_M
ClientSampleId :
DFTPP69