


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM082819\
Data File : BM022379.D
Acq On : 28 Aug 2019 08:49
Operator : HP/JU
Sample : DFTPP09
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP09

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM082219MA.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 28 19:12:04 2019



AutoFind: Scans 2447, 2448, 2449; Background Corrected with Scan 2442

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	47474	PASS
68	69	0.00	2	1.8	816	PASS
69	198	0.00	100	33.1	44269	PASS
70	69	0.00	2	0.3	152	PASS
127	198	10	80	38.1	50968	PASS
197	198	0.00	2	0.8	1015	PASS
198	198	100	100	100.0	133816	PASS
199	198	5	9	6.8	9161	PASS
275	198	10	60	35.4	47400	PASS
365	198	1	100	8.5	11315	PASS
441	198	0.01	100	12.4	16618	PASS
442	198	50	100	77.3	103416	PASS
443	442	15	24	19.0	19644	PASS

m/z	Abundance
207.00	553.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
37.00	346.0
38.10	863.0
39.00	3994.0
43.90	329.0
49.00	538.0
50.00	16744.0
51.00	51488.0
52.00	2826.0
56.00	1175.0
57.00	2964.0
61.10	611.0
62.00	630.0
63.00	2045.0
65.00	745.0
68.00	951.0
69.00	45304.0
73.10	779.0
74.00	7368.0
75.00	9369.0
76.10	3445.0
77.10	58720.0
78.00	3922.0
79.00	5143.0
80.00	3576.0
81.00	4394.0
81.90	1260.0
83.00	701.0
84.90	793.0
86.00	2043.0
86.90	560.0
90.90	921.0
92.00	1122.0
93.00	10195.0
94.00	677.0
96.00	419.0
98.00	8811.0
99.00	5456.0
101.00	2282.0
102.90	816.0
104.00	1888.0
105.00	1978.0
106.00	566.0
107.00	20432.0
108.00	2625.0
109.10	500.0
110.00	25768.0
111.00	4615.0
112.00	698.0
115.90	1097.0
117.00	33536.0
118.00	2186.0
120.00	345.0
122.00	1634.0
123.00	2279.0
124.00	1207.0
124.90	838.0
127.00	50136.0
128.00	4386.0
129.00	24440.0
130.00	2320.0
130.90	434.0
133.90	859.0
135.00	2266.0
136.00	932.0
137.00	1445.0
139.90	404.0
140.90	4329.0
142.00	1287.0
143.00	800.0
145.90	948.0
146.90	2394.0
148.00	8213.0
149.00	1478.0
150.10	386.0
150.90	501.0
153.00	1358.0
153.90	925.0
155.00	2260.0
156.00	3287.0
157.00	537.0
158.00	752.0
159.00	733.0
160.00	1690.0
161.00	2136.0
161.90	578.0
164.90	2015.0
166.10	1036.0
167.00	13510.0
168.00	6497.0
168.90	873.0
169.90	305.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

171.00	405.0
172.00	813.0
173.00	1311.0
174.00	1973.0
175.00	3829.0
176.00	828.0
176.90	1485.0
178.00	775.0
178.90	7170.0
180.00	4833.0
181.00	2105.0
182.00	337.0
184.00	619.0
185.00	3926.0
186.00	23032.0
187.00	7162.0
188.00	805.0
189.00	2020.0
191.00	909.0
192.00	2307.0
193.00	2925.0
193.90	614.0
196.00	3801.0
197.00	1010.0
197.90	121264.0
198.90	8477.0
199.90	836.0
201.30	556.0
203.00	1756.0
204.00	7666.0
205.00	12019.0
206.00	44144.0
207.00	7003.0
208.00	2152.0
209.00	906.0
209.90	846.0
210.90	1800.0
214.90	630.0
216.00	1279.0
216.90	15013.0
218.00	2117.0
221.00	4100.0
221.90	1877.0
222.90	3237.0
224.00	27640.0
225.00	7067.0
226.00	806.0
227.00	13510.0
228.00	2142.0
229.00	2626.0
229.90	461.0
231.00	1082.0
234.00	755.0
234.90	1129.0
235.90	874.0
237.00	1397.0
239.00	533.0
239.90	399.0
240.90	866.0
242.00	1694.0
243.00	1655.0
244.00	19024.0
245.00	2292.0
246.00	5463.0
247.00	1128.0
248.90	628.0
252.80	655.0
254.10	755.0
255.00	103736.0
256.00	16640.0
257.00	1254.0
258.00	8631.0
259.00	1546.0
264.90	4243.0
266.00	494.0
272.00	393.0
273.00	2958.0
274.00	9002.0
275.00	38624.0
276.00	5539.0
277.00	5321.0
277.90	799.0
281.00	315.0
283.00	694.0
284.00	419.0
285.00	694.0
293.00	1152.0
295.00	340.0
296.00	18648.0
297.00	2911.0
303.00	1913.0
303.90	510.0
314.00	719.0
314.90	2053.0
316.00	931.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

321.00	417.0
323.00	3764.0
324.00	701.0
326.90	1033.0
328.00	503.0
333.00	404.0
333.90	2858.0
335.00	721.0
340.90	507.0
345.90	946.0
352.00	1075.0
352.90	604.0
354.00	1079.0
365.00	7966.0
365.90	1081.0
370.90	394.0
372.00	1654.0
373.10	498.0
382.80	329.0
402.00	582.0
403.00	725.0
404.00	473.0
420.90	629.0
421.90	504.0
423.00	3848.0
424.00	1029.0
441.10	10446.0
442.00	65696.0
443.00	12894.0
444.10	1088.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
37.00	435.0
38.00	1117.0
39.10	4230.0
49.00	541.0
50.00	18880.0
51.10	57856.0
52.00	2713.0
55.00	311.0
56.00	1505.0
57.00	3485.0
61.00	637.0
62.00	823.0
63.00	2580.0
65.00	801.0
68.00	907.0
69.00	55464.0
70.00	456.0
73.00	869.0
74.10	8557.0
75.00	10336.0
76.10	4077.0
77.00	70240.0
78.00	4878.0
79.00	5764.0
80.00	3974.0
81.00	5070.0
82.00	1198.0
83.00	875.0
85.00	822.0
85.90	1788.0
86.90	675.0
91.00	1340.0
92.00	1603.0
93.00	12856.0
94.00	754.0
98.00	10358.0
99.00	6240.0
100.00	608.0
101.00	2606.0
102.90	1022.0
104.00	2241.0
105.00	2048.0
105.90	638.0
107.00	25672.0
108.00	3114.0
109.00	785.0
110.00	31776.0
111.00	6098.0
111.90	845.0
113.00	362.0
116.00	1574.0
117.00	42256.0
118.00	2857.0
122.00	1666.0
123.00	3028.0
124.00	1217.0
125.00	1064.0
127.00	65576.0
128.00	5541.0
129.00	32408.0
129.90	2551.0
131.00	638.0
132.00	306.0
133.90	818.0
135.00	2703.0
135.90	1537.0
137.00	1924.0
138.00	319.0
140.10	573.0
141.00	5430.0
142.00	1605.0
143.00	1037.0
143.90	334.0
145.90	1140.0
147.00	2996.0
148.00	10015.0
149.00	1607.0
149.90	352.0
151.00	540.0
152.00	378.0
152.90	1730.0
154.00	1304.0
155.00	3351.0
156.00	4104.0
156.90	563.0
157.90	1202.0
159.00	977.0
160.00	1822.0
161.00	2812.0
162.00	716.0
165.00	2683.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

166.00	1703.0
167.00	16544.0
168.00	8759.0
169.00	1297.0
169.90	417.0
171.00	503.0
171.90	1636.0
173.00	1544.0
174.00	2214.0
175.00	4756.0
176.00	988.0
177.00	2330.0
178.00	882.0
178.90	9568.0
180.00	6346.0
181.00	2808.0
182.00	416.0
183.90	815.0
185.00	4249.0
186.00	32088.0
187.00	10166.0
188.00	860.0
188.90	2892.0
190.00	355.0
190.90	1388.0
192.00	3303.0
193.00	4036.0
193.90	694.0
196.00	4887.0
197.00	1223.0
197.90	172416.0
198.90	11721.0
200.00	1132.0
201.40	515.0
202.90	2209.0
204.00	10819.0
205.00	16239.0
206.00	62032.0
207.00	9357.0
207.90	3138.0
209.00	836.0
209.90	1006.0
211.00	2943.0
211.90	452.0
214.90	891.0
216.00	1693.0
217.00	20552.0
218.00	2636.0
221.00	5911.0
221.90	2411.0
223.00	4535.0
224.00	38168.0
225.00	9771.0
226.00	1222.0
227.00	20768.0
228.00	3177.0
229.00	3607.0
229.90	471.0
231.00	1404.0
232.00	300.0
234.00	1280.0
235.00	1326.0
235.90	1221.0
237.00	1842.0
238.90	1093.0
240.00	445.0
241.00	933.0
242.00	2247.0
243.00	2508.0
244.00	28016.0
245.00	3710.0
246.00	8355.0
246.90	1416.0
247.90	371.0
248.90	1001.0
250.90	399.0
253.00	795.0
254.10	1112.0
255.00	154176.0
256.00	23416.0
256.90	1888.0
258.00	11517.0
259.00	1800.0
260.00	364.0
260.90	318.0
265.00	5190.0
266.00	842.0
270.90	332.0
272.00	535.0
273.00	4839.0
274.00	13340.0
275.00	60912.0
276.00	8508.0
277.00	7932.0
277.90	1372.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

281.00	337.0
283.00	790.0
283.90	603.0
285.00	1435.0
288.80	325.0
293.00	1244.0
293.90	365.0
295.00	387.0
296.00	27992.0
297.00	3816.0
302.00	482.0
303.00	2646.0
304.00	677.0
307.90	399.0
309.00	353.0
314.00	1294.0
315.00	3206.0
315.90	1385.0
316.90	355.0
321.00	716.0
322.00	399.0
323.00	6770.0
324.00	1053.0
326.90	1337.0
328.00	765.0
332.00	489.0
332.90	828.0
334.00	4665.0
335.00	1136.0
341.00	655.0
346.00	1734.0
346.90	320.0
352.00	1923.0
353.00	1314.0
354.00	2024.0
355.00	365.0
365.00	13867.0
366.00	2004.0
371.00	490.0
372.00	3478.0
373.00	566.0
383.00	832.0
389.90	405.0
402.00	998.0
403.00	1567.0
403.90	546.0
421.00	1209.0
422.00	1040.0
423.10	7897.0
424.00	1814.0
441.00	21104.0
442.00	127032.0
443.10	24360.0
444.00	1900.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
38.10	419.0
39.00	1992.0
49.00	403.0
50.00	10480.0
51.00	33080.0
52.00	1412.0
56.00	955.0
57.00	1754.0
60.90	371.0
62.00	519.0
63.00	1416.0
65.10	556.0
68.10	591.0
69.00	32040.0
73.00	694.0
74.10	4797.0
75.00	6100.0
76.00	2271.0
77.10	40216.0
78.10	2283.0
79.00	3429.0
80.00	2289.0
81.00	2601.0
82.00	796.0
82.90	454.0
84.90	381.0
86.00	1172.0
86.90	358.0
91.10	744.0
92.00	763.0
93.00	6772.0
94.10	457.0
98.00	6513.0
99.00	3548.0
100.00	364.0
101.00	1850.0
102.90	544.0
104.00	1239.0
104.90	1296.0
107.00	14319.0
107.90	1799.0
109.00	391.0
110.00	18784.0
111.00	3227.0
112.00	502.0
116.00	968.0
117.00	25136.0
118.00	1457.0
121.90	1379.0
123.00	1788.0
123.90	762.0
124.90	582.0
127.00	37192.0
128.00	2944.0
129.00	18800.0
130.00	1909.0
131.00	315.0
133.90	822.0
135.00	1937.0
136.00	803.0
136.90	1481.0
140.90	3185.0
141.90	1065.0
143.00	700.0
146.00	737.0
147.00	1852.0
148.00	6089.0
149.00	1091.0
151.00	348.0
153.00	1222.0
153.90	782.0
155.00	1952.0
156.00	2863.0
156.90	343.0
157.90	722.0
158.90	357.0
160.00	1192.0
161.00	1735.0
161.90	361.0
165.00	1369.0
166.00	1160.0
167.00	10940.0
168.00	5254.0
169.00	940.0
170.90	420.0
172.00	666.0
172.90	1132.0
174.00	1909.0
175.00	3020.0
176.00	559.0
176.90	1500.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

178.00	626.0
178.90	6418.0
180.00	4194.0
181.00	1770.0
183.90	584.0
185.00	3300.0
186.00	20760.0
187.00	5917.0
188.00	643.0
188.90	1872.0
189.90	336.0
190.90	728.0
192.00	2087.0
193.00	2496.0
194.00	398.0
196.00	2690.0
197.00	813.0
197.90	107768.0
198.90	7287.0
200.00	648.0
201.50	362.0
203.00	1449.0
204.00	6923.0
205.00	10704.0
206.00	41152.0
207.00	6011.0
207.90	2045.0
208.90	554.0
210.00	846.0
210.90	1893.0
214.90	552.0
216.00	1228.0
216.90	13586.0
218.00	1541.0
220.90	3565.0
221.90	1696.0
223.00	3070.0
224.00	25920.0
225.00	6635.0
226.00	698.0
227.00	14485.0
228.00	1908.0
229.00	2328.0
231.00	796.0
234.00	967.0
235.00	963.0
236.00	753.0
237.00	874.0
238.90	757.0
239.90	485.0
241.00	740.0
242.00	1537.0
243.00	1748.0
244.00	18560.0
245.00	2665.0
246.00	5214.0
247.00	842.0
248.90	559.0
252.90	600.0
253.90	662.0
255.00	102848.0
256.00	15368.0
257.00	1195.0
258.00	8672.0
258.90	1527.0
264.90	3682.0
265.90	642.0
271.00	376.0
272.00	450.0
273.00	3394.0
274.00	8598.0
275.00	42664.0
276.00	5871.0
277.00	5206.0
277.90	1196.0
282.90	639.0
284.00	417.0
285.00	817.0
292.90	1345.0
294.00	308.0
295.00	348.0
296.00	21136.0
297.00	2884.0
302.00	329.0
303.00	2175.0
304.00	511.0
313.90	845.0
315.00	2793.0
316.00	1274.0
317.00	306.0
321.00	414.0
323.00	5019.0
324.00	931.0
326.90	1467.0
328.10	536.0

Instrument :

BNA_M

ClientSampleId :

DFTPP09

332.00	432.0
333.00	471.0
334.00	3892.0
334.90	945.0
341.00	543.0
345.90	1346.0
352.00	1514.0
353.00	983.0
354.00	1287.0
355.10	361.0
365.00	12114.0
366.00	1253.0
372.00	2753.0
373.00	639.0
383.00	845.0
389.90	369.0
402.00	923.0
403.00	1446.0
403.90	466.0
421.00	1137.0
422.00	967.0
423.00	7207.0
424.00	1686.0
441.00	18304.0
442.00	117520.0
443.10	21680.0
444.00	2159.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09