

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082823\
 Data File : BM041558.D
 Acq On : 29 Aug 2023 14:01
 Operator : MA/JU
 Sample : PB155102BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS102

Quant Time: Aug 29 23:48:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 28 15:20:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	10131	0.400	ng/ul	0.00
4) Naphthalene-d8	10.939	136	26474	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.749	164	12043	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.498	188	21421	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.688	240	5732	0.400	ng/ul	# 0.00
23) Perylene-d12	24.246	264	6065	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.444	96	9214	0.731	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.517	152	13573	0.388	ng/ul	0.00
18) Fluoranthene-d10	19.520	212	12951	0.552	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.482	88	22639	1.734	ng/ul#	86
5) Naphthalene	10.988	128	29715	0.351	ng/ul	100
7) 2-Methylnaphthalene	12.588	142	18058	0.352	ng/ul	99
8) 1-Methylnaphthalene	12.803	142	19206	0.370	ng/ul	99
10) Acenaphthylene	14.476	152	29359	0.369	ng/ul	100
11) Acenaphthene	14.809	153	20162	0.361	ng/ul	99
12) Fluorene	15.795	166	22023	0.365	ng/ul	100
14) Pentachlorophenol	17.122	266	4114	0.467	ng/ul	100
15) Phenanthrene	17.540	178	31874	0.363	ng/ul	100
16) Anthracene	17.633	178	28217	0.350	ng/ul	99
19) Fluoranthene	19.552	202	28873	0.544	ng/ul	100
20) Pyrene	19.919	202	27872	0.519	ng/ul	99
21) Benzo(a)anthracene	21.670	228	17148	0.420	ng/ul	99
22) Chrysene	21.726	228	17898	0.442	ng/ul	99
24) Benzo(b)fluoranthene	23.453	252	17790	0.461	ng/ul	96
25) Benzo(k)fluoranthene	23.506	252	17264	0.458	ng/ul	97
26) Benzo(a)pyrene	24.132	252	15505	0.445	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	26.954	276	20096	0.429	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.988	278	13742	0.408	ng/ul	97
29) Benzo(g,h,i)perylene	27.799	276	17113	0.436	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082823\
Data File : BM041558.D
Acq On : 29 Aug 2023 14:01
Operator : MA/JU
Sample : PB155102BS
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS102

Quant Time: Aug 29 23:48:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 28 15:20:06 2023
Response via : Initial Calibration

