


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $4.52 \times 10^{-1} \times \text{Amt} - 2.53 \times 10^{-1}$
Coef of Det (r^2) = 0.991 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM082919.M
Calibration Table Last Updated: Thu Aug 29 16:51:32 2019


2,4-Dinitrophenol

Response Ratio



Resp Ratio = $2.29 \times 10^{-1} \times \text{Amt} - 9.49 \times 10^{-2}$
Coef of Det (r^2) = 0.994 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM082919.M
Calibration Table Last Updated: Thu Aug 29 16:51:32 2019



Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM082919.M
Calibration Table Last Updated: Thu Aug 29 16:51:32 2019