

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082923\
 Data File : BM041588.D
 Acq On : 30 Aug 2023 21:15
 Operator : MA/JU
 Sample : PB155125BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK125

Quant Time: Aug 31 02:18:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 30 01:50:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.114	152	5740	0.400	ng/ul	0.00
4) Naphthalene-d8	10.944	136	13497	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.749	164	5890	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.502	188	10292	0.400	ng/ul	0.00
17) Chrysene-d12	21.688	240	3382	0.400	ng/ul	0.00
23) Perylene-d12	24.248	264	4230	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.452	96	51494	6.079	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.517	152	4909	0.301	ng/ul	0.00
18) Fluoranthene-d10	19.520	212	5296	0.318	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082923\
Data File : BM041588.D
Acq On : 30 Aug 2023 21:15
Operator : MA/JU
Sample : PB155125BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK125

Quant Time: Aug 31 02:18:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 30 01:50:06 2023
Response via : Initial Calibration

