

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082923\
 Data File : BM041621.D
 Acq On : 31 Aug 2023 18:43
 Operator : MA/JU
 Sample : PB155158BL
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK158

Quant Time: Sep 01 02:09:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 30 01:50:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	5613	0.400	ng/ul	0.00
4) Naphthalene-d8	10.939	136	13100	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.745	164	5566	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.498	188	9598	0.400	ng/ul	0.00
17) Chrysene-d12	21.682	240	3016	0.400	ng/ul	0.00
23) Perylene-d12	24.240	264	3973	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.448	96	54403	6.567	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.511	152	4714	0.298	ng/ul	-0.01
18) Fluoranthene-d10	19.520	212	5044	0.339	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082923\
Data File : BM041621.D
Acq On : 31 Aug 2023 18:43
Operator : MA/JU
Sample : PB155158BL
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK158

Quant Time: Sep 01 02:09:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 30 01:50:06 2023
Response via : Initial Calibration

