

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082923\
 Data File : BM041635.D
 Acq On : 01 Sep 2023 03:17
 Operator : MA/JU
 Sample : PB155158BS
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS158

Quant Time: Sep 01 05:14:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 01 04:06:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.106	152	6772	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.933	136	14335	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.740	164	5472	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.493	188	10220	0.400	ng/ul	-0.01
17) Chrysene-d12	21.676	240	3862	0.400	ng/ul	-0.01
23) Perylene-d12	24.231	264	4100	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.452	96	6242	0.625	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.511	152	6846	0.396	ng/ul	-0.01
18) Fluoranthene-d10	19.515	212	7154	0.376	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.486	88	15596	1.463	ng/ul#	86
5) Naphthalene	10.988	128	17668	0.385	ng/ul	100
7) 2-Methylnaphthalene	12.583	142	10139	0.390	ng/ul	100
8) 1-Methylnaphthalene	12.803	142	10674	0.406	ng/ul	99
10) Acenaphthylene	14.472	152	15488	0.406	ng/ul	100
11) Acenaphthene	14.805	153	10719	0.407	ng/ul	99
12) Fluorene	15.790	166	11619	0.406	ng/ul	98
14) Pentachlorophenol	17.118	266	2919	0.743	ng/ul	98
15) Phenanthrene	17.536	178	17584	0.394	ng/ul	99
16) Anthracene	17.628	178	15595	0.402	ng/ul	99
19) Fluoranthene	19.543	202	18144	0.358	ng/ul	99
20) Pyrene	19.910	202	17923	0.352	ng/ul	100
21) Benzo(a)anthracene	21.659	228	12132	0.352	ng/ul	100
22) Chrysene	21.717	228	12830	0.347	ng/ul	99
24) Benzo(b)fluoranthene	23.436	252	12971	0.341	ng/ul	99
25) Benzo(k)fluoranthene	23.488	252	12453	0.344	ng/ul	99
26) Benzo(a)pyrene	24.114	252	11046	0.326	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	26.924	276	14798	0.327	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.954	278	10054	0.330	ng/ul	97
29) Benzo(g,h,i)perylene	27.766	276	13013	0.342	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082923\
Data File : BM041635.D
Acq On : 01 Sep 2023 03:17
Operator : MA/JU
Sample : PB155158BS
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS158

Quant Time: Sep 01 05:14:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 01 04:06:38 2023
Response via : Initial Calibration

