

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM083019\
 Data File : BM022446.D
 Acq On : 30 Aug 2019 18:17
 Operator : HP/JU
 Sample : K4594-02RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 30 19:30:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 16:51:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	17278	20.00	ng	-0.01
21) Naphthalene-d8	10.29	136	59297	20.00	ng	0.00
39) Acenaphthene-d10	14.18	164	35654	20.00	ng	0.00
64) Phenanthrene-d10	16.94	188	78283	20.00	ng	0.00
76) Chrysene-d12	21.16	240	134947	20.00	ng	-0.01
87) Perylene-d12	23.35	264	167662	20.00	ng	-0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	102139	103.77	ng	0.00
7) Phenol-d6	6.72	99	126021	95.74	ng	0.00
23) Nitrobenzene-d5	8.69	82	88763	65.57	ng	0.00
42) 2,4,6-Tribromophenol	15.69	330	60777	83.30	ng	0.00
45) 2-Fluorobiphenyl	12.79	172	164467	52.26	ng	0.00
79) Terphenyl-d14	19.59	244	372402	36.42	ng	-0.01


Target Compounds				Qvalue		
10) Phenol	6.75	94	6609	4.998	ng	82
31) Naphthalene	10.34	128	9740	3.158	ng	97
50) Dimethylphthalate	13.66	163	20107	6.484	ng	# 97
52) Acenaphthene	14.25	154	30960	14.786	ng	97
55) Dibenzofuran	14.59	168	22748	6.449	ng	98
58) Fluorene	15.24	166	34369	10.906	ng	97
71) Phenanthrene	16.99	178	615416	134.868	ng	99
72) Anthracene	17.09	178	108322	24.128	ng	100
73) Carbazole	17.37	167	66663	17.964	ng	99
75) Fluoranthene	19.03	202	1277798	210.277	ng	99
78) Pyrene	19.39	202	971415	105.627	ng	99
81) Benzo(a)anthracene	21.15	228	607500	62.720	ng	99
83) Chrysene	21.20	228	553395	60.025	ng	97
84) Bis(2-ethylhexyl)phthalate	21.06	149	11958	2.744	ng	98
86) Indeno(1,2,3-cd)pyrene	25.53	276	405676	43.746	ng	98
88) Benzo(b)fluoranthene	22.70	252	870705	74.754	ng	100
89) Benzo(k)fluoranthene	22.70	252	870705	75.355	ng	98
90) Benzo(a)pyrene	23.26	252	575062	55.624	ng	99
91) Dibenzo(a,h)anthracene	25.53	278	106108	10.293	ng	# 90
92) Benzo(a,h,i)perylene	26.20	276	385746	42.933	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM083019\
Data File : BM022446.D
Acq On : 30 Aug 2019 18:17
Operator : HP/JU
Sample : K4594-02RE
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Aug 30 19:30:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 16:51:32 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.52 min Scan# 770
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampled :

Tot Ion:152 Resp: 17278
Ion Ratio Lower Upper
152 100
150 159.7 125.4 188.0
115 65.8 50.3 75.5




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 103.766 ng
RT: 5.14 min Scan# 366
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion:112 Resp: 102139
Ion Ratio Lower Upper
112 100
64 60.1 45.4 68.2
63 40.5 34.5 51.7



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 63.00 (62.70 to 63.70): BM0





#7

Phenol-d6

Concen: 95.740 ng

RT: 6.72 min Scan# 634

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampled :


Tot Ion: 99 Resp: 126021

Ion Ratio Lower Upper

99 100

42 30.2 26.3 39.5




71 54.0 46.6 69.8



Abundance Ion 99.00 (98.70 to 99.70): BM022196.D (-635) (-)

Ion 42.00 (41.70 to 42.70): BM022196.D (-635) (-)

Ion 71.00 (70.70 to 71.70): BM022196.D (-635) (-)



#10

Phenol

Concen: 4.998 ng

RT: 6.75 min Scan# 639

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion: 94 Resp: 6609

Ion Ratio Lower Upper

94 100

65 47.1 17.7 57.7



66 74.0 39.6 79.6




Abundance Ion 94.00 (93.70 to 94.70): BM022196.D (-640) (-)

Ion 65.00 (64.70 to 65.70): BM022196.D (-640) (-)

Ion 66.00 (65.70 to 66.70): BM022196.D (-640) (-)






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.29 min Scan# 1242
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampled :

Tot Ion:136	Resp:	59297
Ion Ratio	Lower	Upper
136	100	
137	10.9	8.6 13.0
54	8.2	7.0 10.6
68	4.1	2.9 4.3






Abundance

Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BM

Ion 68.00 (67.70 to 68.70): BM



#23
Nitrobenzene-d5
Concen: 65.566 ng
RT: 8.69 min Scan# 970
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion: 82	Resp:	88763
Ion Ratio	Lower	Upper
82	100	
128	36.5	27.9 41.9
54	56.1	45.2 67.8






Abundance

Ion 82.00 (81.70 to 82.70): BM

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BM









#31
Naphthalene
Concen: 3.158 ng
RT: 10.34 min Scan# 1250
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :

Tot Ion:128 Resp: 9740
Ion Ratio Lower Upper
128 100
129 12.3 9.0 13.4
127 12.6 11.0 16.4




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.18 min Scan# 1903
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion:164 Resp: 35654
Ion Ratio Lower Upper
164 100
162 108.8 80.6 121.0
160 46.1 37.3 55.9



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#42
 2,4,6-Tribromophenol
 Concen: 83.304 ng
 RT: 15.69 min Scan# 2160
 Delta R.T. -0.01 min
 Lab File: BM022446.D
 Acq: 30 Aug 2019 18:17

Instrument :
 BNA_M
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.8	77.8	116.8
141	47.8	35.6	53.4






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 52.257 ng
 RT: 12.79 min Scan# 1666
 Delta R.T. -0.01 min
 Lab File: BM022446.D
 Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.1	28.5	42.7
170	23.0	18.9	28.3






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#50
Dimethylphthalate
Concen: 6.484 ng
RT: 13.66 min Scan# 1814
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17




Instrument :
BNA_M
ClientSampleId :


Tot Ion:163 Resp: 20107
Ion Ratio Lower Upper
163 100
194 6.1 3.4 5.2#
164 11.0 8.2 12.2



#52
Acenaphthene
Concen: 14.786 ng
RT: 14.25 min Scan# 1914
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion:154 Resp: 30960
Ion Ratio Lower Upper
154 100
153 116.7 94.3 141.5
152 61.5 45.4 68.2









#55
Dibenzofuran
Concen: 6.449 ng
RT: 14.59 min Scan# 1973
Delta R.T. -0.00 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampled :

Tot Ion:168 Resp: 22748
Ion Ratio Lower Upper
168 100
139 45.7 36.1 54.1
169 15.0 10.4 15.6




Abundance Ion 168.00 (167.70 to 168.70): E
20000 Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#58
Fluorene
Concen: 10.906 ng
RT: 15.24 min Scan# 2084
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion:166 Resp: 34369
Ion Ratio Lower Upper
166 100
165 98.6 81.1 121.7
167 14.9 11.1 16.7



Abundance Ion 166.00 (165.70 to 166.70): E
30000 Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 16.94 min Scan# 2373

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :


Tot Ion:188 Resp: 78283

Ion Ratio Lower Upper

188 100

94 5.7 4.4 6.6




80 7.2 5.0 7.6



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0



#71

Phenanthrene

Concen: 134.868 ng

RT: 16.99 min Scan# 2381

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:178 Resp: 615416

Ion Ratio Lower Upper

178 100

176 20.1 15.4 23.2



179 15.5 12.3 18.5




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Anthracene

Concen: 24.128 ng

RT: 17.09 min Scan# 2397

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :


Tot Ion:178 Resp: 108322

Ion Ratio Lower Upper

178 100

176 18.9 15.0 22.4




179 14.7 11.8 17.6



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#73

Carbazole

Concen: 17.964 ng

RT: 17.37 min Scan# 2446

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:167 Resp: 66663

Ion Ratio Lower Upper

167 100

166 22.3 17.8 26.6



139 15.6 11.6 17.4




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#75

Fluoranthene

Concen: 210.277 ng

RT: 19.03 min Scan# 2727

Delta R.T. -0.00 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :


Tot Ion:202 Resp: 1277798

Ion Ratio Lower Upper

202 100

101 4.5 0.0 25.1




203 17.2 0.0 37.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.16 min Scan# 3090

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:240 Resp: 134947

Ion Ratio Lower Upper

240 100

120 4.9 3.9 5.9



236 27.4 21.4 32.2




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Pvrene

Concen: 105.627 ng

RT: 19.39 min Scan# 2789

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :


Tot Ion:202 Resp: 971415

Ion Ratio Lower Upper

202 100

200 21.5 17.7 26.5

203 17.7 13.8 20.8






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#79

Terphenyl-d14

Concen: 36.420 ng

RT: 19.59 min Scan# 2823

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:244 Resp: 372402

Ion Ratio Lower Upper

244 100

212 8.9 7.0 10.4

122 6.3 3.9 5.9#






Abundance

Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#81

Benzo(a)anthracene

Concen: 62.720 ng

RT: 21.15 min Scan# 3088

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :


Tot Ion:228 Resp: 607500

Ion Ratio Lower Upper




228 100

226 28.7 22.6 34.0

229 20.6 15.9 23.9



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#83

Chrysene

Concen: 60.025 ng

RT: 21.20 min Scan# 3097

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:228 Resp: 553395

Ion Ratio Lower Upper



228 100


226 30.0 24.5 36.7

229 21.9 15.7 23.5



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#84

Bis(2-ethylhexyl)phthalate

Concen: 2.744 ng

RT: 21.06 min Scan# 3073

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :


Tot Ion:149 Resp: 11958

Ion Ratio Lower Upper

149 100

167 23.5 19.6 29.4

279 5.7 4.4 6.6






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#86

Indeno(1,2,3-cd)pyrene

Concen: 43.746 ng

RT: 25.53 min Scan# 3833

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:276 Resp: 405676

Ion Ratio Lower Upper

276 100

138 11.3 9.1 13.7

277 23.7 20.0 30.0






Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Pervlene-d12
Concen: 20.000 ng
RT: 23.35 min Scan# 3462
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17




Instrument :
BNA_M
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
264	100		
260	25.1	20.2	30.2
265	24.0	18.2	27.2



#88
Benzo(b)fluoranthene
Concen: 74.754 ng
RT: 22.70 min Scan# 3352
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.9	26.9
125	4.7	3.5	5.3









#89
Benzo(k)fluoranthene
Concen: 75.355 ng
RT: 22.70 min Scan# 3352
Delta R.T. -0.05 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17




Instrument :
BNA_M
ClientSampleId :


Tot Ion:252 Resp: 870705
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.4
125 4.7 3.7 5.5



#90
Benzo(a)pyrene
Concen: 55.624 ng
RT: 23.26 min Scan# 3447
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion:252 Resp: 575062
Ion Ratio Lower Upper
252 100
253 23.3 18.2 27.2
125 4.6 3.8 5.8









#91
Dibenzo(a,h)anthracene
Concen: 10.293 ng
RT: 25.53 min Scan# 3833
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampled :

Tot Ion:278 Resp: 106108
Ion Ratio Lower Upper
278 100
139 0.0 6.6 9.8#
279 27.0 19.3 28.9



#92
Benzo(g,h,i)perylene
Concen: 42.933 ng
RT: 26.20 min Scan# 3947
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion:276 Resp: 385746
Ion Ratio Lower Upper
276 100
277 25.2 19.4 29.0
138 11.0 8.8 13.2

