

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM083019\
 Data File : BM022446.D
 Acq On : 30 Aug 2019 18:17
 Operator : HP/JU
 Sample : K4594-02RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 S704A-N-1.0-1.5-082819RE

Manual Integrations
 APPROVED

mohammad
 8/31/2019 2:43:57 PM

Quant Time: Aug 30 19:52:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 16:51:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	17278	20.00	ng	-0.01
21) Naphthalene-d8	10.29	136	59297	20.00	ng	0.00
39) Acenaphthene-d10	14.18	164	35654	20.00	ng	0.00
64) Phenanthrene-d10	16.94	188	78283	20.00	ng	0.00
76) Chrysene-d12	21.16	240	134947	20.00	ng	-0.01
87) Perylene-d12	23.35	264	167662	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	102139	103.77	ng	0.00
7) Phenol-d6	6.72	99	126021	95.74	ng	0.00
23) Nitrobenzene-d5	8.69	82	88763	65.57	ng	0.00
42) 2,4,6-Tribromophenol	15.69	330	60777	83.30	ng	0.00
45) 2-Fluorobiphenyl	12.79	172	164467	52.26	ng	0.00
79) Terphenyl-d14	19.59	244	372402	36.42	ng	-0.01
Target Compounds						Qvalue
10) Phenol	6.75	94	6609	4.998	ng	82
31) Naphthalene	10.34	128	9740	3.158	ng	97
50) Dimethylphthalate	13.66	163	20107	6.484	ng	# 97
52) Acenaphthene	14.25	154	30960	14.786	ng	97
55) Dibenzofuran	14.59	168	22748	6.449	ng	98
58) Fluorene	15.24	166	34369	10.906	ng	97
71) Phenanthrene	16.99	178	615416	134.868	ng	99
72) Anthracene	17.09	178	108322	24.128	ng	100
73) Carbazole	17.37	167	66663	17.964	ng	99
75) Fluoranthene	19.03	202	1277798	210.277	ng	99
78) Pyrene	19.39	202	971415	105.627	ng	99
81) Benzo(a)anthracene	21.15	228	607500	62.720	ng	99
83) Chrysene	21.20	228	553395	60.025	ng	97
84) Bis(2-ethylhexyl)phthalate	21.06	149	11958	2.744	ng	98
86) Indeno(1,2,3-cd)pyrene	25.53	276	405676	43.746	ng	98
88) Benzo(b)fluoranthene	22.70	252	870705	74.754	ng	100
89) Benzo(k)fluoranthene	22.74	252	234053m	20.256	ng	
90) Benzo(a)pyrene	23.26	252	575062	55.624	ng	99
91) Dibenzo(a,h)anthracene	25.53	278	106108	10.293	ng	# 90
92) Benzo(a,h,i)perylene	26.20	276	385746	42.933	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM083019\
Data File : BM022446.D
Acq On : 30 Aug 2019 18:17
Operator : HP/JU
Sample : K4594-02RE
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM

Quant Time: Aug 30 19:52:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM082919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 16:51:32 2019
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.52 min Scan# 770
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
152	100		
150	159.7	125.4	188.0
115	65.8	50.3	75.5

Manual Integrations
APPROVED




mohammad
8/31/2019 2:43:57 PM




#5

2-Fluorophenol
Concen: 103.766 ng
RT: 5.14 min Scan# 366
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
112	100		
64	60.1	45.4	68.2
63	40.5	34.5	51.7





#7

Phenol-d6

Concen: 95.740 ng

RT: 6.72 min Scan# 634

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M





ClientSampleId :

S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
99	100		
42	30.2	26.3	39.5
71	54.0	46.6	69.8

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM



#10

Phenol

Concen: 4.998 ng




RT: 6.75 min Scan# 639


Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
94	100		
65	47.1	17.7	57.7
66	74.0	39.6	79.6









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.29 min Scan# 1242
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	10.9	8.6	13.0	
54	8.2	7.0	10.6	
68	4.1	2.9	4.3	




Manual Integrations
APPROVED


mohammad
8/31/2019 2:43:57 PM



#23
Nitrobenzene-d5
Concen: 65.566 ng
RT: 8.69 min Scan# 970
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	36.5	27.9	41.9	
54	56.1	45.2	67.8	






#31
Naphthalene
Concen: 3.158 ng
RT: 10.34 min Scan# 1250
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
128	100		
129	12.3	9.0	13.4
127	12.6	11.0	16.4

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM






Abundance

Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.18 min Scan# 1903
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17




Tgt Ion	Ratio	Lower	Upper
164	100		
162	108.8	80.6	121.0
160	46.1	37.3	55.9


Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E






#42
 2,4,6-Tribromophenol
 Concen: 83.304 ng
 RT: 15.69 min Scan# 2160
 Delta R.T. -0.01 min
 Lab File: BM022446.D
 Acq: 30 Aug 2019 18:17

Instrument :
 BNA_M
 ClientSampleId :
 S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.8	77.8	116.8
141	47.8	35.6	53.4

Manual Integrations
 APPROVED

mohammad
 8/31/2019 2:43:57 PM






Abundance

Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 52.257 ng
 RT: 12.79 min Scan# 1666
 Delta R.T. -0.01 min
 Lab File: BM022446.D
 Acq: 30 Aug 2019 18:17




Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.1	28.5	42.7
170	23.0	18.9	28.3


Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 6.484 ng
RT: 13.66 min Scan# 1814
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
163	100		
194	6.1	3.4	5.2#
164	11.0	8.2	12.2

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM






Abundance

Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#52
Acenaphthene
Concen: 14.786 ng
RT: 14.25 min Scan# 1914
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17




Tgt Ion	Ratio	Lower	Upper
154	100		
153	116.7	94.3	141.5
152	61.5	45.4	68.2


Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E






#55
Dibenzofuran
Concen: 6.449 ng
RT: 14.59 min Scan# 1973
Delta R.T. -0.00 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
168	100		
139	45.7	36.1	54.1
169	15.0	10.4	15.6

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM






Abundance

Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#58
Fluorene
Concen: 10.906 ng
RT: 15.24 min Scan# 2084
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17




Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.6	81.1	121.7
167	14.9	11.1	16.7


Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.94 min Scan# 2373
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Resp	Lower	Upper
188	100	78283		
94	5.7		4.4	6.6
80	7.2		5.0	7.6




Manual Integrations
APPROVED


mohammad
8/31/2019 2:43:57 PM



#71
Phenanthrene
Concen: 134.868 ng
RT: 16.99 min Scan# 2381
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	615416		
176	20.1		15.4	23.2
179	15.5		12.3	18.5









#72
 Anthracene
 Concen: 24.128 ng
 RT: 17.09 min Scan# 2397
 Delta R.T. -0.01 min
 Lab File: BM022446.D
 Acq: 30 Aug 2019 18:17

Instrument :
 BNA_M
 ClientSampleId :
 S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	14.7	11.8	17.6




Manual Integrations
 APPROVED


mohammad
 8/31/2019 2:43:57 PM



#73
 Carbazole
 Concen: 17.964 ng
 RT: 17.37 min Scan# 2446
 Delta R.T. -0.01 min
 Lab File: BM022446.D
 Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
167	100		
166	22.3	17.8	26.6
139	15.6	11.6	17.4





#75

Fluoranthene

Concen: 210.277 ng

RT: 19.03 min Scan# 2727

Delta R.T. -0.00 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M

ClientSampleId :

S704A-N-1.0-1.5-082819RE

Tot Ion:202 Resp: 1277798

Ion Ratio Lower Upper

202 100


101 4.5 0.0 25.1

203 17.2 0.0 37.1

Manual Integrations
APPROVED

mohammad




8/31/2019 2:43:57 PM



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.16 min Scan# 3090

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17


Tgt Ion:240 Resp: 134947

Ion Ratio Lower Upper

240 100

120 4.9 3.9 5.9



236 27.4 21.4 32.2




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 105.627 ng
RT: 19.39 min Scan# 2789
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Resp	Lower	Upper
202	100			
200	21.5	17.7	26.5	
203	17.7	13.8	20.8	




Manual Integrations
APPROVED


mohammad
8/31/2019 2:43:57 PM



#79
Terphenyl-d14
Concen: 36.420 ng
RT: 19.59 min Scan# 2823
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Resp	Lower	Upper
244	100			
212	8.9	7.0	10.4	
122	6.3	3.9	5.9	





#81

Benzo(a)anthracene

Concen: 62.720 ng

RT: 21.15 min Scan# 3088

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M





ClientSampleId :

S704A-N-1.0-1.5-082819RE

Tot Ion:	228	Resp:	607500
Ion Ratio	Lower	Upper	
228	100		
226	28.7	22.6	34.0
229	20.6	15.9	23.9

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM



#83

Chrysene

Concen: 60.025 ng




RT: 21.20 min Scan# 3097


Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Tgt Ion:	228	Resp:	553395
Ion Ratio	Lower	Upper	
228	100		
226	30.0	24.5	36.7
229	21.9	15.7	23.5





#84

Bis(2-ethylhexyl)phthalate

Concen: 2.744 ng

RT: 21.06 min Scan# 3073

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M


ClientSampleId :

S704A-N-1.0-1.5-082819RE

Tot Ion:	149	Resp:	11958
Ion Ratio	Lower	Upper	
149	100		
167	23.5	19.6	29.4
279	5.7	4.4	6.6

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#86

Indeno(1,2,3-cd)pyrene

Concen: 43.746 ng

RT: 25.53 min Scan# 3833

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17




Tgt Ion:	276	Resp:	405676
Ion Ratio	Lower	Upper	
276	100		
138	11.3	9.1	13.7
277	23.7	20.0	30.0


Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 23.35 min Scan# 3462

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M





ClientSampleId :

S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
264	100		
260	25.1	20.2	30.2
265	24.0	18.2	27.2

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM



#88

Benzo(b)fluoranthene

Concen: 74.754 ng




RT: 22.70 min Scan# 3352


Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.9	26.9
125	4.7	3.5	5.3









#89
Benzo(k)fluoranthene
Concen: 20.256 ng m
RT: 22.74 min Scan# 3359
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Instrument :
BNA_M
ClientSampleId :
S704A-N-1.0-1.5-082819RE

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.0	25.4
125	5.1	3.7	5.5




Manual Integrations
APPROVED


mohammad
8/31/2019 2:43:57 PM



#90
Benzo(a)pyrene
Concen: 55.624 ng
RT: 23.26 min Scan# 3447
Delta R.T. -0.01 min
Lab File: BM022446.D
Acq: 30 Aug 2019 18:17

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.3	18.2	27.2
125	4.6	3.8	5.8





#91

Dibenzo(a,h)anthracene

Concen: 10.293 ng

RT: 25.53 min Scan# 3833

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Instrument :

BNA_M





ClientSampleId :

S704A-N-1.0-1.5-082819RE

Tot Ion:	278	Resp:	106108
Ion Ratio	Lower	Upper	
278	100		
139	0.0	6.6	9.8#
279	27.0	19.3	28.9

Manual Integrations
APPROVED

mohammad
8/31/2019 2:43:57 PM



#92

Benzo(g,h,i)perylene

Concen: 42.933 ng

RT: 26.20 min Scan# 3947

Delta R.T. -0.01 min

Lab File: BM022446.D

Acq: 30 Aug 2019 18:17

Tgt Ion:	276	Resp:	385746
Ion Ratio	Lower	Upper	
276	100		
277	25.2	19.4	29.0
138	11.0	8.8	13.2

