

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM083022\
 Data File : BM036463.D
 Acq On : 30 Aug 2022 16:35
 Operator : CG/JU
 Sample : PB147326BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS326

Quant Time: Aug 31 01:00:37 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM083022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 17:45:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	3472	0.400	ng/ul	0.00
4) Naphthalene-d8	10.539	136	15438	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.385	164	9460	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.131	188	20716	0.400	ng/ul	0.00
17) Chrysene-d12	21.321	240	18722	0.400	ng/ul	0.00
23) Perylene-d12	23.616	264	16112	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.143	96	3909	0.775	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.140	152	9486	0.394	ng/ul	0.00
18) Fluoranthene-d10	19.163	212	22334	0.383	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.176	88	10387	2.167	ng/ul#	78
5) Naphthalene	10.589	128	15599	0.369	ng/ul	97
7) 2-Methylnaphthalene	12.211	142	10140	0.392	ng/ul	93
8) 1-Methylnaphthalene	12.431	142	10632	0.404	ng/ul	99
10) Acenaphthylene	14.108	152	14800	0.412	ng/ul	99
11) Acenaphthene	14.450	153	12231	0.385	ng/ul	99
12) Fluorene	15.440	166	13467	0.381	ng/ul	98
14) Pentachlorophenol	16.798	266	2872	0.770	ng/ul	98
15) Phenanthrene	17.174	178	22709	0.376	ng/ul	99
16) Anthracene	17.271	178	20541	0.404	ng/ul	99
19) Fluoranthene	19.196	202	26355	0.392	ng/ul	98
20) Pyrene	19.558	202	27022	0.389	ng/ul	98
21) Benzo(a)anthracene	21.307	228	23112	0.409	ng/ul	99
22) Chrysene	21.356	228	25617	0.405	ng/ul	99
24) Benzo(b)fluoranthene	22.917	252	24892	0.388	ng/ul	93
25) Benzo(k)fluoranthene	22.964	252	24240	0.379	ng/ul#	90
26) Benzo(a)pyrene	23.513	252	23016	0.439	ng/ul#	85
27) Indeno(1,2,3-cd)pyrene	25.987	276	24568	0.394	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.007	278	20182	0.377	ng/ul#	90
29) Benzo(g,h,i)perylene	26.718	276	21827	0.367	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM083022\
Data File : BM036463.D
Acq On : 30 Aug 2022 16:35
Operator : CG/JU
Sample : PB147326BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS326

Quant Time: Aug 31 01:00:37 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM083022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 29 17:45:20 2022
Response via : Initial Calibration

