

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090623\
 Data File : BM041694.D
 Acq On : 07 Sep 2023 14:30
 Operator : MA/JU
 Sample : 04166-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BHAS5

Quant Time: Sep 08 02:53:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM090523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 06 02:20:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	4934	0.400	ng/ul	0.00
4) Naphthalene-d8	10.818	136	11639	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.643	164	5143	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.392	188	9117	0.400	ng/ul	0.00
17) Chrysene-d12	21.565	240	3801	0.400	ng/ul	0.00
23) Perylene-d12	24.000	264	4863	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.402	96	37394	4.959	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.401	152	5093	0.340	ng/ul	0.00
18) Fluoranthene-d10	19.413	212	6880	0.373	ng/ul	0.00
Target Compounds						
2) 1,4-Dioxane	3.440	88	1064	0.136	ng/ul#	Qvalue 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090623\
 Data File : BM041694.D
 Acq On : 07 Sep 2023 14:30
 Operator : MA/JU
 Sample : 04166-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BHAS5

Quant Time: Sep 08 02:53:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM090523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 06 02:20:07 2023
 Response via : Initial Calibration

