

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090822\
 Data File : BM036523.D
 Acq On : 07 Sep 2022 23:37
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020169

Quant Time: Sep 08 00:57:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 02 01:33:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	265300	20.000	ng/u1	0.00
20) Naphthalene-d8	10.522	136	1058924	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.374	164	773359	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.115	188	1564857	20.000	ng/u1	0.00
79) Chrysene-d12	21.309	240	1425458	20.000	ng/u1	0.00
88) Perylene-d12	23.597	264	1171740	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.128	96	44466	6.388	ng/uL	0.00
4) Pyridine-d5	3.557	84	333821	17.389	ng/u1	0.00
7) Phenol-d5	6.904	99	397065	17.534	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	282008	19.784	ng/u1	0.00
11) 2-Chlorophenol-d4	7.251	132	355719	19.992	ng/u1	0.00
15) 4-Methylphenol-d8	8.445	113	354278	20.953	ng/u1	0.00
21) Nitrobenzene-d5	8.892	128	187874	23.401	ng/u1	0.00
24) 2-Nitrophenol-d4	9.610	143	199386	24.483	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.145	165	354431	24.357	ng/u1	0.00
31) 4-Chloroaniline-d4	10.669	131	461512	20.459	ng/u1	0.00
46) Dimethylphthalate-d6	13.798	166	1003334	20.507	ng/u1	0.00
49) Acenaphthylene-d8	14.063	160	1234834	20.236	ng/u1	0.00
54) 4-Nitrophenol-d4	14.598	143	192228	20.735	ng/u1	0.00
60) Fluorene-d10	15.368	176	899530	20.346	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.504	200	152205	18.919	ng/u1	0.00
73) Anthracene-d10	17.215	188	1386359	19.924	ng/u1	0.00
81) Pyrene-d10	19.515	212	1535253	20.762	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.450	264	1159944	20.191	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.163	88	48927	6.673	ng/uL	99
5) Pyridine	3.575	79	343642	17.329	ng/u1	94
6) Benzaldehyde	6.869	77	209485	22.520	ng/u1	93
8) Phenol	6.928	94	430745	18.295	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.157	93	369987	19.647	ng/u1	98
12) 2-Chlorophenol	7.286	128	365883	19.665	ng/u1	98
13) 2-Methylphenol	8.181	108	353250	20.354	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.251	45	539336	21.406	ng/u1	99
16) Acetophenone	8.557	105	578487	21.272	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.539	70	300113	22.172	ng/u1	97
18) 4-Methylphenol	8.510	108	387205	20.842	ng/u1	99
19) Hexachloroethane	8.792	117	157915	20.287	ng/u1	92
22) Nitrobenzene	8.933	77	438790	23.066	ng/u1	99
23) Isophorone	9.463	82	834846	24.713	ng/u1	97
25) 2-Nitrophenol	9.639	139	216596	23.872	ng/u1	94
26) 2,4-Dimethylphenol	9.710	107	426155	23.486	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.945	93	513017	23.327	ng/u1	99
29) 2,4-Dichlorophenol	10.175	162	359904	24.175	ng/u1	98
30) Naphthalene	10.569	128	1084848	19.519	ng/u1	100
32) 4-Chloroaniline	10.692	127	469236	20.451	ng/u1	100
33) Hexachlorobutadiene	10.845	225	195968	20.953	ng/u1	99
34) Caprolactam	11.486	113	69969	15.686	ng/u1	92
35) 4-Chloro-3-methylphenol	11.827	107	345206	22.933	ng/u1	99
36) 2-Methylnaphthalene	12.186	142	823066	23.603	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090822\
 Data File : BM036523.D
 Acq On : 07 Sep 2022 23:37
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020169

Quant Time: Sep 08 00:57:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 02 01:33:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.410	142	851283	24.241	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.557	216	409540	19.159	ng/ul	98
40) Hexachlorocyclopentadiene	12.527	237	212926	16.791	ng/ul	99
41) 2,4,6-Trichlorophenol	12.810	196	277620	20.640	ng/ul	99
42) 2,4,5-Trichlorophenol	12.880	196	292039	20.469	ng/ul	98
43) 1,1'-Biphenyl	13.210	154	1119965	19.255	ng/ul	99
44) 2-Chloronaphthalene	13.245	162	863044	19.336	ng/ul	100
45) 2-Nitroaniline	13.468	65	261713	21.654	ng/ul	96
47) Dimethylphthalate	13.845	163	1037461	20.600	ng/ul	99
48) 2,6-Dinitrotoluene	13.968	165	211316	22.239	ng/ul	90
50) Acenaphthylene	14.092	152	1431087	19.887	ng/ul	99
51) 3-Nitroaniline	14.298	138	226293	21.776	ng/ul	100
52) Acenaphthene	14.439	153	924699	19.715	ng/ul	98
53) 2,4-Dinitrophenol	14.510	184	104818	20.241	ng/ul	93
55) 4-Nitrophenol	14.615	109	155185	21.186	ng/ul	83
56) Dibenzofuran	14.774	168	1287702	19.786	ng/ul	97
57) 2,4-Dinitrotoluene	14.757	165	302162	22.422	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	15.004	232	236203	21.805	ng/ul#	92
59) Diethylphthalate	15.204	149	1077711	21.709	ng/ul	99
61) Fluorene	15.421	166	1050623	20.353	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.421	204	501814	20.811	ng/ul	99
63) 4-Nitroaniline	15.462	138	216695	22.354	ng/ul#	86
66) 4,6-Dinitro-2-methylph...	15.515	198	151993	18.686	ng/ul#	90
67) N-Nitrosodiphenylamine	15.639	169	867809	19.398	ng/ul	98
68) 4-Bromophenyl-phenylether	16.309	248	288206	19.890	ng/ul	95
69) Hexachlorobenzene	16.415	284	345390	19.670	ng/ul	95
70) Atrazine	16.598	200	311196	20.358	ng/ul	98
71) Pentachlorophenol	16.774	266	208743	21.563	ng/ul	99
72) Phenanthrene	17.162	178	1614816	19.629	ng/ul	98
74) Anthracene	17.251	178	1653647	20.050	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.168	216	421947	17.971	ng/ul	97
76) Pentachlorobenzene	14.686	250	412927	18.835	ng/ul	98
77) Carbazole	17.527	167	1510656	19.563	ng/ul	99
78) Di-n-butylphthalate	18.092	149	1902263	22.522	ng/ul	99
80) Fluoranthene	19.180	202	1870074	20.926	ng/ul	98
82) Pyrene	19.545	202	1902039	20.396	ng/ul	97
83) Butylbenzylphthalate	20.450	149	853102	24.453	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.233	252	563024	19.181	ng/ul	98
85) Benzo(a)anthracene	21.291	228	1748210	20.408	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.227	149	1291460	26.811	ng/ul	98
87) Chrysene	21.344	228	1639221	19.475	ng/ul	98
89) Di-n-octyl phthalate	22.115	149	2008566	28.836	ng/ul	100
90) Benzo(b)fluoranthene	22.903	252	1564781	21.904	ng/ul	98
91) Benzo(k)fluoranthene	22.950	252	1402156	20.630	ng/ul	98
93) Benzo(a)pyrene	23.497	252	1415680	20.143	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.962	276	1284297	15.542	ng/ul	93
95) Dibenzo(a,h)anthracene	25.974	278	1111884	15.620	ng/ul	96
96) Benzo(g,h,i)perylene	26.685	276	1184748	16.766	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 6.673 ng/uL
RT: 3.163 min Scan# 30
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion: 88 Resp: 48927
Ion Ratio Lower Upper
88 100
43 34.0 27.8 41.8
58 77.9 61.8 92.8



#5
Pyridine
Concen: 17.329 ng/uL
RT: 3.575 min Scan# 100
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion: 79 Resp: 343642
Ion Ratio Lower Upper
79 100
52 78.5 67.8 101.6
51 37.1 31.4 47.0









#6
Benzaldehyde
Concen: 22.520 ng/ul
RT: 6.869 min Scan# 660
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion: 77 Resp: 209485
Ion Ratio Lower Upper
77 100
105 94.9 84.6 127.0
106 96.2 80.3 120.5



#8
Phenol
Concen: 18.295 ng/ul
RT: 6.928 min Scan# 670
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion: 94 Resp: 430745
Ion Ratio Lower Upper
94 100
65 29.3 21.3 31.9
66 38.4 27.4 41.0





#10

Bis(2-Chloroethyl)ether

Concen: 19.647 ng/ul

RT: 7.157 min Scan# 709

Delta R.T. -0.000 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion: 93 Resp: 369987

Ion Ratio Lower Upper

93 100

63 78.2 64.0 96.0

95 32.0 25.8 38.6



#12

2-Chlorophenol

Concen: 19.665 ng/ul

RT: 7.286 min Scan# 731

Delta R.T. -0.000 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37




Tgt Ion: 128 Resp: 365883


Ion Ratio Lower Upper

128 100

64 48.1 36.9 55.3

130 32.2 25.6 38.4









#13
2-Methylphenol
Concen: 20.354 ng/ul
RT: 8.181 min Scan# 883
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:108 Resp: 353250
Ion Ratio Lower Upper
108 100
107 89.7 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.406 ng/ul
RT: 8.251 min Scan# 895
Delta R.T. -0.012 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion: 45 Resp: 539336
Ion Ratio Lower Upper
45 100
77 13.9 10.6 15.8
79 10.2 8.3 12.5









#16
Acetophenone
Concen: 21.272 ng/ul
RT: 8.557 min Scan# 947
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:105 Resp: 578487
Ion Ratio Lower Upper
105 100
77 76.8 59.8 89.6
51 28.1 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 22.172 ng/ul
RT: 8.539 min Scan# 944
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion: 70 Resp: 300113
Ion Ratio Lower Upper
70 100
42 63.0 49.6 74.4
101 9.2 8.1 12.1
130 20.1 18.4 27.6









#18
4-Methylphenol
Concen: 20.842 ng/ul
RT: 8.510 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:108 Resp: 387205
Ion Ratio Lower Upper
108 100
107 111.5 90.3 135.5



#19
Hexachloroethane
Concen: 20.287 ng/ul
RT: 8.792 min Scan# 987
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:117 Resp: 157915
Ion Ratio Lower Upper
117 100
201 93.7 81.8 122.6
199 59.1 51.6 77.4





#22

Nitrobenzene

Concen: 23.066 ng/ul

RT: 8.933 min Scan# 1101

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion: 77 Resp: 438790

Ion Ratio Lower Upper

77 100

123 46.1 37.4 56.0

65 13.1 10.8 16.2



#23

Isophorone

Concen: 24.713 ng/ul

RT: 9.463 min Scan# 1101

Delta R.T. -0.000 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37




Tgt Ion: 82 Resp: 834846


Ion Ratio Lower Upper

82 100

95 6.7 5.5 8.3

138 17.3 15.3 22.9









#25
2-Nitrophenol
Concen: 23.872 ng/ul
RT: 9.639 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:139 Resp: 216596
Ion Ratio Lower Upper
139 100
65 54.5 39.8 59.6
109 29.7 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 23.486 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:107 Resp: 426155
Ion Ratio Lower Upper
107 100
121 50.5 42.7 64.1
122 86.9 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 23.327 ng/ul

RT: 9.945 min Scan# 1183

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion: 93 Resp: 513017

Ion Ratio Lower Upper

93 100

95 31.9 26.1 39.1

123 11.5 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 24.175 ng/ul

RT: 10.175 min Scan# 1222

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37




Tgt Ion: 162 Resp: 359904


Ion Ratio Lower Upper

162 100

164 66.3 51.8 77.6

98 35.9 29.5 44.3









#30
Naphthalene
Concen: 19.519 ng/ul
RT: 10.569 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:128 Resp: 1084848
Ion Ratio Lower Upper
128 100
129 11.2 8.8 13.2
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 20.451 ng/ul
RT: 10.692 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:127 Resp: 469236
Ion Ratio Lower Upper
127 100
129 31.5 25.3 37.9









#33
Hexachlorobutadiene
Concen: 20.953 ng/ul
RT: 10.845 min Scan# 1336
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:225 Resp: 195968
Ion Ratio Lower Upper
225 100
223 62.5 50.0 75.0
227 64.2 49.9 74.9



#34
Caprolactam
Concen: 15.686 ng/ul
RT: 11.486 min Scan# 1445
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:113 Resp: 69969
Ion Ratio Lower Upper
113 100
55 171.6 150.2 225.4
56 124.5 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 22.933 ng/ul
RT: 11.827 min Scan# 1503
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:107 Resp: 345206
Ion Ratio Lower Upper
107 100
144 26.1 20.3 30.5
142 81.0 64.4 96.6



#36
2-Methylnaphthalene
Concen: 23.603 ng/ul
RT: 12.186 min Scan# 1564
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:142 Resp: 823066
Ion Ratio Lower Upper
142 100
141 89.6 70.7 106.1





#37

1-Methylnaphthalene

Concen: 24.241 ng/ul

RT: 12.410 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion:142 Resp: 851283

Ion Ratio Lower Upper

142 100

141 95.2 75.2 112.8

116 3.7 2.6 3.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.159 ng/ul

RT: 12.557 min Scan# 1627

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Tgt Ion:216 Resp: 409540




Ion Ratio Lower Upper


216 100

214 79.6 62.4 93.6

179 21.4 17.3 25.9


108 20.1 14.0 21.0








#40
Hexachlorocyclopentadiene
Concen: 16.791 ng/ul
RT: 12.527 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37


Instrument :
BNA_M
ClientSampleId :
SSTD020169





Tgt Ion:237 Resp: 212926
Ion Ratio Lower Upper
237 100
235 60.3 48.4 72.6
272 12.5 11.6 17.4




#41
2,4,6-Trichlorophenol
Concen: 20.640 ng/ul
RT: 12.810 min Scan# 1670
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37



Tgt Ion:196 Resp: 277620
Ion Ratio Lower Upper
196 100
198 96.2 77.8 116.6
200 30.7 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.469 ng/ul
RT: 12.880 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:196 Resp: 292039
Ion Ratio Lower Upper
196 100
198 98.9 76.8 115.2
200 31.3 24.9 37.3



#43
1,1'-Biphenyl
Concen: 19.255 ng/ul
RT: 13.210 min Scan# 1738
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:154 Resp: 1119965
Ion Ratio Lower Upper
154 100
153 40.0 32.4 48.6
76 12.8 10.1 15.1





#44

2-Chloronaphthalene

Concen: 19.336 ng/ul

RT: 13.245 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion:162 Resp: 863044

Ion Ratio Lower Upper

162 100

164 32.2 25.4 38.0

127 39.5 31.6 47.4



#45

2-Nitroaniline

Concen: 21.654 ng/ul

RT: 13.468 min Scan# 1782

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37




Tgt Ion: 65 Resp: 261713


Ion Ratio Lower Upper

65 100

92 67.4 54.5 81.7

138 108.0 91.2 136.8









#47
Dimethylphthalate
Concen: 20.600 ng/ul
RT: 13.845 min Scan# 1846
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:163 Resp: 1037461
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.239 ng/ul
RT: 13.968 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:165 Resp: 211316
Ion Ratio Lower Upper
165 100
89 62.3 42.5 63.7
121 22.5 17.2 25.8









#50
Acenaphthylene
Concen: 19.887 ng/ul
RT: 14.092 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:152 Resp: 1431087
Ion Ratio Lower Upper
152 100
151 20.4 16.1 24.1
153 13.5 10.6 15.8



#51
3-Nitroaniline
Concen: 21.776 ng/ul
RT: 14.298 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:138 Resp: 226293
Ion Ratio Lower Upper
138 100
108 9.3 8.1 12.1
92 112.9 90.4 135.6









#52
Acenaphthene
Concen: 19.715 ng/ul
RT: 14.439 min Scan# 1946
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:153 Resp: 924699
Ion Ratio Lower Upper
153 100
152 48.2 37.1 55.7
154 87.3 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 20.241 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:184 Resp: 104818
Ion Ratio Lower Upper
184 100
63 74.4 54.2 81.4
154 65.2 49.3 73.9









#55
4-Nitrophenol
Concen: 21.186 ng/ul
RT: 14.615 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:109 Resp: 155185
Ion Ratio Lower Upper
109 100
139 132.1 124.9 187.3
65 131.9 122.2 183.4



#56
Dibenzofuran
Concen: 19.786 ng/ul
RT: 14.774 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:168 Resp: 1287702
Ion Ratio Lower Upper
168 100
139 39.5 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 22.422 ng/ul
RT: 14.757 min Scan# 2001
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:165 Resp: 302162
Ion Ratio Lower Upper
165 100
63 53.2 35.9 53.9
182 2.5 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.805 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:232 Resp: 236203
Ion Ratio Lower Upper
232 100
131 52.6 36.7 55.1
130 2.6 1.7 2.5#
166 32.8 24.4 36.6









#59
Diethylphthalate
Concen: 21.709 ng/ul
RT: 15.204 min Scan# 2077
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:149 Resp: 1077711
Ion Ratio Lower Upper
149 100
177 20.4 16.8 25.2
150 12.3 10.2 15.4



#61
Fluorene
Concen: 20.353 ng/ul
RT: 15.421 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:166 Resp: 1050623
Ion Ratio Lower Upper
166 100
165 99.2 77.2 115.8
167 13.4 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 20.811 ng/ul
RT: 15.421 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:204 Resp: 501814
Ion Ratio Lower Upper
204 100
206 32.4 25.9 38.9
141 61.7 50.6 75.8



#63
4-Nitroaniline
Concen: 22.354 ng/ul
RT: 15.462 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:138 Resp: 216695
Ion Ratio Lower Upper
138 100
92 51.2 39.2 58.8
108 81.5 51.6 77.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 18.686 ng/ul
RT: 15.515 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:198 Resp: 151993
Ion Ratio Lower Upper
198 100
182 4.4 4.2 6.4
77 30.5 19.9 29.9#



#67
N-Nitrosodiphenylamine
Concen: 19.398 ng/ul
RT: 15.639 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:169 Resp: 867809
Ion Ratio Lower Upper
169 100
168 68.0 53.8 80.6
167 38.0 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 19.890 ng/ul
RT: 16.309 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:248 Resp: 288206
Ion Ratio Lower Upper
248 100
250 98.1 77.3 115.9
141 82.6 59.8 89.8



#69
Hexachlorobenzene
Concen: 19.670 ng/ul
RT: 16.415 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:284 Resp: 345390
Ion Ratio Lower Upper
284 100
142 40.4 28.6 42.8
249 30.5 24.6 37.0









#70
Atrazine
Concen: 20.358 ng/ul
RT: 16.598 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:200 Resp: 311196
Ion Ratio Lower Upper
200 100
173 27.0 21.3 31.9
215 46.8 39.0 58.6



#71
Pentachlorophenol
Concen: 21.563 ng/ul
RT: 16.774 min Scan# 2344
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:266 Resp: 208743
Ion Ratio Lower Upper
266 100
264 61.6 49.5 74.3
268 63.0 51.0 76.4









#72
Phenanthrene
Concen: 19.629 ng/ul
RT: 17.162 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:178 Resp: 1614816
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 19.7 14.9 22.3



#74
Anthracene
Concen: 20.050 ng/ul
RT: 17.251 min Scan# 2425
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:178 Resp: 1653647
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.2
176 19.0 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.971 ng/uL
RT: 13.168 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:216 Resp: 421947
Ion Ratio Lower Upper
216 100
214 77.8 60.9 91.3
218 49.2 37.5 56.3



#76
Pentachlorobenzene
Concen: 18.835 ng/uL
RT: 14.686 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:250 Resp: 412927
Ion Ratio Lower Upper
250 100
248 63.6 50.4 75.6
252 65.8 51.1 76.7









#77
 Carbazole
 Concen: 19.563 ng/ul
 RT: 17.527 min Scan# 2472
 Delta R.T. -0.006 min
 Lab File: BM036523.D
 Acq: 07 Sep 2022 23:37




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020169


Tgt Ion:167 Resp: 1510656
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.4 26.0
 139 13.6 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 22.522 ng/ul
 RT: 18.092 min Scan# 2568
 Delta R.T. -0.006 min
 Lab File: BM036523.D
 Acq: 07 Sep 2022 23:37

Tgt Ion:149 Resp: 1902263
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 5.3 3.7 5.5





#80

Fluoranthene

Concen: 20.926 ng/ul

RT: 19.180 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion:202 Resp: 1870074

Ion Ratio Lower Upper

202 100

101 11.9 10.1 15.1

100 9.2 7.8 11.6



#82

Pyrene

Concen: 20.396 ng/ul

RT: 19.545 min Scan# 2815

Delta R.T. -0.000 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37




Tgt Ion:202 Resp: 1902039


Ion Ratio Lower Upper

202 100

101 13.8 11.9 17.9

100 11.5 10.2 15.4









#83
Butylbenzylphthalate
Concen: 24.453 ng/ul
RT: 20.450 min Scan# 2969
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:149 Resp: 853102
Ion Ratio Lower Upper
149 100
91 73.8 53.2 79.8
206 18.0 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 19.181 ng/ul
RT: 21.233 min Scan# 3102
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:252 Resp: 563024
Ion Ratio Lower Upper
252 100
254 64.2 52.4 78.6
126 12.1 10.8 16.2









#85
Benzo(a)anthracene
Concen: 20.408 ng/ul
RT: 21.291 min Scan# 3112
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:228 Resp: 1748210
Ion Ratio Lower Upper
228 100
229 20.0 15.6 23.4
226 27.1 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 26.811 ng/ul
RT: 21.227 min Scan# 3101
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:149 Resp: 1291460
Ion Ratio Lower Upper
149 100
167 25.9 21.5 32.3
279 3.6 3.4 5.2





#87

Chrysene

Concen: 19.475 ng/ul

RT: 21.344 min Scan# 3121

Delta R.T. -0.000 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Instrument :

BNA_M

ClientSampleId :

SSTD020169





Tgt Ion:228 Resp: 1639221

Ion Ratio Lower Upper

228 100

226 30.7 23.7 35.5

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 28.836 ng/ul




RT: 22.115 min Scan# 3252


Delta R.T. -0.006 min

Lab File: BM036523.D

Acq: 07 Sep 2022 23:37

Tgt Ion:149 Resp: 2008566









#90
Benzo(b)fluoranthene
Concen: 21.904 ng/ul
RT: 22.903 min Scan# 3386
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:252 Resp: 1564781
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 11.3 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 20.630 ng/ul
RT: 22.950 min Scan# 3394
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:252 Resp: 1402156
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 10.9 10.6 15.8









#93
Benzo(a)pyrene
Concen: 20.143 ng/ul
RT: 23.497 min Scan# 34
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37




Instrument :
BNA_M
ClientSampleId :
SSTD020169


Tgt Ion:252 Resp: 1415680
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 12.8 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 15.542 ng/ul
RT: 25.962 min Scan# 3906
Delta R.T. -0.006 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:276 Resp: 1284297
Ion Ratio Lower Upper
276 100
138 25.9 25.6 38.4
277 25.7 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 15.620 ng/ul
RT: 25.974 min Scan# 3907
Delta R.T. -0.012 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Instrument :
BNA_M
ClientSampleId :
SSTD020169

Tgt Ion:278 Resp: 1111884
Ion Ratio Lower Upper
278 100
139 16.5 16.2 24.2
279 23.2 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 16.766 ng/ul
RT: 26.685 min Scan# 4029
Delta R.T. -0.012 min
Lab File: BM036523.D
Acq: 07 Sep 2022 23:37

Tgt Ion:276 Resp: 1184748
Ion Ratio Lower Upper
276 100
138 25.0 21.3 31.9
277 24.2 18.6 28.0

