

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090921\
 Data File : BM032116.D
 Acq On : 09 Sep 2021 19:08
 Operator : CG/JU
 Sample : M3670-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SAND-BLAST-MEDIA

Quant Time: Sep 10 03:09:46 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM090821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 08 19:23:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.351	152	63652	20.000	ng	-0.01
21) Naphthalene-d8	10.110	136	234597	20.000	ng	# 0.00
39) Acenaphthene-d10	14.016	164	144225	20.000	ng	0.00
64) Phenanthrene-d10	16.780	188	280882	20.000	ng	0.00
76) Chrysene-d12	20.997	240	258590	20.000	ng	-0.01
86) Perylene-d12	23.091	264	279643	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.999	112	285790	80.117	ng	0.00
7) Phenol-d6	6.557	99	339246	62.513	ng	-0.01
23) Nitrobenzene-d5	8.510	82	439746	83.538	ng	0.00
42) 2,4,6-Tribromophenol	15.527	330	95891	83.661	ng	0.00
45) 2-Fluorobiphenyl	12.622	172	888058	80.734	ng	0.00
79) Terphenyl-d14	19.445	244	1342410	84.701	ng	0.00
Target Compounds						
60) Diethylphthalate	14.880	149	32006	3.148	ng	98
80) Butylbenzylphthalate	20.156	149	1216	3.876	ng	99
84) Bis(2-ethylhexyl)phtha...	20.945	149	5419	3.420	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090921\
Data File : BM032116.D
Acq On : 09 Sep 2021 19:08
Operator : CG/JU
Sample : M3670-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SAND-BLAST-MEDIA

Quant Time: Sep 10 03:09:46 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM090821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 08 19:23:17 2021
Response via : Initial Calibration

