

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090923\
 Data File : BM041766.D
 Acq On : 10 Sep 2023 01:33
 Operator : MA/JU
 Sample : 04247-14
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBJ10

Quant Time: Sep 10 04:51:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM090523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 09 23:26:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.997	152	440	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.808	136	940	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.630	164	425	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.380	188	796	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.546	240	260	0.400	ng/ul	# 0.00
23) Perylene-d12	23.969	264	295	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	2369	3.523	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.386	152	381	0.315	ng/ul	0.00
18) Fluoranthene-d10	19.400	212	482	0.382	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.830	128	574	0.186	ng/ul#	1
7) 2-Methylnaphthalene	12.458	142	927	0.529	ng/ul	97
8) 1-Methylnaphthalene	12.678	142	3526	2.003	ng/ul	94
10) Acenaphthylene	14.348	152	675	0.257	ng/ul#	1
11) Acenaphthene	14.699	153	3086	1.511	ng/ul#	78
12) Fluorene	15.680	166	4009	1.750	ng/ul#	91
15) Phenanthrene	17.422	178	13495	4.015	ng/ul	97
16) Anthracene	17.553	178	3387	1.176	ng/ul#	1
19) Fluoranthene	19.428	202	563	0.183	ng/ul#	61
20) Pyrene	19.795	202	2974	0.976	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM090923\
Data File : BM041766.D
Acq On : 10 Sep 2023 01:33
Operator : MA/JU
Sample : 04247-14
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBJ10

Quant Time: Sep 10 04:51:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM090523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 09 23:26:58 2023
Response via : Initial Calibration

