

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091019\
 Data File : BM022651.D
 Acq On : 11 Sep 2019 11:25
 Operator : HP/JU
 Sample : K4706-03
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BF569

Manual Integrations
 APPROVED

mohammad
 9/16/2019 8:54:42 AM

Quant Time: Sep 11 15:05:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 10 15:06:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	18915	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	81260	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.48	164	47279	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	108297m	0.40	ng/ul	0.00
16) Chrysene-d12	21.39	240	91748m	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	83553	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.23	152	37342	0.28	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	97393m	0.33	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.69	128	656323	2.835	ng/ul	100
5) 2-Methylnaphthalene	12.30	142	98675	0.622	ng/ul	100
8) Acenaphthene	14.54	153	19735	0.108	ng/ul	98
9) Fluorene	15.52	166	14321	0.068	ng/ul	99
12) Phenanthrene	17.25	178	22847	0.066	ng/ul	98
13) Anthracene	17.34	178	8690m	0.029	ng/ul	
15) Fluoranthene	19.27	202	9552m	0.026	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091019\
Data File : BM022651.D
Acq On : 11 Sep 2019 11:25
Operator : HP/JU
Sample : K4706-03
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BF569

Manual Integrations
APPROVED

mohammad
9/16/2019 8:54:42 AM

Quant Time: Sep 11 15:05:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 10 15:06:45 2019
Response via : Initial Calibration

