

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091019\
 Data File : BM022652.D
 Acq On : 11 Sep 2019 12:02
 Operator : HP/JU
 Sample : K4706-04
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BF570

Manual Integrations
 APPROVED

mohammad
 9/16/2019 8:54:45 AM

Quant Time: Sep 11 15:09:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 10 15:06:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	20018	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	86637	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.48	164	51176	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	114693m	0.40	ng/ul	0.00
16) Chrysene-d12	21.39	240	95649m	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	76624m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.23	152	39474	0.28	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	101427m	0.32	ng/ul	0.00
Target Compounds						
					Ovalue	
3) Naphthalene	10.69	128	113787	0.461	ng/ul	100
5) 2-Methylnaphthalene	12.30	142	33247	0.196	ng/ul	100
8) Acenaphthene	14.54	153	16979	0.086	ng/ul	99
9) Fluorene	15.52	166	17364	0.076	ng/ul	100
12) Phenanthrene	17.25	178	65243	0.179	ng/ul	99
15) Fluoranthene	19.27	202	15426m	0.039	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091019\
Data File : BM022652.D
Acq On : 11 Sep 2019 12:02
Operator : HP/JU
Sample : K4706-04
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BF570

Manual Integrations
APPROVED

mohammad
9/16/2019 8:54:45 AM

Quant Time: Sep 11 15:09:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 10 15:06:45 2019
Response via : Initial Calibration

