

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091019\
Data File : BM022647.D
Acq On : 11 Sep 2019 08:22
Operator : HP/JU
Sample : PB122840BL
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK40

Manual Integrations
APPROVED

mohammad
9/16/2019 8:54:28 AM

Quant Time: Sep 11 08:58:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 11 07:37:53 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	17163	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	71463	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.48	164	38713	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	85731m	0.40	ng/ul	0.00
16) Chrysene-d12	21.39	240	69475m	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	68201	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.23	152	32899	0.28	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	66679m	0.28	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091019\
Data File : BM022647.D
Acq On : 11 Sep 2019 08:22
Operator : HP/JU
Sample : PB122840BL
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK40

Manual Integrations
APPROVED

mohammad
9/16/2019 8:54:28 AM

Quant Time: Sep 11 08:58:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 11 07:37:53 2019
Response via : Initial Calibration

