

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM091118\
 Data File : BM016740.D
 Acq On : 12 Sep 2018 01:44
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02057

Quant Time: Sep 12 02:42:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM091018MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 10 15:28:07 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.487	0.480	1.4	108	0.00
3 S	1,4-Dioxane-d8	0.468	0.452	3.4	110	0.00
4	Benzaldehyde	0.949	1.075	-13.3	99	0.00
5 S	Phenol-d5	1.687	1.627	3.6	101	0.00
6	Phenol	1.684	1.648	2.1	103	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.045	1.009	3.4	100	0.00
8	Bis(2-Chloroethyl)ether	1.296	1.270	2.0	101	0.00
9 S	2-Chlorophenol-d4	1.377	1.337	2.9	101	0.00
10	2-Chlorophenol	1.383	1.370	0.9	103	0.00
11	2-Methylphenol	1.262	1.222	3.2	102	0.00
12	2,2'-oxybis(1-Chloropropane	2.050	1.945	5.1	99	0.00
13 S	4-Methylphenol-d8	1.305	1.287	1.4	104	0.00
14	Acetophenone	2.023	2.016	0.3	104	0.00
15 P	N-Nitroso-di-n-propylamine	1.012	0.979	3.3	101	0.00
16	4-Methylphenol	1.334	1.318	1.2	103	0.00
17	Hexachloroethane	0.543	0.530	2.4	102	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
19 S	Nitrobenzene-d5	0.131	0.139	-6.1	111	0.00
20	Nitrobenzene	0.332	0.337	-1.5	106	0.00
21	Isophorone	0.661	0.620	6.2	99	0.00
22 S	2-Nitrophenol-d4	0.119	0.124	-4.2	111	0.00
23 C	2-Nitrophenol	0.138	0.147	-6.5	111	0.00
24	2,4-Dimethylphenol	0.356	0.348	2.2	103	0.00
25	Bis(2-Chloroethoxy)methane	0.408	0.402	1.5	105	0.00
26 S	2,4-Dichlorophenol-d3	0.304	0.302	0.7	105	0.00
27 C	2,4-Dichlorophenol	0.292	0.292	0.0	105	0.00
28	Naphthalene	1.028	1.016	1.2	104	0.00
29 S	4-Chloroaniline-d4	0.367	0.379	-3.3	100	0.00
30	4-Chloroaniline	0.367	0.381	-3.8	100	0.00
31 C	Hexachlorobutadiene	0.197	0.196	0.5	104	0.00
32	Caprolactam	0.092	0.086	6.5	105	0.00
33 C	4-Chloro-3-methylphenol	0.306	0.305	0.3	105	0.00
34	2-Methylnaphthalene	0.725	0.724	0.1	105	0.00
35	1-Methylnaphthalene	0.725	0.724	0.1	105	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
37	1,2,4,5-Tetrachlorobenzene	0.687	0.671	2.3	107	0.00
38	Hexachlorocyclopentadiene	0.428	0.403	5.8	108	0.00
39 C	2,4,6-Trichlorophenol	0.391	0.389	0.5	109	0.00
40	2,4,5-Trichlorophenol	0.427	0.426	0.2	111	0.00
41	1,1'-Biphenyl	1.664	1.646	1.1	109	0.00
42	2-Chloronaphthalene	1.302	1.297	0.4	110	0.00
43	2-Nitroaniline	0.287	0.317	-10.5	121	0.00
44 S	Dimethylphthalate-d6	1.606	1.605	0.1	112	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM091118\
 Data File : BM016740.D
 Acq On : 12 Sep 2018 01:44
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02057

Quant Time: Sep 12 02:42:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM091018MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 10 15:28:07 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.569	1.560	0.6	112	0.00
46	2,6-Dinitrotoluene	0.241	0.278	-15.4	127	0.00
47 S	Acenaphthylene-d8	2.074	2.055	0.9	109	0.00
48	Acenaphthylene	1.963	1.969	-0.3	110	0.00
49	3-Nitroaniline	0.262	0.280	-6.9	122	0.00
50 C	Acenaphthene	1.394	1.397	-0.2	111	0.00
51	2,4-Dinitrophenol	0.108	0.109	-0.9	148	0.00
52 S	4-Nitrophenol-d4	0.264	0.274	-3.8	130	0.00
53	4-Nitrophenol	0.203	0.208	-2.5	119	0.00
54	Dibenzofuran	1.912	1.923	-0.6	112	0.00
55	2,4-Dinitrotoluene	0.354	0.418	-18.1	129	0.00
56	2,3,4,6-Tetrachlorophenol	0.345	0.353	-2.3	114	0.00
57	Diethylphthalate	1.555	1.558	-0.2	112	0.00
58 S	Fluorene-d10	1.402	1.427	-1.8	113	0.00
59	Fluorene	1.571	1.588	-1.1	113	0.00
60	4-Chlorophenyl-phenylether	0.797	0.820	-2.9	115	0.00
61	4-Nitroaniline	0.322	0.346	-7.5	124	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.082	0.082	0.0	147	0.00
64	4,6-Dinitro-2-methylphenol	0.092	0.090	2.2	138	0.00
65	N-Nitrosodiphenylamine	0.619	0.611	1.3	115	0.00
66	4-Bromophenyl-phenylether	0.226	0.222	1.8	115	0.00
67	Hexachlorobenzene	0.244	0.242	0.8	117	0.00
68	Atrazine	0.219	0.210	4.1	114	0.00
69 C	Pentachlorophenol	0.132	0.121	8.3	118	0.00
70	Phenanthrene	1.111	1.119	-0.7	118	0.00
71 S	Anthracene-d10	0.965	0.971	-0.6	118	0.00
72	Anthracene	1.141	1.151	-0.9	120	0.00
73	1,2,3,4-Tetrachlorobenzene	0.331	0.315	4.8	109	0.00
74	Pentachlorobenzene	0.299	0.296	1.0	115	0.00
75	Carbazole	1.003	1.008	-0.5	119	0.00
76	Di-n-butylphthalate	1.167	1.137	2.6	115	0.00
77 C	Fluoranthene	1.273	1.312	-3.1	123	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
79 S	Pyrene-d10	0.970	0.924	4.7	124	0.00
80	Pyrene	1.228	1.180	3.9	125	0.00
81	Butylbenzylphthalate	0.454	0.414	8.8	121	0.00
82	3,3'-Dichlorobenzidine	0.396	0.368	7.1	119	0.00
83	Benzo(a)anthracene	1.238	1.226	1.0	132	0.00
84	Bis(2-ethylhexyl)phthalate	0.714	0.655	8.3	123	0.00
85	Chrysene	1.155	1.148	0.6	131	0.00
86 I	Perylene-d12	1.000	1.000	0.0	134	0.00
87	Di-n-octyl phthalate	1.202	1.031	14.2	119	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM091118\
Data File : BM016740.D
Acq On : 12 Sep 2018 01:44
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02057

Quant Time: Sep 12 02:42:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM091018MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 10 15:28:07 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.197	1.155	3.5	130	0.00
89	Benzo(k)fluoranthene	1.149	1.181	-2.8	137	0.00
90 S	Benzo(a)pyrene-d12	1.043	1.022	2.0	132	0.00
91 C	Benzo(a)pyrene	1.141	1.138	0.3	133	0.00
92	Indeno(1,2,3-cd)pyrene	1.361	1.347	1.0	133	0.00
93	Dibenzo(a,h)anthracene	1.139	1.142	-0.3	135	0.00
94	Benzo(g,h,i)perylene	1.129	1.114	1.3	133	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0