

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091219\
 Data File : BM022672.D
 Acq On : 12 Sep 2019 19:01
 Operator : HP/JU
 Sample : PB122880BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK80

Manual Integrations
 APPROVED

mohammad
 9/16/2019 8:59:04 AM

Quant Time: Sep 13 01:24:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 10 15:06:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	9744	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	41688	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.48	164	23393	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	53729m	0.40	ng/ul	0.00
16) Chrysene-d12	21.39	240	42981m	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	37209m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.23	152	20666	0.30	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	48425m	0.33	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM091219\
Data File : BM022672.D
Acq On : 12 Sep 2019 19:01
Operator : HP/JU
Sample : PB122880BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK80

Manual Integrations
APPROVED

mohammad
9/16/2019 8:59:04 AM

Quant Time: Sep 13 01:24:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 10 15:06:45 2019
Response via : Initial Calibration

