

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091323\
 Data File : BM041830.D
 Acq On : 13 Sep 2023 14:14
 Operator : MA/JU
 Sample : SSTD3.235
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2027

Quant Time: Sep 13 16:12:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM091323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 13 14:45:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	3639	0.400	ng/ul	0.00
4) Naphthalene-d8	10.790	136	9447	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.615	164	4796	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.367	188	10033	0.400	ng/ul	0.00
17) Chrysene-d12	21.539	240	6557	0.400	ng/ul	0.00
23) Perylene-d12	23.956	264	7142	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	14847	3.208	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.414	88	15622	3.216	ng/ul#	67
14) Pentachlorophenol	16.995	266	20095	3.516	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091323\
Data File : BM041830.D
Acq On : 13 Sep 2023 14:14
Operator : MA/JU
Sample : SSTD3.235
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2027

Quant Time: Sep 13 16:12:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM091323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 13 14:45:01 2023
Response via : Initial Calibration

