

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091622\
 Data File : BM036571.D
 Acq On : 16 Sep 2022 20:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020175

Quant Time: Sep 16 23:02:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	414044	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	1844363	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	1167310	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	2349973	20.000	ng/u1	0.00
79) Chrysene-d12	21.550	240	2072401	20.000	ng/u1	0.00
88) Perylene-d12	23.991	264	1720530	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	81629	7.391	ng/uL	0.00
4) Pyridine-d5	3.816	84	586892	18.220	ng/u1	0.00
7) Phenol-d5	7.175	99	687728	18.587	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	435342	19.374	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	534819	20.077	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	535070	19.571	ng/u1	0.00
21) Nitrobenzene-d5	9.192	128	257427	20.621	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	239516	21.420	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	519867	20.061	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	808030	18.754	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1542221	19.589	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1825789	19.551	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	279788	18.323	ng/u1	0.00
60) Fluorene-d10	15.639	176	1376096	19.504	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	179656	19.323	ng/u1	0.00
73) Anthracene-d10	17.486	188	2086179	19.416	ng/u1	0.00
81) Pyrene-d10	19.762	212	2233127	21.903	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	1673189	19.905	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	85180	7.159	ng/uL	95
5) Pyridine	3.834	79	594068	18.357	ng/u1	95
6) Benzaldehyde	7.157	77	387624	21.634	ng/u1	96
8) Phenol	7.198	94	742873	18.707	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.446	93	605344	19.592	ng/u1	98
12) 2-Chlorophenol	7.587	128	577568	20.043	ng/u1	98
13) 2-Methylphenol	8.463	108	557115	19.086	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.563	45	717729	19.599	ng/u1	99
16) Acetophenone	8.851	105	941066	19.844	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.840	70	465015	20.322	ng/u1	97
18) 4-Methylphenol	8.792	108	609596	19.313	ng/u1	99
19) Hexachloroethane	9.116	117	234168	20.063	ng/u1	96
22) Nitrobenzene	9.234	77	669307	20.194	ng/u1	97
23) Isophorone	9.763	82	1322824	19.574	ng/u1	100
25) 2-Nitrophenol	9.945	139	287582	20.823	ng/u1	98
26) 2,4-Dimethylphenol	10.004	107	675729	19.837	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.245	93	796586	19.776	ng/u1	99
29) 2,4-Dichlorophenol	10.481	162	538681	19.773	ng/u1	100
30) Naphthalene	10.892	128	1951856	19.426	ng/u1	100
32) 4-Chloroaniline	10.992	127	831504	18.693	ng/u1	100
33) Hexachlorobutadiene	11.181	225	325193	19.517	ng/u1	100
34) Caprolactam	11.751	113	97199	10.999	ng/u1	100
35) 4-Chloro-3-methylphenol	12.104	107	590886	19.649	ng/u1	99
36) 2-Methylnaphthalene	12.492	142	1312613	19.608	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091622\
 Data File : BM036571.D
 Acq On : 16 Sep 2022 20:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020175

Quant Time: Sep 16 23:02:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.710	142	1313586	19.519	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.851	216	642698	19.452	ng/ul	99
40) Hexachlorocyclopentadiene	12.839	237	339487	19.716	ng/ul	98
41) 2,4,6-Trichlorophenol	13.086	196	404609	20.001	ng/ul	99
42) 2,4,5-Trichlorophenol	13.157	196	441275	20.345	ng/ul	98
43) 1,1'-Biphenyl	13.492	154	1691253	19.654	ng/ul	99
44) 2-Chloronaphthalene	13.533	162	1322643	19.588	ng/ul	99
45) 2-Nitroaniline	13.727	65	356309	21.439	ng/ul	98
47) Dimethylphthalate	14.110	163	1640386	19.642	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	293648	21.627	ng/ul	99
50) Acenaphthylene	14.374	152	2075078	19.435	ng/ul	99
51) 3-Nitroaniline	14.545	138	320296	19.669	ng/ul#	98
52) Acenaphthene	14.716	153	1443706	19.325	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	118410	19.654	ng/ul	94
55) 4-Nitrophenol	14.839	109	242627	18.587	ng/ul	98
56) Dibenzofuran	15.045	168	1954798	19.484	ng/ul	98
57) 2,4-Dinitrotoluene	15.004	165	431314	21.851	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.269	232	361639	19.919	ng/ul	100
59) Diethylphthalate	15.469	149	1644257	19.668	ng/ul	99
61) Fluorene	15.692	166	1664680	19.482	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.686	204	803001	19.650	ng/ul	98
63) 4-Nitroaniline	15.704	138	310116	19.601	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.763	198	198258	18.896	ng/ul	97
67) N-Nitrosodiphenylamine	15.898	169	1367775	19.762	ng/ul	99
68) 4-Bromophenyl-phenylether	16.580	248	417701	18.034	ng/ul	97
69) Hexachlorobenzene	16.692	284	524758	19.155	ng/ul	99
70) Atrazine	16.845	200	456888	18.707	ng/ul	98
71) Pentachlorophenol	17.027	266	288725	17.742	ng/ul	98
72) Phenanthrene	17.427	178	2530195	19.547	ng/ul	100
74) Anthracene	17.515	178	2558560	19.447	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.457	216	636627	19.807	ng/ul	99
76) Pentachlorobenzene	14.963	250	685502	19.719	ng/ul	99
77) Carbazole	17.786	167	2270393	18.639	ng/ul	100
78) Di-n-butylphthalate	18.357	149	2711740	19.641	ng/ul	100
80) Fluoranthene	19.433	202	2762389	22.111	ng/ul	99
82) Pyrene	19.792	202	2909968	22.004	ng/ul	97
83) Butylbenzylphthalate	20.692	149	1041663	20.787	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.468	252	805519	19.893	ng/ul	98
85) Benzo(a)anthracene	21.533	228	2600597	20.028	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.468	149	1587811	21.691	ng/ul	99
87) Chrysene	21.592	228	2407852	19.536	ng/ul	100
89) Di-n-octyl phthalate	22.415	149	2339057	22.550	ng/ul	100
90) Benzo(b)fluoranthene	23.250	252	2267075	20.779	ng/ul	99
91) Benzo(k)fluoranthene	23.297	252	2173294	19.935	ng/ul	99
93) Benzo(a)pyrene	23.886	252	1903994	19.851	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.544	276	2047402	18.279	ng/ul	97
95) Dibenzo(a,h)anthracene	26.562	278	1869043	18.400	ng/ul	98
96) Benzo(g,h,i)perylene	27.321	276	1790393	18.438	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091622\
 Data File : BM036571.D
 Acq On : 16 Sep 2022 20:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020175

Quant Time: Sep 16 23:02:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.159 ng/uL
RT: 3.428 min Scan# 58
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion: 88 Resp: 85180
Ion Ratio Lower Upper
88 100
43 26.2 23.1 34.7
58 67.1 56.6 85.0



#5
Pyridine
Concen: 18.357 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion: 79 Resp: 594068
Ion Ratio Lower Upper
79 100
52 65.9 57.2 85.8
51 34.2 26.8 40.2









#6
Benzaldehyde
Concen: 21.634 ng/ul
RT: 7.157 min Scan# 692
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion: 77 Resp: 387624
Ion Ratio Lower Upper
77 100
105 98.3 75.1 112.7
106 94.9 73.0 109.6



#8
Phenol
Concen: 18.707 ng/ul
RT: 7.198 min Scan# 699
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion: 94 Resp: 742873
Ion Ratio Lower Upper
94 100
65 28.0 22.2 33.4
66 36.8 28.9 43.3









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.592 ng/ul
 RT: 7.446 min Scan# 741
 Delta R.T. 0.000 min
 Lab File: BM036571.D
 Acq: 16 Sep 2022 20:09

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020175

Tgt Ion: 93 Resp: 605344




Ion	Ratio	Lower	Upper
93	100		
63	72.3	59.1	88.7
95	31.7	26.1	39.1




#12
 2-Chlorophenol
 Concen: 20.043 ng/ul
 RT: 7.587 min Scan# 765
 Delta R.T. 0.000 min
 Lab File: BM036571.D
 Acq: 16 Sep 2022 20:09

Tgt Ion:128 Resp: 577568

Ion	Ratio	Lower	Upper
128	100		
64	45.8	37.8	56.8
130	32.8	26.0	39.0





#13

2-Methylphenol

Concen: 19.086 ng/ul

RT: 8.463 min Scan# 914

Delta R.T. 0.000 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :





SSTD020175

Tgt Ion:108 Resp: 557115

Ion Ratio Lower Upper

108 100

107 86.0 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.599 ng/ul

RT: 8.563 min Scan# 931

Delta R.T. 0.000 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09




Tgt Ion: 45 Resp: 717729


Ion Ratio Lower Upper

45 100

77 16.5 13.1 19.7

79 12.7 9.5 14.3









#16
Acetophenone
Concen: 19.844 ng/ul
RT: 8.851 min Scan# 980
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:105 Resp: 941066
Ion Ratio Lower Upper
105 100
77 77.5 61.6 92.4
51 25.9 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.322 ng/ul
RT: 8.840 min Scan# 978
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion: 70 Resp: 465015
Ion Ratio Lower Upper
70 100
42 51.7 43.0 64.4
101 9.5 7.7 11.5
130 20.6 15.1 22.7









#18
4-Methylphenol
Concen: 19.313 ng/ul
RT: 8.792 min Scan# 970
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:108 Resp: 609596
Ion Ratio Lower Upper
108 100
107 111.1 89.4 134.0



#19
Hexachloroethane
Concen: 20.063 ng/ul
RT: 9.116 min Scan# 1025
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:117 Resp: 234168
Ion Ratio Lower Upper
117 100
201 89.6 68.7 103.1
199 57.4 43.0 64.6









#22
Nitrobenzene
Concen: 20.194 ng/ul
RT: 9.234 min Scan# 1045
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion: 77 Resp: 669307
Ion Ratio Lower Upper
77 100
123 45.7 34.9 52.3
65 13.3 10.7 16.1



#23
Isophorone
Concen: 19.574 ng/ul
RT: 9.763 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion: 82 Resp: 1322824
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.2
138 16.9 13.5 20.3









#25
2-Nitrophenol
Concen: 20.823 ng/ul
RT: 9.945 min Scan# 1166
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:139 Resp: 287582
Ion Ratio Lower Upper
139 100
65 52.1 40.9 61.3
109 31.0 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 19.837 ng/ul
RT: 10.004 min Scan# 1176
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:107 Resp: 675729
Ion Ratio Lower Upper
107 100
121 49.3 39.4 59.2
122 86.1 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.776 ng/ul
 RT: 10.245 min Scan# 1217
 Delta R.T. -0.006 min
 Lab File: BM036571.D
 Acq: 16 Sep 2022 20:09




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020175


Tgt Ion: 93 Resp: 796586
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.2 39.4
 123 11.2 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 19.773 ng/ul
 RT: 10.481 min Scan# 1257
 Delta R.T. 0.000 min
 Lab File: BM036571.D
 Acq: 16 Sep 2022 20:09

Tgt Ion:162 Resp: 538681
 Ion Ratio Lower Upper
 162 100
 164 65.6 52.1 78.1
 98 36.4 29.1 43.7









#30
Naphthalene
Concen: 19.426 ng/ul
RT: 10.892 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:128 Resp: 1951856
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 18.693 ng/ul
RT: 10.992 min Scan# 1344
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:127 Resp: 831504
Ion Ratio Lower Upper
127 100
129 32.1 25.4 38.2









#33
Hexachlorobutadiene
Concen: 19.517 ng/ul
RT: 11.181 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:225 Resp: 325193
Ion Ratio Lower Upper
225 100
223 62.2 49.8 74.8
227 63.5 51.0 76.6



#34
Caprolactam
Concen: 10.999 ng/ul
RT: 11.751 min Scan# 1473
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:113 Resp: 97199
Ion Ratio Lower Upper
113 100
55 159.4 127.5 191.3
56 118.2 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 19.649 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:107 Resp: 590886
Ion Ratio Lower Upper
107 100
144 25.4 21.5 32.3
142 79.6 63.9 95.9



#36
2-Methylnaphthalene
Concen: 19.608 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:142 Resp: 1312613
Ion Ratio Lower Upper
142 100
141 87.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 19.519 ng/ul

RT: 12.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :

SSTD020175





Tgt Ion:142 Resp: 1313586

Ion Ratio Lower Upper

142 100

141 91.3 73.4 110.2

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.452 ng/ul

RT: 12.851 min Scan# 1660

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Tgt Ion:216 Resp: 642698




Ion Ratio Lower Upper


216 100

214 78.7 62.4 93.6

179 21.6 16.7 25.1

108 20.5 15.4 23.0









#40
Hexachlorocyclopentadiene
Concen: 19.716 ng/ul
RT: 12.839 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:237 Resp: 339487
Ion Ratio Lower Upper
237 100
235 62.3 50.9 76.3
272 14.0 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 20.001 ng/ul
RT: 13.086 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:196 Resp: 404609
Ion Ratio Lower Upper
196 100
198 95.9 77.3 115.9
200 30.6 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 20.345 ng/ul
RT: 13.157 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:196 Resp: 441275
Ion Ratio Lower Upper
196 100
198 96.9 75.6 113.4
200 30.3 24.7 37.1



#43
1,1'-Biphenyl
Concen: 19.654 ng/ul
RT: 13.492 min Scan# 1769
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:154 Resp: 1691253
Ion Ratio Lower Upper
154 100
153 39.9 31.4 47.2
76 13.0 10.4 15.6









#44
2-Chloronaphthalene
Concen: 19.588 ng/ul
RT: 13.533 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:162 Resp: 1322643
Ion Ratio Lower Upper
162 100
164 33.0 26.2 39.2
127 39.8 32.6 49.0



#45
2-Nitroaniline
Concen: 21.439 ng/ul
RT: 13.727 min Scan# 1809
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion: 65 Resp: 356309
Ion Ratio Lower Upper
65 100
92 69.3 54.4 81.6
138 104.8 86.1 129.1









#47
Dimethylphthalate
Concen: 19.642 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:163 Resp: 1640386
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 9.9 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 21.627 ng/ul
RT: 14.222 min Scan# 1893
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:165 Resp: 293648
Ion Ratio Lower Upper
165 100
89 66.8 52.8 79.2
121 24.0 19.4 29.0









#50
Acenaphthylene
Concen: 19.435 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:152 Resp: 2075078
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 12.8 10.5 15.7



#51
3-Nitroaniline
Concen: 19.669 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:138 Resp: 320296
Ion Ratio Lower Upper
138 100
108 12.0 12.5 18.7#
92 117.5 95.4 143.2









#52
Acenaphthene
Concen: 19.325 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:153 Resp: 1443706
Ion Ratio Lower Upper
153 100
152 48.8 38.2 57.2
154 88.2 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 19.654 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:184 Resp: 118410
Ion Ratio Lower Upper
184 100
63 68.4 58.8 88.2
154 63.4 55.0 82.4









#55
4-Nitrophenol
Concen: 18.587 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:109 Resp: 242627
Ion Ratio Lower Upper
109 100
139 130.5 102.2 153.4
65 133.2 105.7 158.5



#56
Dibenzofuran
Concen: 19.484 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:168 Resp: 1954798
Ion Ratio Lower Upper
168 100
139 38.0 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 21.851 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:165 Resp: 431314
Ion Ratio Lower Upper
165 100
63 44.6 39.7 59.5
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.919 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:232 Resp: 361639
Ion Ratio Lower Upper
232 100
131 49.1 39.1 58.7
130 2.5 1.8 2.6
166 31.9 25.4 38.0









#59
Diethylphthalate
Concen: 19.668 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:149 Resp: 1644257
Ion Ratio Lower Upper
149 100
177 21.8 17.3 25.9
150 12.1 10.1 15.1



#61
Fluorene
Concen: 19.482 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:166 Resp: 1664680
Ion Ratio Lower Upper
166 100
165 98.5 78.5 117.7
167 13.4 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 19.650 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:204 Resp: 803001
Ion Ratio Lower Upper
204 100
206 33.2 25.7 38.5
141 62.8 49.4 74.0



#63
4-Nitroaniline
Concen: 19.601 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:138 Resp: 310116
Ion Ratio Lower Upper
138 100
92 51.2 42.1 63.1
108 79.3 66.2 99.2









#66
4,6-Dinitro-2-methylphenol
Concen: 18.896 ng/ul
RT: 15.763 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:198 Resp: 198258
Ion Ratio Lower Upper
198 100
182 4.3 4.2 6.2
77 28.3 24.0 36.0



#67
N-Nitrosodiphenylamine
Concen: 19.762 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:169 Resp: 1367775
Ion Ratio Lower Upper
169 100
168 67.3 54.6 81.8
167 36.4 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 18.034 ng/ul
RT: 16.580 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:248 Resp: 417701
Ion Ratio Lower Upper
248 100
250 98.0 78.9 118.3
141 76.1 65.7 98.5



#69
Hexachlorobenzene
Concen: 19.155 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:284 Resp: 524758
Ion Ratio Lower Upper
284 100
142 37.1 30.8 46.2
249 30.0 24.2 36.2





#70

Atrazine

Concen: 18.707 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :

SSTD020175





Tgt Ion:200 Resp: 456888

Ion Ratio Lower Upper

200 100

173 27.7 21.8 32.6

215 47.7 39.4 59.2



#71

Pentachlorophenol

Concen: 17.742 ng/ul

RT: 17.027 min Scan# 2370

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09




Tgt Ion:266 Resp: 288725


Ion Ratio Lower Upper

266 100

264 64.3 49.7 74.5

268 63.2 51.8 77.6





#72

Phenanthrene

Concen: 19.547 ng/ul

RT: 17.427 min Scan# 2438

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :

SSTD020175





Tgt Ion:178 Resp: 2530195

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 19.5 15.7 23.5



#74

Anthracene

Concen: 19.447 ng/ul

RT: 17.515 min Scan# 2453

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09




Tgt Ion:178 Resp: 2558560


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.1 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.807 ng/uL
RT: 13.457 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:216 Resp: 636627
Ion Ratio Lower Upper
216 100
214 78.4 62.6 94.0
218 48.5 38.2 57.2



#76
Pentachlorobenzene
Concen: 19.719 ng/uL
RT: 14.963 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:250 Resp: 685502
Ion Ratio Lower Upper
250 100
248 63.5 50.2 75.2
252 63.8 50.5 75.7









#77
 Carbazole
 Concen: 18.639 ng/ul
 RT: 17.786 min Scan# 2499
 Delta R.T. 0.000 min
 Lab File: BM036571.D
 Acq: 16 Sep 2022 20:09




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020175


Tgt Ion:167 Resp: 2270393
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.2 25.8
 139 13.2 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 19.641 ng/ul
 RT: 18.357 min Scan# 2596
 Delta R.T. -0.006 min
 Lab File: BM036571.D
 Acq: 16 Sep 2022 20:09

Tgt Ion:149 Resp: 2711740
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 5.4 4.3 6.5









#80
Fluoranthene
Concen: 22.111 ng/ul
RT: 19.433 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:202 Resp: 2762389
Ion Ratio Lower Upper
202 100
101 11.6 9.8 14.6
100 8.8 7.4 11.2



#82
Pyrene
Concen: 22.004 ng/ul
RT: 19.792 min Scan# 2840
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:202 Resp: 2909968
Ion Ratio Lower Upper
202 100
101 14.6 10.9 16.3
100 12.0 8.7 13.1





#83

Butylbenzylphthalate

Concen: 20.787 ng/ul

RT: 20.692 min Scan# 2993

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :

SSTD020175





Tgt Ion:149 Resp: 1041663

Ion Ratio Lower Upper

149 100

91 75.1 55.2 82.8

206 19.9 16.2 24.2



#84

3,3'-Dichlorobenzidine

Concen: 19.893 ng/ul

RT: 21.468 min Scan# 3125

Delta R.T. 0.000 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09




Tgt Ion:252 Resp: 805519


Ion Ratio Lower Upper

252 100

254 64.8 53.0 79.4

126 11.7 10.4 15.6









#85
Benzo(a)anthracene
Concen: 20.028 ng/ul
RT: 21.533 min Scan# 3136
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:228 Resp: 2600597
Ion Ratio Lower Upper
228 100
229 20.0 15.8 23.8
226 27.2 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.691 ng/ul
RT: 21.468 min Scan# 3125
Delta R.T. -0.012 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:149 Resp: 1587811
Ion Ratio Lower Upper
149 100
167 27.1 22.0 33.0
279 3.5 3.3 4.9









#87
Chrysene
Concen: 19.536 ng/ul
RT: 21.592 min Scan# 3146
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:228 Resp: 2407852
Ion Ratio Lower Upper
228 100
226 30.3 24.4 36.6
229 19.7 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 22.550 ng/ul
RT: 22.415 min Scan# 3286
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:149 Resp: 2339057









#90
Benzo(b)fluoranthene
Concen: 20.779 ng/ul
RT: 23.250 min Scan# 3428
Delta R.T. 0.000 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09




Instrument :
BNA_M
ClientSampleId :
SSTD020175


Tgt Ion:252 Resp: 2267075
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 10.3 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 19.935 ng/ul
RT: 23.297 min Scan# 3436
Delta R.T. -0.006 min
Lab File: BM036571.D
Acq: 16 Sep 2022 20:09

Tgt Ion:252 Resp: 2173294
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 10.7 8.2 12.4





#93

Benzo(a)pyrene

Concen: 19.851 ng/ul

RT: 23.886 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :

SSTD020175





Tgt Ion:252 Resp: 1903994

Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 11.8 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.279 ng/ul

RT: 26.544 min Scan# 3988

Delta R.T. -0.006 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09




Tgt Ion:276 Resp: 2047402


Ion Ratio Lower Upper

276 100

138 27.1 20.2 30.4

277 25.3 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 18.400 ng/ul

RT: 26.562 min Scan# 3991

Delta R.T. -0.012 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Instrument :

BNA_M

ClientSampleId :

SSTD020175





Tgt Ion:278 Resp: 1869043

Ion Ratio Lower Upper

278 100

139 16.7 14.5 21.7

279 24.0 19.7 29.5



#96

Benzo(g,h,i)perylene

Concen: 18.438 ng/ul

RT: 27.321 min Scan# 4120

Delta R.T. -0.012 min

Lab File: BM036571.D

Acq: 16 Sep 2022 20:09

Tgt Ion:276 Resp: 1790393

Ion Ratio Lower Upper

276 100

138 23.8 18.7 28.1

277 24.2 19.2 28.8

