

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
 Data File : BM041904.D
 Acq On : 19 Sep 2023 13:11
 Operator : MA/JU
 Sample : 04346-06
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PT-ACIDS-WP

Quant Time: Sep 19 14:12:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 19 04:49:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	283256	20.000	ng	-0.01
21) Naphthalene-d8	10.780	136	1185994	20.000	ng	0.00
39) Acenaphthene-d10	14.610	164	738889	20.000	ng	0.00
64) Phenanthrene-d10	17.362	188	1563013	20.000	ng	0.00
76) Chrysene-d12	21.539	240	1447377	20.000	ng	0.00
86) Perylene-d12	23.962	264	1586679	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	2891478	170.025	ng	0.00
7) Phenol-d6	7.110	99	3891657	164.813	ng	0.00
23) Nitrobenzene-d5	9.157	82	2446706	105.038	ng	0.00
42) 2,4,6-Tribromophenol	16.104	330	1543863	167.463	ng	0.00
45) 2-Fluorobiphenyl	13.227	172	5192448	103.145	ng	0.00
79) Terphenyl-d14	19.974	244	8277963	104.165	ng	0.00
Target Compounds						Qvalue
8) 2-Chlorophenol	7.516	128	2762692	145.914	ng	99
9) Benzaldehyde	7.140	77	36112	2.705	ng	# 4
10) Phenol	7.140	94	3049488	125.817	ng	99
11) bis(2-Chloroethyl)ether	7.516	93	61737	3.054	ng	# 1
15) Benzyl Alcohol	8.398	79	1112538	59.680	ng	# 1
17) 2-Methylphenol	8.398	107	2419506	137.283	ng	99
20) 3+4-Methylphenols	8.722	107	3254522	134.534	ng	90
26) 2-Nitrophenol	9.898	139	603885	57.991	ng	99
27) 2,4-Dimethylphenol	9.933	122	2637008	140.806	ng	98
28) bis(2-Chloroethoxy)met...	9.933	93	102361	3.809	ng	# 1
29) 2,4-Dichlorophenol	10.416	162	2791425	154.018	ng	99
32) Benzoic acid	9.933	122	2637008	178.709	ng	# 14
35) Caprolactam	12.051	113	85066	13.132	ng	# 16
36) 4-Chloro-3-methylphenol	12.051	107	2950840	146.094	ng	98
43) 2,4,6-Trichlorophenol	13.039	196	2269517	155.627	ng	99
44) 2,4,5-Trichlorophenol	13.104	196	1921847	112.606	ng	98
52) Acenaphthene	14.757	154	860320	21.709	ng	# 5
53) 3-Nitroaniline	14.757	138	40899	3.157	ng	# 1
54) 2,4-Dinitrophenol	14.757	184	1421896	196.671	ng	99
55) Dibenzofuran	15.227	168	660222	10.321	ng	# 43
59) 2,3,4,6-Tetrachlorophenol	15.227	232	2432280	173.629	ng	95
63) Azobenzene	15.745	77	385474	6.998	ng	# 1
65) 4,6-Dinitro-2-methylph...	15.751	198	1538812	171.448	ng	93
69) Atrazine	16.992	200	391324	28.427	ng	# 35
70) Pentachlorophenol	16.992	266	2182896	169.415	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
Data File : BM041904.D
Acq On : 19 Sep 2023 13:11
Operator : MA/JU
Sample : 04346-06
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP

Quant Time: Sep 19 14:12:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 19 04:49:34 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.957 min Scan# 81
Delta R.T. -0.012 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:152 Resp: 283256
Ion Ratio Lower Upper
152 100
150 158.8 125.9 188.9
115 59.6 46.8 70.2



#5
2-Fluorophenol
Concen: 170.025 ng
RT: 5.498 min Scan# 393
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:112 Resp: 2891478
Ion Ratio Lower Upper
112 100
64 60.6 48.1 72.1
63 32.3 25.0 37.6





#7

Phenol-d6

Concen: 164.813 ng

RT: 7.110 min Scan# 60

Delta R.T. -0.000 min

Lab File: BM041904.D

Acq: 19 Sep 2023 13:11

Instrument :





BNA_M

ClientSampleId :

PT-ACIDS-WP

Tgt Ion: 99 Resp: 3891657

Ion	Ratio	Lower	Upper
99	100		
42	20.4	16.6	25.0
71	33.9	26.8	40.2



#8

2-Chlorophenol

Concen: 145.914 ng

RT: 7.516 min Scan# 736




Delta R.T. -0.006 min


Lab File: BM041904.D

Acq: 19 Sep 2023 13:11

Tgt Ion: 128 Resp: 2762692

Ion	Ratio	Lower	Upper
128	100		
130	32.1	12.4	52.4
64	51.2	31.7	71.7









#9
Benzaldehyde
Concen: 2.705 ng
RT: 7.140 min Scan# 61
Delta R.T. 0.011 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion: 77 Resp: 36112
Ion Ratio Lower Upper
77 100
105 1.4 75.1 115.1#
106 0.0 71.6 111.6#



#10
Phenol
Concen: 125.817 ng
RT: 7.140 min Scan# 672
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion: 94 Resp: 3049488
Ion Ratio Lower Upper
94 100
65 29.4 9.1 49.1
66 36.9 17.3 57.3









#11
bis(2-Chloroethyl)ether
Concen: 3.054 ng
RT: 7.516 min Scan# 716
Delta R.T. 0.117 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion: 93 Resp: 61737
Ion Ratio Lower Upper
93 100
63 1483.9 59.5 99.5#
95 2.9 12.9 52.9#



#15
Benzyl Alcohol
Concen: 59.680 ng
RT: 8.398 min Scan# 886
Delta R.T. 0.182 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion: 79 Resp: 1112538
Ion Ratio Lower Upper
79 100
108 251.6 61.4 92.0#
77 105.9 52.0 78.0#









#17
2-Methylphenol
Concen: 137.283 ng
RT: 8.398 min Scan# 886
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:107 Resp: 2419506
Ion Ratio Lower Upper
107 100
108 114.0 91.9 137.9
77 48.0 39.0 58.6
79 45.3 36.7 55.1



#20
3+4-Methylphenols
Concen: 134.534 ng
RT: 8.722 min Scan# 941
Delta R.T. -0.012 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:107 Resp: 3254522
Ion Ratio Lower Upper
107 100
108 81.3 71.5 111.5
77 31.9 14.9 54.9
79 22.1 10.7 50.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.780 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:136 Resp: 1185994
Ion Ratio Lower Upper
136 100
137 10.4 8.6 13.0
54 8.9 7.0 10.4
68 6.4 5.3 7.9



#23
Nitrobenzene-d5
Concen: 105.038 ng
RT: 9.157 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion: 82 Resp: 2446706
Ion Ratio Lower Upper
82 100
128 42.0 33.8 50.6
54 52.7 42.8 64.2









#26
2-Nitrophenol
Concen: 57.991 ng
RT: 9.898 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:139 Resp: 603885
Ion Ratio Lower Upper
139 100
109 29.3 23.0 34.4
65 54.9 44.5 66.7



#27
2,4-Dimethylphenol
Concen: 140.806 ng
RT: 9.933 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:122 Resp: 2637008
Ion Ratio Lower Upper
122 100
107 115.0 93.8 140.6
121 56.8 46.4 69.6









#28
bis(2-Chloroethoxy)methane
Concen: 3.809 ng
RT: 9.933 min Scan# 1192
Delta R.T. -0.265 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion: 93 Resp: 102361
Ion Ratio Lower Upper
93 100
95 21.1 25.9 38.9#
123 205.9 9.2 13.8#



#29
2,4-Dichlorophenol
Concen: 154.018 ng
RT: 10.416 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:162 Resp: 2791425
Ion Ratio Lower Upper
162 100
164 64.9 45.2 85.2
98 35.6 16.7 56.7









#32
Benzoic acid
Concen: 178.709 ng
RT: 9.933 min Scan# 1173
Delta R.T. -0.153 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:122 Resp: 2637008
Ion Ratio Lower Upper
122 100
105 3.6 103.7 143.7#
77 37.8 79.3 119.3#



#35
Caprolactam
Concen: 13.132 ng
RT: 12.051 min Scan# 1507
Delta R.T. 0.247 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:113 Resp: 85066
Ion Ratio Lower Upper
113 100
55 83.9 158.5 198.5#
56 8.0 115.8 155.8#









#36
4-Chloro-3-methylphenol
Concen: 146.094 ng
RT: 12.051 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:107 Resp: 2950840
Ion Ratio Lower Upper
107 100
144 25.8 20.2 30.2
142 79.3 61.7 92.5



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.610 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:164 Resp: 738889
Ion Ratio Lower Upper
164 100
162 100.6 79.3 118.9
160 44.9 35.6 53.4









#42
2,4,6-Tribromophenol
Concen: 167.463 ng
RT: 16.104 min Scan# 2197
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:330 Resp: 1543863
Ion Ratio Lower Upper
330 100
332 96.8 77.8 116.6
141 32.2 27.0 40.4



#43
2,4,6-Trichlorophenol
Concen: 155.627 ng
RT: 13.039 min Scan# 1675
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:196 Resp: 2269517
Ion Ratio Lower Upper
196 100
198 97.4 76.9 115.3
200 31.9 24.7 37.1









#44
2,4,5-Trichlorophenol
Concen: 112.606 ng
RT: 13.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:196 Resp: 1921847
Ion Ratio Lower Upper
196 100
198 96.6 77.2 115.8
97 47.4 41.1 61.7
132 25.0 21.8 32.6



#45
2-Fluorobiphenyl
Concen: 103.145 ng
RT: 13.227 min Scan# 1707
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:172 Resp: 5192448
Ion Ratio Lower Upper
172 100
171 35.2 28.2 42.4
170 23.8 19.1 28.7





#52

Acenaphthene

Concen: 21.709 ng

RT: 14.757 min Scan# 1967

Delta R.T. 0.076 min

Lab File: BM041904.D

Acq: 19 Sep 2023 13:11

Instrument :

BNA_M

ClientSampleId :

PT-ACIDS-WP





Tgt Ion:154 Resp: 860320

Ion Ratio Lower Upper

154 100

153 0.4 90.3 135.5#

152 0.3 43.4 65.2#



#53

3-Nitroaniline

Concen: 3.157 ng

RT: 14.757 min Scan# 1967

Delta R.T. 0.200 min

Lab File: BM041904.D

Acq: 19 Sep 2023 13:11




Tgt Ion:138 Resp: 40899


Ion Ratio Lower Upper

138 100

108 208.4 8.0 12.0#

92 908.5 93.6 140.4#









#54
2,4-Dinitrophenol
Concen: 196.671 ng
RT: 14.757 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:184 Resp: 1421896
Ion Ratio Lower Upper
184 100
63 69.9 55.8 83.8
154 59.5 48.9 73.3



#55
Dibenzofuran
Concen: 10.321 ng
RT: 15.227 min Scan# 2047
Delta R.T. 0.212 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:168 Resp: 660222
Ion Ratio Lower Upper
168 100
139 0.0 29.4 44.2#
169 30.3 10.8 16.2#









#59
2,3,4,6-Tetrachlorophenol
Concen: 173.629 ng
RT: 15.227 min Scan# 2047
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:232 Resp: 2432280
Ion Ratio Lower Upper
232 100
131 39.6 34.6 51.8
130 2.1 1.8 2.8
166 27.8 23.8 35.6



#63
Azobenzene
Concen: 6.998 ng
RT: 15.745 min Scan# 2135
Delta R.T. -0.206 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion: 77 Resp: 385474
Ion Ratio Lower Upper
77 100
182 12.8 4.9 44.9
105 169.4 0.7 40.7#
51 199.3 11.8 51.8#









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.362 min Scan# 2411
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:188 Resp: 1563013
Ion Ratio Lower Upper
188 100
94 9.5 7.6 11.4
80 10.9 8.3 12.5



#65
4,6-Dinitro-2-methylphenol
Concen: 171.448 ng
RT: 15.751 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:198 Resp: 1538812
Ion Ratio Lower Upper
198 100
51 45.7 32.0 72.0
105 40.5 23.6 63.6





#69

Atrazine

Concen: 28.427 ng

RT: 16.992 min Scan# 21

Delta R.T. 0.170 min

Lab File: BM041904.D

Acq: 19 Sep 2023 13:11

Instrument :

BNA_M

ClientSampleId :

PT-ACIDS-WP





Tgt Ion:200 Resp: 391324

Ion Ratio Lower Upper

200 100

173 0.0 6.5 46.5#

215 0.1 28.9 68.9#



#70

Pentachlorophenol

Concen: 169.415 ng

RT: 16.992 min Scan# 2347

Delta R.T. -0.000 min

Lab File: BM041904.D

Acq: 19 Sep 2023 13:11




Tgt Ion:266 Resp: 2182896


Ion Ratio Lower Upper

266 100

268 63.4 50.5 75.7

264 63.0 50.6 76.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.539 min Scan# 3121
Delta R.T. -0.006 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11




Instrument :
BNA_M
ClientSampleId :
PT-ACIDS-WP


Tgt Ion:240 Resp: 1447377
Ion Ratio Lower Upper
240 100
120 11.0 8.1 12.1
236 25.7 20.7 31.1



#79
Terphenyl-d14
Concen: 104.165 ng
RT: 19.974 min Scan# 2854
Delta R.T. -0.000 min
Lab File: BM041904.D
Acq: 19 Sep 2023 13:11

Tgt Ion:244 Resp: 8277963
Ion Ratio Lower Upper
244 100
212 7.1 5.7 8.5
122 9.4 7.4 11.0





#86

Perylene-d12

Concen: 20.000 ng

RT: 23.962 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM041904.D


Acq: 19 Sep 2023 13:11

Instrument :

BNA_M

ClientSampleId :

PT-ACIDS-WP



Tgt Ion:264 Resp: 1586679

Ion Ratio Lower Upper

264 100

260 24.0 18.6 28.0

265 22.4 17.0 25.6

