

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
 Data File : BM041915.D
 Acq On : 19 Sep 2023 22:39
 Operator : MA/JU
 Sample : PB155451BSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB155451BSD

Quant Time: Sep 20 02:12:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 19 04:49:34 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/20/2023
 Supervised By :mohammad ahmed 09/21/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	292747	20.000	ng	0.00
21) Naphthalene-d8	10.781	136	1214431	20.000	ng	0.00
39) Acenaphthene-d10	14.610	164	728056	20.000	ng	0.00
64) Phenanthrene-d10	17.363	188	1465385	20.000	ng	0.00
76) Chrysene-d12	21.533	240	1218985	20.000	ng	-0.01
86) Perylene-d12	23.950	264	1218362	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.505	112	1943097	110.554	ng	0.00
7) Phenol-d6	7.110	99	2555840	104.731	ng	0.00
23) Nitrobenzene-d5	9.157	82	2361611	99.011	ng	0.00
42) 2,4,6-Tribromophenol	16.104	330	906929	99.839	ng	0.00
45) 2-Fluorobiphenyl	13.228	172	4951302	99.819	ng	0.00
79) Terphenyl-d14	19.974	244	7153050	106.874	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.411	88	257430	35.909	ng	98
3) Pyridine	3.816	79	751102	33.727	ng	99
4) n-Nitrosodimethylamine	3.740	42	492230	46.499	ng	97
6) Aniline	7.304	93	1250599	38.680	ng	100
8) 2-Chlorophenol	7.522	128	1031490	52.713	ng	98
9) Benzaldehyde	7.122	77	323359	23.438	ng	99
10) Phenol	7.140	94	1338865	53.449	ng	99
11) bis(2-Chloroethyl)ether	7.399	93	992833	47.515	ng	98
12) 1,3-Dichlorobenzene	7.846	146	1023347	48.978	ng	99
13) 1,4-Dichlorobenzene	7.998	146	1050387	49.476	ng	99
14) 1,2-Dichlorobenzene	8.310	146	1021836	50.072	ng	98
15) Benzyl Alcohol	8.216	79	951935	49.410	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.487	45	1534063	48.005	ng	98
17) 2-Methylphenol	8.398	107	881838	48.413	ng	99
18) Hexachloroethane	9.034	117	388608	48.697	ng	99
19) n-Nitroso-di-n-propyla...	8.781	70	830431	45.777	ng	98
20) 3+4-Methylphenols	8.728	107	1189945	47.595	ng	97
22) Acetophenone	8.810	105	1503013	49.508	ng	# 100
24) Nitrobenzene	9.204	77	1181736	49.916	ng	100
25) Isophorone	9.710	82	2237040	47.519	ng	100
26) 2-Nitrophenol	9.898	139	551485	51.719	ng	100
27) 2,4-Dimethylphenol	9.934	122	950155	49.546	ng	99
28) bis(2-Chloroethoxy)met...	10.192	93	1365114	49.614	ng	100
29) 2,4-Dichlorophenol	10.416	162	995297	53.630	ng	99
30) 1,2,4-Trichlorobenzene	10.634	180	974356	51.137	ng	98
31) Naphthalene	10.834	128	2979535	48.734	ng	100
32) Benzoic acid	10.087	122	745481	49.338	ng	99
33) 4-Chloroaniline	10.957	127	671049	24.523	ng	99
34) Hexachlorobutadiene	11.063	225	574188	49.337	ng	100
35) Caprolactam	11.798	113	309051m	46.593	ng	
36) 4-Chloro-3-methylphenol	12.057	107	1038618	50.217	ng	98
37) 2-Methylnaphthalene	12.433	142	2079569	46.261	ng	99
38) 1-Methylnaphthalene	12.657	142	1996816	47.297	ng	100
40) 1,2,4,5-Tetrachloroben...	12.786	216	1078242	51.067	ng	100
41) Hexachlorocyclopentadiene	12.739	237	1365719	115.034	ng	99
43) 2,4,6-Trichlorophenol	13.039	196	773534	53.833	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
 Data File : BM041915.D
 Acq On : 19 Sep 2023 22:39
 Operator : MA/JU
 Sample : PB155451BSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB155451BSD

Quant Time: Sep 20 02:12:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 19 04:49:34 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/20/2023
 Supervised By :mohammad ahmed 09/21/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.104	196	834131	49.601	ng	98
46) 1,1'-Biphenyl	13.445	154	2721782	50.669	ng	100
47) 2-Chloronaphthalene	13.492	162	2062863	52.732	ng	100
48) 2-Nitroaniline	13.722	65	712713	50.986	ng	99
49) Acenaphthylene	14.339	152	3385429	52.146	ng	99
50) Dimethylphthalate	14.075	163	2646168	49.163	ng	99
51) 2,6-Dinitrotoluene	14.210	165	573106	50.244	ng	98
52) Acenaphthene	14.675	154	1951991	49.990	ng	99
53) 3-Nitroaniline	14.551	138	460134	36.041	ng	99
54) 2,4-Dinitrophenol	14.751	184	727062	102.061	ng	97
55) Dibenzofuran	15.010	168	3109838	49.338	ng	99
56) 4-Nitrophenol	14.839	139	1043641	103.072	ng	98
57) 2,4-Dinitrotoluene	14.998	165	766700	50.805	ng	98
58) Fluorene	15.663	166	2455287	49.409	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.222	232	710422	51.468	ng	99
60) Diethylphthalate	15.422	149	2602621	48.610	ng	99
61) 4-Chlorophenyl-phenyle...	15.651	204	1285778	49.321	ng	97
62) 4-Nitroaniline	15.721	138	572177	46.606	ng	99
63) Azobenzene	15.945	77	2685189	49.470	ng	99
65) 4,6-Dinitro-2-methylph...	15.745	198	443034	52.650	ng	98
66) n-Nitrosodiphenylamine	15.874	169	2185624	52.809	ng	99
67) 4-Bromophenyl-phenylether	16.545	248	763962	50.591	ng	99
68) Hexachlorobenzene	16.633	284	889381	55.626	ng	98
69) Atrazine	16.816	200	773153	59.906	ng	99
70) Pentachlorophenol	16.992	266	1068870	88.482	ng	99
71) Phenanthrene	17.410	178	3828667	52.388	ng	100
72) Anthracene	17.498	178	3903775	52.363	ng	100
73) Carbazole	17.780	167	3592508	52.553	ng	99
74) Di-n-butylphthalate	18.298	149	4530295	51.621	ng	100
75) Fluoranthene	19.415	202	4372625	50.963	ng	100
77) Benzidine	19.615	184	1408880	56.623	ng	99
78) Pyrene	19.780	202	4524760	53.346	ng	100
80) Butylbenzylphthalate	20.656	149	1952656	51.761	ng	96
81) Benzo(a)anthracene	21.521	228	4125463	51.891	ng	100
82) 3,3'-Dichlorobenzidine	21.456	252	1277138	45.614	ng	99
83) Chrysene	21.574	228	3762796	50.494	ng	100
84) Bis(2-ethylhexyl)phtha...	21.398	149	2857286	51.055	ng	100
85) Di-n-octyl phthalate	22.327	149	4869340	51.689	ng	100
87) Indeno(1,2,3-cd)pyrene	26.462	276	3894153	53.532	ng	100
88) Benzo(b)fluoranthene	23.209	252	3847659	54.092	ng	100
89) Benzo(k)fluoranthene	23.262	252	3674346	52.273	ng	100
90) Benzo(a)pyrene	23.844	252	3267113	48.948	ng	100
91) Dibenzo(a,h)anthracene	26.480	278	3270338	53.926	ng	100
92) Benzo(g,h,i)perylene	27.250	276	3080169	53.157	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
 Data File : BM041915.D
 Acq On : 19 Sep 2023 22:39
 Operator : MA/JU
 Sample : PB155451BSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 PB155451BSD

Quant Time: Sep 20 02:12:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 19 04:49:34 2023
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 09/20/2023
 Supervised By :mohammad ahmed 09/21/2023





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.963 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:152 Resp: 29274
Ion Ratio Lower Upper
152 100
150 159.6 125.9 188.9
115 59.4 46.8 70.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#2
1,4-Dioxane
Concen: 35.909 ng
RT: 3.411 min Scan# 38
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion: 88 Resp: 257430
Ion Ratio Lower Upper
88 100
58 76.0 62.4 93.6
43 32.7 25.8 38.6





#3
Pyridine
Concen: 33.727 ng
RT: 3.816 min Scan# 108
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion: 79 Resp: 751102
Ion Ratio Lower Upper
79 100
52 72.1 57.2 85.8
51 37.2 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#4
n-Nitrosodimethylamine
Concen: 46.499 ng
RT: 3.740 min Scan# 94
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion: 42 Resp: 492230
Ion Ratio Lower Upper
42 100
74 120.5 98.8 148.2
44 6.0 5.1 7.7





#5
2-Fluorophenol
Concen: 110.554 ng
RT: 5.505 min Scan# 394
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:112 Resp: 1943097
Ion Ratio Lower Upper
112 100
64 59.4 48.1 72.1
63 31.6 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#6
Aniline
Concen: 38.680 ng
RT: 7.304 min Scan# 700
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion: 93 Resp: 1250599
Ion Ratio Lower Upper
93 100
66 39.5 31.4 47.2
65 20.6 16.2 24.4









#7
Phenol-d6
Concen: 104.731 ng
RT: 7.110 min Scan# 667
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Instrument :
BNA_M
ClientSampleId :
PB155451BSD

Tgt Ion: 99 Resp: 2555840
Ion Ratio Lower Upper
99 100
42 20.7 16.6 25.0
71 33.9 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#8
2-Chlorophenol
Concen: 52.713 ng
RT: 7.522 min Scan# 737
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:128 Resp: 1031490
Ion Ratio Lower Upper
128 100
130 32.4 12.4 52.4
64 48.9 31.7 71.7





#9
Benzaldehyde
Concen: 23.438 ng
RT: 7.122 min Scan# 609
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion: 77 Resp: 323359
Ion Ratio Lower Upper
77 100
105 94.0 75.1 115.1
106 91.6 71.6 111.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#10
Phenol
Concen: 53.449 ng
RT: 7.140 min Scan# 672
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion: 94 Resp: 1338865
Ion Ratio Lower Upper
94 100
65 29.0 9.1 49.1
66 36.4 17.3 57.3





#11
bis(2-Chloroethyl)ether
Concen: 47.515 ng
RT: 7.399 min Scan# 716
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion: 93 Resp: 992833
Ion Ratio Lower Upper
93 100
63 77.3 59.5 99.5
95 32.3 12.9 52.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#12
1,3-Dichlorobenzene
Concen: 48.978 ng
RT: 7.846 min Scan# 792
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:146 Resp: 1023347
Ion Ratio Lower Upper
146 100
148 64.4 51.7 77.5
75 33.1 27.2 40.8









#13
1,4-Dichlorobenzene
Concen: 49.476 ng
RT: 7.998 min Scan# 819
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Instrument :
BNA_M
ClientSampleId :
PB155451BSD

Tgt Ion:146 Resp: 105038
Ion Ratio Lower Upper
146 100
148 64.0 51.9 77.9
111 40.3 32.1 48.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#14
1,2-Dichlorobenzene
Concen: 50.072 ng
RT: 8.310 min Scan# 871
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:146 Resp: 1021836
Ion Ratio Lower Upper
146 100
148 63.0 51.5 77.3
111 41.7 34.0 51.0





#15
Benzyl Alcohol
Concen: 49.410 ng
RT: 8.216 min Scan# 801
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion: 79 Resp: 951935
Ion Ratio Lower Upper
79 100
108 75.9 61.4 92.0
77 65.1 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#16
2,2'-oxybis(1-Chloropropane)
Concen: 48.005 ng
RT: 8.487 min Scan# 901
Delta R.T. 0.012 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion: 45 Resp: 1534063
Ion Ratio Lower Upper
45 100
77 12.9 0.0 33.4
79 10.4 0.0 29.6





#17
2-Methylphenol
Concen: 48.413 ng
RT: 8.398 min Scan# 886
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:107 Resp: 881838
Ion Ratio Lower Upper
107 100
108 114.0 91.9 137.9
77 48.0 39.0 58.6
79 45.8 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#18
Hexachloroethane
Concen: 48.697 ng
RT: 9.034 min Scan# 994
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:117 Resp: 388608
Ion Ratio Lower Upper
117 100
119 95.1 77.2 115.8
201 102.3 81.9 122.9





#19
n-Nitroso-di-n-propylamine
Concen: 45.777 ng
RT: 8.781 min Scan# 951
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion: 70 Resp: 83043
Ion Ratio Lower Upper
70 100
42 62.8 49.2 73.8
101 9.5 7.0 10.6
130 19.5 15.0 22.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#20
3+4-Methylphenols
Concen: 47.595 ng
RT: 8.728 min Scan# 942
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:107 Resp: 1189945
Ion Ratio Lower Upper
107 100
108 88.9 71.5 111.5
77 34.2 14.9 54.9
79 28.9 10.7 50.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.781 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:136 Resp: 121443
Ion Ratio Lower Upper
136 100
137 10.9 8.6 13.0
54 8.7 7.0 10.4
68 6.6 5.3 7.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#22
Acetophenone
Concen: 49.508 ng
RT: 8.810 min Scan# 956
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion:105 Resp: 1503013
Ion Ratio Lower Upper
105 100
71 0.1 0.2 0.2#
51 30.6 24.3 36.5
120 23.1 18.2 27.4





#23

Nitrobenzene-d5

Concen: 99.011 ng

RT: 9.157 min Scan# 1015

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD

Tgt Ion: 82 Resp: 236161

Ion Ratio Lower Upper

82 100

128 41.7 33.8 50.6





54 53.9 42.8 64.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#24

Nitrobenzene

Concen: 49.916 ng

RT: 9.204 min Scan# 1023

Delta R.T. 0.000 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion: 77 Resp: 1181736


Ion Ratio Lower Upper

77 100

123 43.3 34.6 52.0

65 13.3 11.1 16.7





#25

Isophorone

Concen: 47.519 ng

RT: 9.710 min Scan# 1110

Delta R.T. -0.006 min

Lab File: BM041915.D


Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD



Tgt Ion: 82 Resp: 2237040

Ion Ratio Lower Upper

82 100

95 6.7 5.4 8.0




138 16.4 13.1 19.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#26

2-Nitrophenol

Concen: 51.719 ng

RT: 9.898 min Scan# 1141

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:139 Resp: 551485


Ion Ratio Lower Upper

139 100

109 28.8 23.0 34.4


65 56.0 44.5 66.7





#27
2,4-Dimethylphenol
Concen: 49.546 ng
RT: 9.934 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:122 Resp: 95015
Ion Ratio Lower Upper
122 100
107 115.7 93.8 140.6
121 57.3 46.4 69.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#28
bis(2-Chloroethoxy)methane
Concen: 49.614 ng
RT: 10.192 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion: 93 Resp: 1365114
Ion Ratio Lower Upper
93 100
95 32.3 25.9 38.9
123 11.3 9.2 13.8





#29
2,4-Dichlorophenol
Concen: 53.630 ng
RT: 10.416 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:162 Resp: 99529
Ion Ratio Lower Upper
162 100
164 63.4 45.2 85.2
98 36.7 16.7 56.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#30
1,2,4-Trichlorobenzene
Concen: 51.137 ng
RT: 10.634 min Scan# 1266
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:180 Resp: 974356
Ion Ratio Lower Upper
180 100
182 95.2 77.7 116.5
145 29.6 24.6 37.0





#31
Naphthalene
Concen: 48.734 ng
RT: 10.834 min Scan# 1173
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:128 Resp: 297953
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.1 10.4 15.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#32
Benzoic acid
Concen: 49.338 ng
RT: 10.087 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:122 Resp: 745481
Ion Ratio Lower Upper
122 100
105 124.4 103.7 143.7
77 101.3 79.3 119.3





#33
4-Chloroaniline
Concen: 24.523 ng
RT: 10.957 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:127 Resp: 671049
Ion Ratio Lower Upper
127 100
129 32.0 25.7 38.5
65 34.9 27.3 40.9
92 20.3 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#34
Hexachlorobutadiene
Concen: 49.337 ng
RT: 11.063 min Scan# 1339
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:225 Resp: 574188
Ion Ratio Lower Upper
225 100
223 63.5 50.7 76.1
227 64.6 51.4 77.0





#35

Caprolactam

Concen: 46.593 ng m

RT: 11.798 min Scan# 1465

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD

Tgt Ion:113 Resp: 30905

Ion Ratio Lower Upper

113 100

55 177.7 158.5 198.5





56 135.6 115.8 155.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#36

4-Chloro-3-methylphenol

Concen: 50.217 ng

RT: 12.057 min Scan# 1508

Delta R.T. 0.000 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:107 Resp: 1038618


Ion Ratio Lower Upper

107 100

144 25.8 20.2 30.2


142 79.5 61.7 92.5





#37
2-Methylnaphthalene
Concen: 46.261 ng
RT: 12.433 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:142 Resp: 2079569
Ion Ratio Lower Upper
142 100
141 87.7 70.9 106.3
115 35.5 28.5 42.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#38
1-Methylnaphthalene
Concen: 47.297 ng
RT: 12.657 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:142 Resp: 1996816
Ion Ratio Lower Upper
142 100
141 91.9 73.8 110.6
116 3.8 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.610 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:164 Resp: 728050
Ion Ratio Lower Upper
164 100
162 99.9 79.3 118.9
160 44.9 35.6 53.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#40
1,2,4,5-Tetrachlorobenzene
Concen: 51.067 ng
RT: 12.786 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:216 Resp: 1078242
Ion Ratio Lower Upper
216 100
214 78.7 62.6 93.8
179 20.4 16.3 24.5
108 17.6 14.2 21.4





#41
Hexachlorocyclopentadiene
Concen: 115.034 ng
RT: 12.739 min Scan# 10
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:237 Resp: 1365719
Ion Ratio Lower Upper
237 100
235 62.5 42.0 82.0
272 13.2 0.0 32.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#42
2,4,6-Tribromophenol
Concen: 99.839 ng
RT: 16.104 min Scan# 2196
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:330 Resp: 906929
Ion Ratio Lower Upper
330 100
332 97.0 77.8 116.6
141 34.3 27.0 40.4





#43
2,4,6-Trichlorophenol
Concen: 53.833 ng
RT: 13.039 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:196 Resp: 773534
Ion Ratio Lower Upper
196 100
198 97.2 76.9 115.3
200 31.5 24.7 37.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#44
2,4,5-Trichlorophenol
Concen: 49.601 ng
RT: 13.104 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:196 Resp: 834131
Ion Ratio Lower Upper
196 100
198 95.3 77.2 115.8
97 50.9 41.1 61.7
132 30.9 21.8 32.6





#45

2-Fluorobiphenyl

Concen: 99.819 ng

RT: 13.228 min Scan# 1707

Delta R.T. -0.006 min

Lab File: BM041915.D


Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD



Tgt Ion:172 Resp: 4951302

Ion Ratio Lower Upper

172 100

171 35.3 28.2 42.4




170 23.6 19.1 28.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#46

1,1'-Biphenyl

Concen: 50.669 ng

RT: 13.445 min Scan# 1744

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:154 Resp: 2721782


Ion Ratio Lower Upper

154 100

153 40.1 19.9 59.9


76 13.3 0.0 33.1





#47
2-Chloronaphthalene
Concen: 52.732 ng
RT: 13.492 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:162 Resp: 206286
Ion Ratio Lower Upper
162 100
127 38.8 31.0 46.4
164 32.4 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#48
2-Nitroaniline
Concen: 50.986 ng
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion: 65 Resp: 712713
Ion Ratio Lower Upper
65 100
92 65.3 52.8 79.2
138 98.5 80.3 120.5





#49

Acenaphthylene

Concen: 52.146 ng

RT: 14.339 min Scan# 1896

Delta R.T. -0.006 min

Lab File: BM041915.D


Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD



Tgt Ion:152 Resp: 3385429

Ion Ratio Lower Upper

152 100

151 20.1 15.8 23.8




153 13.2 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#50

Dimethylphthalate

Concen: 49.163 ng

RT: 14.075 min Scan# 1851

Delta R.T. 0.000 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:163 Resp: 2646168


Ion Ratio Lower Upper

163 100

194 4.3 3.4 5.0


164 9.9 8.2 12.4





#51
2,6-Dinitrotoluene
Concen: 50.244 ng
RT: 14.210 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:165 Resp: 573100
Ion Ratio Lower Upper
165 100
63 64.4 50.2 75.2
89 63.1 49.3 73.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#52
Acenaphthene
Concen: 49.990 ng
RT: 14.675 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:154 Resp: 1951991
Ion Ratio Lower Upper
154 100
153 111.1 90.3 135.5
152 53.9 43.4 65.2





#53
3-Nitroaniline
Concen: 36.041 ng
RT: 14.551 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:138 Resp: 460134
Ion Ratio Lower Upper
138 100
108 9.6 8.0 12.0
92 118.0 93.6 140.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#54
2,4-Dinitrophenol
Concen: 102.061 ng
RT: 14.751 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:184 Resp: 727062
Ion Ratio Lower Upper
184 100
63 73.5 55.8 83.8
154 60.6 48.9 73.3





#55
Dibenzenofuran
Concen: 49.338 ng
RT: 15.010 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:168 Resp: 310983
Ion Ratio Lower Upper
168 100
139 36.3 29.4 44.2
169 13.4 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#56
4-Nitrophenol
Concen: 103.072 ng
RT: 14.839 min Scan# 1981
Delta R.T. 0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:139 Resp: 1043641
Ion Ratio Lower Upper
139 100
109 74.3 53.4 93.4
65 105.1 87.4 127.4





#57
2,4-Dinitrotoluene
Concen: 50.805 ng
RT: 14.998 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:165 Resp: 766700
Ion Ratio Lower Upper
165 100
63 55.4 42.3 63.5
89 78.2 61.8 92.6
182 2.7 2.0 3.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#58
Fluorene
Concen: 49.409 ng
RT: 15.663 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Tgt Ion:166 Resp: 2455287
Ion Ratio Lower Upper
166 100
165 96.7 78.2 117.2
167 13.3 11.0 16.6





#59
2,3,4,6-Tetrachlorophenol
Concen: 51.468 ng
RT: 15.222 min Scan# 2046
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:232 Resp: 710421
Ion Ratio Lower Upper
232 100
131 43.9 34.6 51.8
130 2.3 1.8 2.8
166 30.2 23.8 35.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#60
Diethylphthalate
Concen: 48.610 ng
RT: 15.422 min Scan# 2080
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:149 Resp: 2602621
Ion Ratio Lower Upper
149 100
177 22.7 17.9 26.9
150 12.8 10.1 15.1





#61
4-Chlorophenyl-phenylether
Concen: 49.321 ng
RT: 15.651 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:204 Resp: 1285778
Ion Ratio Lower Upper
204 100
206 33.4 26.1 39.1
141 51.4 43.6 65.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#62
4-Nitroaniline
Concen: 46.606 ng
RT: 15.721 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:138 Resp: 572177
Ion Ratio Lower Upper
138 100
92 52.6 32.3 72.3
108 79.8 59.0 99.0





#63

Azobenzene

Concen: 49.470 ng

RT: 15.945 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM041915.D


Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD



Tgt Ion: 77 Resp: 268518

Ion Ratio Lower Upper

77 100

182 24.4 4.9 44.9

105 20.7 0.7 40.7




51 32.8 11.8 51.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.363 min Scan# 2410

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:188 Resp: 1465385


Ion Ratio Lower Upper

188 100

94 9.5 7.6 11.4


80 10.5 8.3 12.5





#65
4,6-Dinitro-2-methylphenol
Concen: 52.650 ng
RT: 15.745 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:198 Resp: 443034
Ion Ratio Lower Upper
198 100
51 53.1 32.0 72.0
105 44.6 23.6 63.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#66
n-Nitrosodiphenylamine
Concen: 52.809 ng
RT: 15.874 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39


Tgt Ion:169 Resp: 2185624
Ion Ratio Lower Upper
169 100
168 67.4 53.6 80.4
167 37.0 29.4 44.0





#67
4-Bromophenyl-phenylether
Concen: 50.591 ng
RT: 16.545 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:248 Resp: 763962
Ion Ratio Lower Upper
248 100
250 96.7 78.0 117.0
141 69.6 55.1 82.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#68
Hexachlorobenzene
Concen: 55.626 ng
RT: 16.633 min Scan# 2286
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:284 Resp: 889381
Ion Ratio Lower Upper
284 100
142 32.7 24.8 37.2
249 29.3 24.0 36.0





#69

Atrazine

Concen: 59.906 ng

RT: 16.816 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD





Tgt Ion:200 Resp: 77315
 Ion Ratio Lower Upper
 200 100
 173 26.4 6.5 46.5
 215 48.1 28.9 68.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#70

Pentachlorophenol

Concen: 88.482 ng




RT: 16.992 min Scan# 2347


Delta R.T. 0.000 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Tgt Ion:266 Resp: 1068870
 Ion Ratio Lower Upper
 266 100
 268 64.0 50.5 75.7
 264 63.0 50.6 76.0









#71
Phenanthrene
Concen: 52.388 ng
RT: 17.410 min Scan# 2418
Delta R.T. 0.000 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Instrument :
BNA_M
ClientSampleId :
PB155451BSD

Tgt Ion:178 Resp: 382866
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.3 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#72
Anthracene
Concen: 52.363 ng
RT: 17.498 min Scan# 2433
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion:178 Resp: 3903775
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.2
179 15.7 12.5 18.7





#73

Carbazole

Concen: 52.553 ng

RT: 17.780 min Scan# 2482

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD

Tgt Ion:167 Resp: 3592508

Ion Ratio Lower Upper

167 100

166 21.6 17.5 26.3





139 12.7 10.0 15.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#74

Di-n-butylphthalate

Concen: 51.621 ng

RT: 18.298 min Scan# 2569

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:149 Resp: 4530295


Ion Ratio Lower Upper

149 100

150 9.5 7.4 11.2


104 5.4 4.2 6.4





#75
Fluoranthene
Concen: 50.963 ng
RT: 19.415 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:202 Resp: 4372629
Ion Ratio Lower Upper
202 100
101 10.4 0.0 30.5
203 17.6 0.0 37.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.533 min Scan# 3119
Delta R.T. -0.012 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:240 Resp: 1218985
Ion Ratio Lower Upper
240 100
120 11.2 8.1 12.1
236 25.6 20.7 31.1





#77

Benzidine

Concen: 56.623 ng

RT: 19.615 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD

Tgt Ion:184 Resp: 1408880

Ion Ratio Lower Upper

184 100

185 15.0 11.2 16.8





183 11.3 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#78

Pyrene

Concen: 53.346 ng

RT: 19.780 min Scan# 2821

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:202 Resp: 4524760


Ion Ratio Lower Upper

202 100

200 21.5 17.3 25.9

203 18.1 14.2 21.4





#79

Terphenyl-d14

Concen: 106.874 ng

RT: 19.974 min Scan# 2854

Delta R.T. 0.000 min

Lab File: BM041915.D


Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD



Tgt Ion:244 Resp: 7153050

Ion Ratio Lower Upper

244 100

212 6.8 5.7 8.5




122 8.6 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#80

Butylbenzylphthalate

Concen: 51.761 ng

RT: 20.656 min Scan# 2970

Delta R.T. 0.000 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:149 Resp: 1952656


Ion Ratio Lower Upper

149 100

91 73.3 61.7 92.5

206 21.9 16.9 25.3









#81
Benzo(a)anthracene
Concen: 51.891 ng
RT: 21.521 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Instrument :
BNA_M
ClientSampleId :
PB155451BSD

Tgt Ion:228 Resp: 412546
Ion Ratio Lower Upper
228 100
226 27.8 22.2 33.4
229 20.0 16.1 24.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#82
3,3'-Dichlorobenzidine
Concen: 45.614 ng
RT: 21.456 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion:252 Resp: 1277138
Ion Ratio Lower Upper
252 100
254 63.6 51.3 76.9
126 10.6 8.6 13.0





#83

Chrysene

Concen: 50.494 ng

RT: 21.574 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD

Tgt Ion:228 Resp: 3762790

Ion Ratio Lower Upper

228 100

226 30.8 24.4 36.6





229 19.9 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#84

Bis(2-ethylhexyl)phthalate

Concen: 51.055 ng

RT: 21.398 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:149 Resp: 2857286


Ion Ratio Lower Upper

149 100

167 28.8 23.0 34.6


279 4.5 3.7 5.5





#85
Di-n-octyl phthalate
Concen: 51.689 ng
RT: 22.327 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD



Tgt Ion:149 Resp: 4869340
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 11.0 8.7 13.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.950 min Scan# 3530
Delta R.T. -0.012 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion:264 Resp: 1218362
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 22.3 17.0 25.6





#87

Indeno(1,2,3-cd)pyrene

Concen: 53.532 ng

RT: 26.462 min Scan# 3957

Delta R.T. -0.018 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39

Instrument :

BNA_M

ClientSampleId :

PB155451BSD

Tgt Ion:276 Resp: 389415

Ion Ratio Lower Upper

276 100

138 24.6 19.8 29.8





277 25.1 20.2 30.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/20/2023

Supervised By :mohammad ahmed 09/21/2023



#88

Benzo(b)fluoranthene

Concen: 54.092 ng

RT: 23.209 min Scan# 3404

Delta R.T. -0.006 min

Lab File: BM041915.D

Acq: 19 Sep 2023 22:39




Tgt Ion:252 Resp: 3847659


Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6


125 9.3 7.8 11.6





#89
Benzo(k)fluoranthene
Concen: 52.273 ng
RT: 23.262 min Scan# 3414
Delta R.T. -0.006 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39




Instrument :
BNA_M
ClientSampleId :
PB155451BSD




Tgt Ion:252 Resp: 3674340
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 9.4 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023




#90
Benzo(a)pyrene
Concen: 48.948 ng
RT: 23.844 min Scan# 3512
Delta R.T. -0.012 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39



Tgt Ion:252 Resp: 3267113
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 10.9 8.6 12.8









#91
Dibenzo(a,h)anthracene
Concen: 53.926 ng
RT: 26.480 min Scan# 3960
Delta R.T. -0.018 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Instrument :
BNA_M
ClientSampleId :
PB155451BSD

Tgt Ion:278 Resp: 3270338
Ion Ratio Lower Upper
278 100
139 15.8 12.8 19.2
279 24.0 19.2 28.8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2023
Supervised By :mohammad ahmed 09/21/2023



#92
Benzo(g,h,i)perylene
Concen: 53.157 ng
RT: 27.250 min Scan# 4091
Delta R.T. -0.018 min
Lab File: BM041915.D
Acq: 19 Sep 2023 22:39

Tgt Ion:276 Resp: 3080169
Ion Ratio Lower Upper
276 100
277 24.0 19.4 29.2
138 21.1 17.0 25.4

