


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
 Data File : BM041911.D
 Acq On : 19 Sep 2023 20:12
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 19 04:49:34 2023






AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.0	264120	PASS
68	69	0.00	2	1.5	3657	PASS
69	198	0.00	100	37.8	249771	PASS
70	69	0.00	2	0.6	1442	PASS
127	198	10	80	41.8	275883	PASS
197	198	0.00	2	0.3	2142	PASS
198	198	100	100	100.0	660693	PASS
199	198	5	9	6.8	45157	PASS
275	198	10	60	24.9	164563	PASS
365	198	1	100	2.9	19038	PASS
441	198	0.01	100	13.9	92125	PASS
442	442	50	100	100.0	597461	PASS
443	442	15	24	19.5	116760	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
Data File : BM041911.D
Acq On : 19 Sep 2023 20:12
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Sep 20 05:55:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 19 04:49:34 2023
Response via : Initial Calibration



TIC: BM041911.D\data.ms

(70) Pentachlorophenol (C)

16.986min (-0.006) 220073.29 ng




response 807320

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	63.39
264.00	63.30	62.80
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091923\
Data File : BM041911.D
Acq On : 19 Sep 2023 20:12
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Sep 20 05:55:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 19 04:49:34 2023
Response via : Initial Calibration



TIC: BM041911.D\data.ms

(77) Benzidine

19.621min (+ 0.000) 228492.28 ng

response 4789898

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.30
183.00	11.40	11.53
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/19/2023	BNA_M	<u>BM041911.D</u>
Compound Name	Response	Retention Time
DDT	2257840	20.809
DDD	43460	20.374
DDE	2161	19.856
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
45621	2303461	1.98

Instrument :
BNA_M
ClientSampleId :
DFTPP