





Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM092019.M
Calibration Table Last Updated: Fri Sep 20 16:34:01 2019

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $1.79 \times 10^{-1} \times \text{Amt} - 6.23 \times 10^{-2}$
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM092019.M
Calibration Table Last Updated: Fri Sep 20 16:34:01 2019

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.17 \times 10^{-1} \times \text{Amt} - 3.89 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM092019.M
Calibration Table Last Updated: Fri Sep 20 16:34:01 2019