

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092121\
 Data File : BM032119.D
 Acq On : 21 Sep 2021 18:11
 Operator : CG/JU
 Sample : SSTD00525
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025

Quant Time: Sep 22 01:52:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 22 01:45:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	314375	20.000	ng/u1	0.00
20) Naphthalene-d8	10.463	136	1315448	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	845688	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.045	188	1674610	20.000	ng/u1	0.00
79) Chrysene-d12	21.203	240	1388442	20.000	ng/u1	0.00
88) Perylene-d12	23.374	264	1302588	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.993	96	15958	2.342	ng/uL	0.00
4) Pyridine-d5	3.416	84	116188	6.005	ng/u1	0.00
7) Phenol-d5	6.846	99	128614	4.934	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	79983	5.303	ng/u1	0.00
11) 2-Chlorophenol-d4	7.204	132	98153	4.893	ng/u1	0.00
15) 4-Methylphenol-d8	8.392	113	99287	4.439	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	34978	3.661	ng/u1	0.00
24) 2-Nitrophenol-d4	9.551	143	32273	2.854	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.092	165	92839	4.092	ng/u1	0.00
31) 4-Chloroaniline-d4	10.592	131	132451	4.139	ng/u1	0.00
46) Dimethylphthalate-d6	13.716	166	290020	4.532	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	378852	4.999	ng/u1	0.00
54) 4-Nitrophenol-d4	14.510	143	29288	2.529	ng/u1	0.00
60) Fluorene-d10	15.304	176	245396	4.362	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.416	200	19386	1.685	ng/u1	0.00
73) Anthracene-d10	17.139	188	376736	4.740	ng/u1	0.00
81) Pyrene-d10	19.421	212	380967	5.858	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.233	264	317459	4.637	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.028	88	21073	2.565	ng/uL#	85
5) Pyridine	3.440	79	118119	5.938	ng/u1#	80
6) Benzaldehyde	6.793	77	97091	8.227	ng/u1	91
8) Phenol	6.875	94	137277	5.233	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.087	93	111698	5.591	ng/u1	94
12) 2-Chlorophenol	7.240	128	105031	5.243	ng/u1	99
13) 2-Methylphenol	8.128	108	104416	5.162	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.198	45	138543	5.336	ng/u1#	89
16) Acetophenone	8.487	105	173663	4.958	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.475	70	91145	5.333	ng/u1#	87
18) 4-Methylphenol	8.451	108	109815	4.889	ng/u1	93
19) Hexachloroethane	8.763	117	45357	5.286	ng/u1	94
22) Nitrobenzene	8.863	77	102942	4.268	ng/u1	92
23) Isophorone	9.381	82	257650	5.804	ng/u1#	96
25) 2-Nitrophenol	9.581	139	37094	3.149	ng/u1	95
26) 2,4-Dimethylphenol	9.651	107	125924	5.252	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.869	93	151800	5.768	ng/u1	98
29) 2,4-Dichlorophenol	10.116	162	95838	4.441	ng/u1#	94
30) Naphthalene	10.516	128	341637	5.188	ng/u1	99
32) 4-Chloroaniline	10.616	127	142036	4.449	ng/u1	98
33) Hexachlorobutadiene	10.822	225	68863	4.810	ng/u1	96
34) Caprolactam	11.322	113	32802	4.560	ng/u1	96
35) 4-Chloro-3-methylphenol	11.751	107	111449	4.925	ng/u1	90
36) 2-Methylnaphthalene	12.128	142	234496	4.516	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092121\
 Data File : BM032119.D
 Acq On : 21 Sep 2021 18:11
 Operator : CG/JU
 Sample : SSTD00525
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025


Quant Time: Sep 22 01:52:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 22 01:45:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.351	142	236812	4.478	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.510	216	123277	5.078	ng/ul	98
40) Hexachlorocyclopentadiene	12.504	237	71557	5.476	ng/ul	98
41) 2,4,6-Trichlorophenol	12.751	196	74967	4.610	ng/ul	94
42) 2,4,5-Trichlorophenol	12.816	196	75275	4.262	ng/ul	98
43) 1,1'-Biphenyl	13.145	154	317407	5.336	ng/ul	98
44) 2-Chloronaphthalene	13.186	162	238284	5.056	ng/ul	97
45) 2-Nitroaniline	13.380	65	40209	3.097	ng/ul	98
47) Dimethylphthalate	13.763	163	295799	4.708	ng/ul	99
48) 2,6-Dinitrotoluene	13.874	165	33490	2.702	ng/ul	96
50) Acenaphthylene	14.033	152	398486	5.715	ng/ul	99
51) 3-Nitroaniline	14.204	138	35957	3.100	ng/ul#	84
52) Acenaphthene	14.380	153	258847	5.111	ng/ul	99
53) 2,4-Dinitrophenol	14.410	184	11488	1.325	ng/ul#	87
55) 4-Nitrophenol	14.521	109	31702	2.965	ng/ul	84
56) Dibenzofuran	14.710	168	362329	4.851	ng/ul	97
57) 2,4-Dinitrotoluene	14.663	165	43329	2.331	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	14.945	232	59733	3.768	ng/ul#	93
59) Diethylphthalate	15.127	149	304655	4.744	ng/ul	99
61) Fluorene	15.357	166	294654	4.663	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.351	204	144797	4.474	ng/ul	97
63) 4-Nitroaniline	15.363	138	40302	3.879	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.427	198	18590	1.662	ng/ul#	88
67) N-Nitrosodiphenylamine	15.563	169	255127	5.305	ng/ul	99
68) 4-Bromophenyl-phenylether	16.239	248	86544	5.043	ng/ul	93
69) Hexachlorobenzene	16.374	284	94367	4.736	ng/ul	96
70) Atrazine	16.498	200	89277	4.927	ng/ul	97
71) Pentachlorophenol	16.710	266	41829	3.210	ng/ul	96
72) Phenanthrene	17.086	178	446415	4.881	ng/ul	98
74) Anthracene	17.174	178	462894	4.945	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.116	216	122750	5.633	ng/ul	97
76) Pentachlorobenzene	14.645	250	119159	5.212	ng/ul	97
77) Carbazole	17.439	167	413128	4.822	ng/ul	99
78) Di-n-butylphthalate	17.998	149	496604	5.059	ng/ul	99
80) Fluoranthene	19.092	202	503619	6.373	ng/ul	99
82) Pyrene	19.451	202	509977	6.095	ng/ul	97
83) Butylbenzylphthalate	20.339	149	162925	4.959	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.121	252	138883	4.443	ng/ul	98
85) Benzo(a)anthracene	21.186	228	435617	5.044	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.115	149	249713	4.713	ng/ul	96
87) Chrysene	21.239	228	423807	5.102	ng/ul	99
89) Di-n-octyl phthalate	21.962	149	386766	4.352	ng/ul	100
90) Benzo(b)fluoranthene	22.727	252	397270	4.787	ng/ul	99
91) Benzo(k)fluoranthene	22.768	252	376353	4.684	ng/ul	99
93) Benzo(a)pyrene	23.274	252	380783	5.081	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.527	276	399952	4.170	ng/ul	98
95) Dibenzo(a,h)anthracene	25.533	278	338844	4.201	ng/ul	99
96) Benzo(g,h,i)perylene	26.191	276	335882	4.229	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 22 01:52:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092121.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 22 01:45:17 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.565 ng/uL
RT: 3.028 min Scan# 75
Delta R.T. 0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion: 88 Resp: 21073
Ion Ratio Lower Upper
88 100
43 63.4 31.5 47.3#
58 83.3 64.1 96.1



#5
Pyridine
Concen: 5.938 ng/uL
RT: 3.440 min Scan# 145
Delta R.T. 0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion: 79 Resp: 118119
Ion Ratio Lower Upper
79 100
52 65.7 69.9 104.9#
51 33.2 33.8 50.6#









#6
Benzaldehyde
Concen: 8.227 ng/ul
RT: 6.793 min Scan# 715
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion: 77 Resp: 97091
Ion Ratio Lower Upper
77 100
105 92.9 82.2 123.4
106 86.7 75.0 112.4



#8
Phenol
Concen: 5.233 ng/ul
RT: 6.875 min Scan# 729
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion: 94 Resp: 137277
Ion Ratio Lower Upper
94 100
65 29.2 24.2 36.4
66 37.5 31.4 47.0









#10
Bis(2-Chloroethyl)ether
Concen: 5.591 ng/ul
RT: 7.087 min Scan# 765
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:	93	Resp:	111698
Ion	Ratio	Lower	Upper
93	100		
63	78.7	68.0	102.0
95	33.8	25.8	38.6



#12
2-Chlorophenol
Concen: 5.243 ng/ul
RT: 7.240 min Scan# 791
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:	128	Resp:	105031
Ion	Ratio	Lower	Upper
128	100		
64	48.6	39.2	58.8
130	31.3	25.3	37.9









#13
2-Methylphenol
Concen: 5.162 ng/ul
RT: 8.128 min Scan# 942
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:108 Resp: 104416
Ion Ratio Lower Upper
108 100
107 88.1 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 5.336 ng/ul
RT: 8.198 min Scan# 954
Delta R.T. -0.012 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion: 45 Resp: 138543
Ion Ratio Lower Upper
45 100
77 16.6 10.2 15.4#
79 14.5 7.8 11.8#









#16
Acetophenone
Concen: 4.958 ng/ul
RT: 8.487 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:105 Resp: 173663
Ion Ratio Lower Upper
105 100
77 77.4 63.7 95.5
51 26.3 25.4 38.0



#17
N-Nitroso-di-n-propylamine
Concen: 5.333 ng/ul
RT: 8.475 min Scan# 1001
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion: 70 Resp: 91145
Ion Ratio Lower Upper
70 100
42 52.5 53.4 80.0#
101 10.5 7.4 11.0
130 19.7 16.8 25.2









#18
4-Methylphenol
Concen: 4.889 ng/ul
RT: 8.451 min Scan# 997
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:108 Resp: 109815
Ion Ratio Lower Upper
108 100
107 116.2 87.0 130.6



#19
Hexachloroethane
Concen: 5.286 ng/ul
RT: 8.763 min Scan# 1050
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:117 Resp: 45357
Ion Ratio Lower Upper
117 100
201 93.6 80.1 120.1
199 59.1 50.8 76.2









#22
Nitrobenzene
Concen: 4.268 ng/ul
RT: 8.863 min Scan# 1067
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion: 77 Resp: 102942
Ion Ratio Lower Upper
77 100
123 36.7 34.8 52.2
65 13.0 10.4 15.6



#23
Isophorone
Concen: 5.804 ng/ul
RT: 9.381 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion: 82 Resp: 257650
Ion Ratio Lower Upper
82 100
95 8.5 5.3 7.9#
138 15.6 13.6 20.4









#25
2-Nitrophenol
Concen: 3.149 ng/ul
RT: 9.581 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:139 Resp: 37094
Ion Ratio Lower Upper
139 100
65 51.6 43.7 65.5
109 33.3 24.0 36.0



#26
2,4-Dimethylphenol
Concen: 5.252 ng/ul
RT: 9.651 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:107 Resp: 125924
Ion Ratio Lower Upper
107 100
121 52.0 41.0 61.6
122 85.5 68.2 102.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 5.768 ng/ul
 RT: 9.869 min Scan# 1238
 Delta R.T. -0.006 min
 Lab File: BM032119.D
 Acq: 21 Sep 2021 18:11




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005025


Tgt Ion: 93	Resp: 151800
Ion Ratio	Lower Upper
93 100	
95 33.8	25.9 38.9
123 11.3	8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 4.441 ng/ul
 RT: 10.116 min Scan# 1280
 Delta R.T. -0.006 min
 Lab File: BM032119.D
 Acq: 21 Sep 2021 18:11

Tgt Ion: 162	Resp: 95838
Ion Ratio	Lower Upper
162 100	
164 63.3	51.6 77.4
98 43.6	28.2 42.4#









#30
Naphthalene
Concen: 5.188 ng/ul
RT: 10.516 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:128 Resp: 341637
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.8 10.6 15.8



#32
4-Chloroaniline
Concen: 4.449 ng/ul
RT: 10.616 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:127 Resp: 142036
Ion Ratio Lower Upper
127 100
129 30.8 25.6 38.4









#33
Hexachlorobutadiene
Concen: 4.810 ng/ul
RT: 10.822 min Scan# 1400
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:225 Resp: 68863
Ion Ratio Lower Upper
225 100
223 58.0 50.7 76.1
227 63.0 50.7 76.1



#34
Caprolactam
Concen: 4.560 ng/ul
RT: 11.322 min Scan# 1485
Delta R.T. -0.012 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:113 Resp: 32802
Ion Ratio Lower Upper
113 100
55 196.8 155.2 232.8
56 134.6 101.9 152.9









#35
4-Chloro-3-methylphenol
Concen: 4.925 ng/ul
RT: 11.751 min Scan# 1558
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:	107	Resp:	111449
Ion Ratio	Lower	Upper	
107	100		
144	23.8	22.3	33.5
142	72.4	65.6	98.4



#36
2-Methylnaphthalene
Concen: 4.516 ng/ul
RT: 12.128 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:	142	Resp:	234496
Ion Ratio	Lower	Upper	
142	100		
141	89.5	70.6	105.8









#37
1-Methylnaphthalene
Concen: 4.478 ng/ul
RT: 12.351 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:142 Resp: 236812
Ion Ratio Lower Upper
142 100
141 94.2 72.8 109.2
116 4.2 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 5.078 ng/ul
RT: 12.510 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:216 Resp: 123277
Ion Ratio Lower Upper
216 100
214 78.5 62.2 93.4
179 22.3 16.6 24.8
108 20.1 13.5 20.3









#40
Hexachlorocyclopentadiene
Concen: 5.476 ng/ul
RT: 12.504 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:237	Resp:	71557
Ion Ratio	Lower	Upper
237	100	
235	62.2	49.0 73.4
272	12.1	11.0 16.4



#41
2,4,6-Trichlorophenol
Concen: 4.610 ng/ul
RT: 12.751 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:196	Resp:	74967
Ion Ratio	Lower	Upper
196	100	
198	91.9	78.8 118.2
200	30.2	26.0 39.0









#42
2,4,5-Trichlorophenol
Concen: 4.262 ng/ul
RT: 12.816 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:196	Resp:	75275
Ion Ratio	Lower	Upper
196	100	
198	96.9	76.1 114.1
200	30.5	25.4 38.2



#43
1,1'-Biphenyl
Concen: 5.336 ng/ul
RT: 13.145 min Scan# 1795
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:154	Resp:	317407
Ion Ratio	Lower	Upper
154	100	
153	41.3	32.3 48.5
76	14.5	10.6 15.8









#44
2-Chloronaphthalene
Concen: 5.056 ng/ul
RT: 13.186 min Scan# 1802
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:162 Resp: 238284
Ion Ratio Lower Upper
162 100
164 32.7 26.2 39.4
127 40.0 29.4 44.0



#45
2-Nitroaniline
Concen: 3.097 ng/ul
RT: 13.380 min Scan# 1835
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion: 65 Resp: 40209
Ion Ratio Lower Upper
65 100
92 66.1 53.9 80.9
138 101.9 83.3 124.9









#47
Dimethylphthalate
Concen: 4.708 ng/ul
RT: 13.763 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:163 Resp: 295799
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 10.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 2.702 ng/ul
RT: 13.874 min Scan# 1919
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:165 Resp: 33490
Ion Ratio Lower Upper
165 100
89 63.0 48.6 72.8
121 25.7 18.1 27.1









#50
Acenaphthylene
Concen: 5.715 ng/ul
RT: 14.033 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:152 Resp: 398486
Ion Ratio Lower Upper
152 100
151 20.7 16.2 24.2
153 12.9 10.2 15.4



#51
3-Nitroaniline
Concen: 3.100 ng/ul
RT: 14.204 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:138 Resp: 35957
Ion Ratio Lower Upper
138 100
108 36.7 7.4 11.2#
92 132.9 96.5 144.7









#52
Acenaphthene
Concen: 5.111 ng/ul
RT: 14.380 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:153 Resp: 258847
Ion Ratio Lower Upper
153 100
152 46.5 37.4 56.2
154 87.8 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 1.325 ng/ul
RT: 14.410 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:184 Resp: 11488
Ion Ratio Lower Upper
184 100
63 63.7 56.9 85.3
154 75.3 49.0 73.6#









#55
4-Nitrophenol
Concen: 2.965 ng/ul
RT: 14.521 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:109 Resp: 31702
Ion Ratio Lower Upper
109 100
139 108.2 100.3 150.5
65 111.5 104.6 156.8



#56
Dibenzofuran
Concen: 4.851 ng/ul
RT: 14.710 min Scan# 2061
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:168 Resp: 362329
Ion Ratio Lower Upper
168 100
139 38.7 29.5 44.3









#57
2,4-Dinitrotoluene
Concen: 2.331 ng/ul
RT: 14.663 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:165 Resp: 43329
Ion Ratio Lower Upper
165 100
63 47.8 37.6 56.4
182 4.3 1.8 2.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.768 ng/ul
RT: 14.945 min Scan# 2101
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:232 Resp: 59733
Ion Ratio Lower Upper
232 100
131 46.6 35.0 52.6
130 3.5 1.7 2.5#
166 36.3 24.4 36.6









#59
Diethylphthalate
Concen: 4.744 ng/ul
RT: 15.127 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:149 Resp: 304655
Ion Ratio Lower Upper
149 100
177 21.9 17.0 25.4
150 12.7 9.9 14.9



#61
Fluorene
Concen: 4.663 ng/ul
RT: 15.357 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:166 Resp: 294654
Ion Ratio Lower Upper
166 100
165 96.6 78.4 117.6
167 13.4 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 4.474 ng/ul
RT: 15.351 min Scan# 2170
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:	204	Resp:	144797
Ion Ratio	Lower	Upper	
204	100		
206	32.6	26.0	39.0
141	59.8	45.6	68.4



#63
4-Nitroaniline
Concen: 3.879 ng/ul
RT: 15.363 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:	138	Resp:	40302
Ion Ratio	Lower	Upper	
138	100		
92	57.9	41.5	62.3
108	100.6	61.4	92.2#









#66
4,6-Dinitro-2-methylphenol
Concen: 1.662 ng/ul
RT: 15.427 min Scan# 2183
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:198 Resp: 18590
Ion Ratio Lower Upper
198 100
182 4.6 3.8 5.6
77 32.3 20.1 30.1#



#67
N-Nitrosodiphenylamine
Concen: 5.305 ng/ul
RT: 15.563 min Scan# 2206
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:169 Resp: 255127
Ion Ratio Lower Upper
169 100
168 68.3 53.8 80.6
167 36.7 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 5.043 ng/ul
RT: 16.239 min Scan# 2321
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:	248	Resp:	86544
Ion Ratio	Lower	Upper	
248	100		
250	97.4	78.6	118.0
141	83.6	56.2	84.2



#69
Hexachlorobenzene
Concen: 4.736 ng/ul
RT: 16.374 min Scan# 2344
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:	284	Resp:	94367
Ion Ratio	Lower	Upper	
284	100		
142	37.8	29.0	43.4
249	34.5	25.0	37.4









#70
Atrazine
Concen: 4.927 ng/ul
RT: 16.498 min Scan# 2365
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:200 Resp: 89277
Ion Ratio Lower Upper
200 100
173 26.0 18.7 28.1
215 47.3 38.9 58.3



#71
Pentachlorophenol
Concen: 3.210 ng/ul
RT: 16.710 min Scan# 2401
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:266 Resp: 41829
Ion Ratio Lower Upper
266 100
264 66.1 48.2 72.4
268 64.5 51.6 77.4









#72
Phenanthrene
Concen: 4.881 ng/ul
RT: 17.086 min Scan# 2465
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:178 Resp: 446415
Ion Ratio Lower Upper
178 100
179 15.5 12.1 18.1
176 19.9 15.2 22.8



#74
Anthracene
Concen: 4.945 ng/ul
RT: 17.174 min Scan# 2480
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:178 Resp: 462894
Ion Ratio Lower Upper
178 100
179 14.9 12.1 18.1
176 18.3 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 5.633 ng/uL
RT: 13.116 min Scan# 1790
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:216 Resp: 122750
Ion Ratio Lower Upper
216 100
214 80.1 61.9 92.9
218 47.3 39.5 59.3



#76
Pentachlorobenzene
Concen: 5.212 ng/uL
RT: 14.645 min Scan# 2050
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:250 Resp: 119159
Ion Ratio Lower Upper
250 100
248 61.7 50.2 75.4
252 61.7 52.2 78.2









#77
 Carbazole
 Concen: 4.822 ng/ul
 RT: 17.439 min Scan# 2525
 Delta R.T. -0.000 min
 Lab File: BM032119.D
 Acq: 21 Sep 2021 18:11




Instrument :
 BNA_M
 ClientSampleId :
 SST005025


Tgt Ion:167	Resp: 413128
Ion Ratio	Lower Upper
167 100	
166 22.1	17.0 25.6
139 12.7	9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 5.059 ng/ul
 RT: 17.998 min Scan# 2620
 Delta R.T. -0.000 min
 Lab File: BM032119.D
 Acq: 21 Sep 2021 18:11

Tgt Ion:149	Resp: 496604		
Ion Ratio	Lower	Upper	
149 100			
150 8.6	7.3	10.9	
104 5.2	4.1	6.1	









#80
Fluoranthene
Concen: 6.373 ng/u1
RT: 19.092 min Scan# 2806
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:202 Resp: 503619
Ion Ratio Lower Upper
202 100
101 11.6 8.8 13.2
100 8.9 6.8 10.2



#82
Pyrene
Concen: 6.095 ng/u1
RT: 19.451 min Scan# 2867
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:202 Resp: 509977
Ion Ratio Lower Upper
202 100
101 14.3 10.0 15.0
100 11.2 8.5 12.7









#83
Butylbenzylphthalate
Concen: 4.959 ng/ul
RT: 20.339 min Scan# 3018
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:	149	Resp:	162925
Ion Ratio	Lower	Upper	
149	100		
91	75.8	57.4	86.2
206	20.2	15.9	23.9



#84
3,3'-Dichlorobenzidine
Concen: 4.443 ng/ul
RT: 21.121 min Scan# 3151
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:	252	Resp:	138883
Ion Ratio	Lower	Upper	
252	100		
254	64.2	49.9	74.9
126	11.4	8.9	13.3









#85
Benzo(a)anthracene
Concen: 5.044 ng/ul
RT: 21.186 min Scan# 3162
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:228 Resp: 435617
Ion Ratio Lower Upper
228 100
229 19.7 16.0 24.0
226 27.7 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 4.713 ng/ul
RT: 21.115 min Scan# 3150
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:149 Resp: 249713
Ion Ratio Lower Upper
149 100
167 28.0 20.7 31.1
279 4.2 3.0 4.4









#87
Chrysene
Concen: 5.102 ng/ul
RT: 21.239 min Scan# 3171
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:	228	Resp:	423807
Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.6	35.4
229	19.8	15.9	23.9



#89
Di-n-octyl phthalate
Concen: 4.352 ng/ul
RT: 21.962 min Scan# 3294
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:149 Resp: 386766









#90
Benzo(b)fluoranthene
Concen: 4.787 ng/ul
RT: 22.727 min Scan# 3424
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:252 Resp: 397270
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 10.1 7.7 11.5



#91
Benzo(k)fluoranthene
Concen: 4.684 ng/ul
RT: 22.768 min Scan# 3431
Delta R.T. -0.000 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:252 Resp: 376353
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 10.7 7.9 11.9









#93
Benzo(a)pyrene
Concen: 5.081 ng/ul
RT: 23.274 min Scan# 3517
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11




Instrument :
BNA_M
ClientSampleId :
SSTD005025


Tgt Ion:252 Resp: 380783
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 12.6 8.6 13.0



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.170 ng/ul
RT: 25.527 min Scan# 3900
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:276 Resp: 399952
Ion Ratio Lower Upper
276 100
138 26.0 19.5 29.3
277 24.8 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 4.201 ng/ul
RT: 25.533 min Scan# 3901
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Instrument :
BNA_M
ClientSampleId :
SSTD005025

Tgt Ion:278 Resp: 338844
Ion Ratio Lower Upper
278 100
139 17.0 12.8 19.2
279 24.1 19.1 28.7



#96
Benzo(g,h,i)perylene
Concen: 4.229 ng/ul
RT: 26.191 min Scan# 4013
Delta R.T. -0.006 min
Lab File: BM032119.D
Acq: 21 Sep 2021 18:11

Tgt Ion:276 Resp: 335882
Ion Ratio Lower Upper
276 100
138 23.1 16.5 24.7
277 24.0 18.5 27.7

