

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092121\
 Data File : BM032120.D
 Acq On : 21 Sep 2021 18:53
 Operator : CG/JU
 Sample : SSTD01026
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026

Quant Time: Sep 22 01:53:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 22 01:45:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	148377	20.000	ng/u1	0.00
20) Naphthalene-d8	10.469	136	696718	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	488664	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.045	188	1070183	20.000	ng/u1	0.00
79) Chrysene-d12	21.203	240	1144073	20.000	ng/u1	0.00
88) Perylene-d12	23.380	264	1265069	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.987	96	15196	4.725	ng/uL	0.00
4) Pyridine-d5	3.411	84	110404	12.090	ng/u1	0.00
7) Phenol-d5	6.846	99	125474	10.198	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	79778	11.206	ng/u1	0.00
11) 2-Chlorophenol-d4	7.204	132	93303	9.854	ng/u1	0.00
15) 4-Methylphenol-d8	8.393	113	103481	9.803	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	36477	7.208	ng/u1	0.00
24) 2-Nitrophenol-d4	9.551	143	34100	5.694	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.092	165	97875	8.146	ng/u1	0.00
31) 4-Chloroaniline-d4	10.592	131	146472	8.642	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	341610	9.239	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	427529	9.763	ng/u1	0.00
54) 4-Nitrophenol-d4	14.510	143	42324	6.324	ng/u1	0.00
60) Fluorene-d10	15.304	176	284595	8.756	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.416	200	27808	3.782	ng/u1	0.00
73) Anthracene-d10	17.139	188	471177	9.276	ng/u1	0.00
81) Pyrene-d10	19.427	212	536042	10.003	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.239	264	622869	9.368	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.022	88	15883	4.097	ng/uL#	87
5) Pyridine	3.434	79	111150	11.839	ng/u1	84
6) Benzaldehyde	6.793	77	92168	16.547	ng/u1	91
8) Phenol	6.875	94	130965	10.578	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.087	93	105211	11.157	ng/u1	93
12) 2-Chlorophenol	7.240	128	96997	10.259	ng/u1	97
13) 2-Methylphenol	8.128	108	98871	10.357	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.204	45	136214	11.115	ng/u1#	90
16) Acetophenone	8.487	105	173614	10.501	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.475	70	97629	12.104	ng/u1#	86
18) 4-Methylphenol	8.451	108	108611	10.245	ng/u1	88
19) Hexachloroethane	8.763	117	39644	9.788	ng/u1	94
22) Nitrobenzene	8.863	77	101738	7.963	ng/u1	94
23) Isophorone	9.387	82	280673	11.937	ng/u1	98
25) 2-Nitrophenol	9.581	139	37369	5.990	ng/u1	95
26) 2,4-Dimethylphenol	9.651	107	128540	10.122	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.869	93	157800	11.321	ng/u1	99
29) 2,4-Dichlorophenol	10.116	162	93526	8.183	ng/u1#	96
30) Naphthalene	10.516	128	337489	9.676	ng/u1	98
32) 4-Chloroaniline	10.616	127	153621	9.086	ng/u1	99
33) Hexachlorobutadiene	10.828	225	61684	8.136	ng/u1	100
34) Caprolactam	11.328	113	41529	10.901	ng/u1	93
35) 4-Chloro-3-methylphenol	11.757	107	121986	10.178	ng/u1	94
36) 2-Methylnaphthalene	12.128	142	241097	8.767	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092121\
 Data File : BM032120.D
 Acq On : 21 Sep 2021 18:53
 Operator : CG/JU
 Sample : SST01026
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010026


Quant Time: Sep 22 01:53:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 22 01:45:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.351	142	247623	8.840	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.510	216	127021	9.055	ng/ul	96
40) Hexachlorocyclopentadiene	12.504	237	71503	9.470	ng/ul	98
41) 2,4,6-Trichlorophenol	12.751	196	80159	8.530	ng/ul	99
42) 2,4,5-Trichlorophenol	12.822	196	84093	8.240	ng/ul	99
43) 1,1'-Biphenyl	13.145	154	339068	9.864	ng/ul	99
44) 2-Chloronaphthalene	13.186	162	255440	9.380	ng/ul	96
45) 2-Nitroaniline	13.386	65	48037	6.403	ng/ul	96
47) Dimethylphthalate	13.763	163	343176	9.452	ng/ul	100
48) 2,6-Dinitrotoluene	13.875	165	40529	5.658	ng/ul#	91
50) Acenaphthylene	14.039	152	443468	11.006	ng/ul	99
51) 3-Nitroaniline	14.210	138	45083	6.726	ng/ul#	85
52) Acenaphthene	14.380	153	287731	9.832	ng/ul	99
53) 2,4-Dinitrophenol	14.410	184	17943	3.583	ng/ul	91
55) 4-Nitrophenol	14.522	109	44252	7.162	ng/ul	92
56) Dibenzofuran	14.716	168	404420	9.371	ng/ul	97
57) 2,4-Dinitrotoluene	14.663	165	57184	5.323	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	14.951	232	73672	8.042	ng/ul#	98
59) Diethylphthalate	15.127	149	360240	9.708	ng/ul	100
61) Fluorene	15.363	166	336111	9.204	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.351	204	161646	8.643	ng/ul	94
63) 4-Nitroaniline	15.363	138	53518	8.913	ng/ul#	83
66) 4,6-Dinitro-2-methylph...	15.427	198	28058	3.924	ng/ul#	87
67) N-Nitrosodiphenylamine	15.563	169	295314	9.609	ng/ul	99
68) 4-Bromophenyl-phenylether	16.239	248	98643	8.995	ng/ul	91
69) Hexachlorobenzene	16.374	284	111788	8.779	ng/ul	94
70) Atrazine	16.504	200	114866	9.920	ng/ul	98
71) Pentachlorophenol	16.710	266	57080	6.855	ng/ul	97
72) Phenanthrene	17.086	178	549399	9.399	ng/ul	99
74) Anthracene	17.174	178	563679	9.423	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.122	216	133979	9.621	ng/ul	99
76) Pentachlorobenzene	14.651	250	136417	9.337	ng/ul	98
77) Carbazole	17.439	167	527411	9.633	ng/ul	98
78) Di-n-butylphthalate	17.998	149	638702	10.181	ng/ul	99
80) Fluoranthene	19.092	202	671133	10.307	ng/ul	97
82) Pyrene	19.457	202	688131	9.982	ng/ul	97
83) Butylbenzylphthalate	20.345	149	264155	9.758	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.121	252	241695	9.384	ng/ul	98
85) Benzo(a)anthracene	21.186	228	691894	9.722	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.115	149	427996	9.804	ng/ul	98
87) Chrysene	21.239	228	681545	9.957	ng/ul	100
89) Di-n-octyl phthalate	21.968	149	740156	8.576	ng/ul	100
90) Benzo(b)fluoranthene	22.727	252	755701	9.376	ng/ul	98
91) Benzo(k)fluoranthene	22.768	252	676719	8.673	ng/ul	99
93) Benzo(a)pyrene	23.280	252	726588	9.983	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.533	276	834526	8.959	ng/ul	97
95) Dibenzo(a,h)anthracene	25.538	278	712209	9.091	ng/ul	99
96) Benzo(g,h,i)perylene	26.197	276	713108	9.245	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 22 01:53:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092121.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 22 01:45:17 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.097 ng/uL
RT: 3.022 min Scan# 74
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 88 Resp: 15883
Ion Ratio Lower Upper
88 100
43 58.4 31.5 47.3#
58 76.0 64.1 96.1



#5
Pyridine
Concen: 11.839 ng/uL
RT: 3.434 min Scan# 144
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion: 79 Resp: 111150
Ion Ratio Lower Upper
79 100
52 71.6 69.9 104.9
51 34.4 33.8 50.6









#6
Benzaldehyde
Concen: 16.547 ng/ul
RT: 6.793 min Scan# 715
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 77 Resp: 92168
Ion Ratio Lower Upper
77 100
105 91.6 82.2 123.4
106 86.2 75.0 112.4



#8
Phenol
Concen: 10.578 ng/ul
RT: 6.875 min Scan# 729
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion: 94 Resp: 130965
Ion Ratio Lower Upper
94 100
65 29.0 24.2 36.4
66 38.1 31.4 47.0









#10
Bis(2-Chloroethyl)ether
Concen: 11.157 ng/ul
RT: 7.087 min Scan# 765
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 93	Resp: 105211
Ion Ratio	Lower Upper
93 100	
63 78.0	68.0 102.0
95 35.8	25.8 38.6



#12
2-Chlorophenol
Concen: 10.259 ng/ul
RT: 7.240 min Scan# 791
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:128	Resp: 96997
Ion Ratio	Lower Upper
128 100	
64 46.9	39.2 58.8
130 30.7	25.3 37.9









#13
2-Methylphenol
Concen: 10.357 ng/ul
RT: 8.128 min Scan# 942
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:108 Resp: 98871
Ion Ratio Lower Upper
108 100
107 94.2 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 11.115 ng/ul
RT: 8.204 min Scan# 955
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion: 45 Resp: 136214
Ion Ratio Lower Upper
45 100
77 16.8 10.2 15.4#
79 13.4 7.8 11.8#









#16
Acetophenone
Concen: 10.501 ng/ul
RT: 8.487 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:105 Resp: 173614
Ion Ratio Lower Upper
105 100
77 78.7 63.7 95.5
51 27.0 25.4 38.0



#17
N-Nitroso-di-n-propylamine
Concen: 12.104 ng/ul
RT: 8.475 min Scan# 1001
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion: 70 Resp: 97629
Ion Ratio Lower Upper
70 100
42 52.1 53.4 80.0#
101 9.0 7.4 11.0
130 18.7 16.8 25.2









#18
4-Methylphenol
Concen: 10.245 ng/ul
RT: 8.451 min Scan# 997
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:108 Resp: 108611
Ion Ratio Lower Upper
108 100
107 121.0 87.0 130.6



#19
Hexachloroethane
Concen: 9.788 ng/ul
RT: 8.763 min Scan# 1050
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:117 Resp: 39644
Ion Ratio Lower Upper
117 100
201 92.8 80.1 120.1
199 60.2 50.8 76.2









#22
Nitrobenzene
Concen: 7.963 ng/ul
RT: 8.863 min Scan# 1067
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion: 77 Resp: 101738
Ion Ratio Lower Upper
77 100
123 39.4 34.8 52.2
65 15.2 10.4 15.6



#23
Isophorone
Concen: 11.937 ng/ul
RT: 9.387 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion: 82 Resp: 280673
Ion Ratio Lower Upper
82 100
95 7.6 5.3 7.9
138 16.1 13.6 20.4









#25
2-Nitrophenol
Concen: 5.990 ng/ul
RT: 9.581 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:139 Resp: 37369
Ion Ratio Lower Upper
139 100
65 52.6 43.7 65.5
109 35.0 24.0 36.0



#26
2,4-Dimethylphenol
Concen: 10.122 ng/ul
RT: 9.651 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:107 Resp: 128540
Ion Ratio Lower Upper
107 100
121 51.2 41.0 61.6
122 83.6 68.2 102.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 11.321 ng/ul
 RT: 9.869 min Scan# 1238
 Delta R.T. -0.006 min
 Lab File: BM032120.D
 Acq: 21 Sep 2021 18:53




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010026


Tgt Ion: 93	Resp: 157800		
Ion Ratio	Lower	Upper	
93	100		
95	32.7	25.9	38.9
123	9.8	8.5	12.7



#29
 2,4-Dichlorophenol
 Concen: 8.183 ng/ul
 RT: 10.116 min Scan# 1280
 Delta R.T. -0.006 min
 Lab File: BM032120.D
 Acq: 21 Sep 2021 18:53

Tgt Ion:162	Resp: 93526
Ion Ratio	Lower Upper
162 100	
164 64.6	51.6 77.4
98 42.5	28.2 42.4#









#30
Naphthalene
Concen: 9.676 ng/ul
RT: 10.516 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:128 Resp: 337489
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.4
127 14.3 10.6 15.8



#32
4-Chloroaniline
Concen: 9.086 ng/ul
RT: 10.616 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:127 Resp: 153621
Ion Ratio Lower Upper
127 100
129 32.8 25.6 38.4









#33
Hexachlorobutadiene
Concen: 8.136 ng/ul
RT: 10.828 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:225 Resp: 61684
Ion Ratio Lower Upper
225 100
223 63.4 50.7 76.1
227 63.8 50.7 76.1



#34
Caprolactam
Concen: 10.901 ng/ul
RT: 11.328 min Scan# 1486
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:113 Resp: 41529
Ion Ratio Lower Upper
113 100
55 183.8 155.2 232.8
56 134.7 101.9 152.9









#35
4-Chloro-3-methylphenol
Concen: 10.178 ng/ul
RT: 11.757 min Scan# 1559
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:	107	Resp:	121986
Ion Ratio	Lower	Upper	
107	100		
144	24.0	22.3	33.5
142	76.5	65.6	98.4



#36
2-Methylnaphthalene
Concen: 8.767 ng/ul
RT: 12.128 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:	142	Resp:	241097
Ion Ratio	Lower	Upper	
142	100		
141	89.2	70.6	105.8









#37
1-Methylnaphthalene
Concen: 8.840 ng/ul
RT: 12.351 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:142 Resp: 247623
Ion Ratio Lower Upper
142 100
141 92.9 72.8 109.2
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.055 ng/ul
RT: 12.510 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:216 Resp: 127021
Ion Ratio Lower Upper
216 100
214 74.0 62.2 93.4
179 20.9 16.6 24.8
108 20.0 13.5 20.3









#40
Hexachlorocyclopentadiene
Concen: 9.470 ng/ul
RT: 12.504 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:237 Resp: 71503
Ion Ratio Lower Upper
237 100
235 62.4 49.0 73.4
272 11.1 11.0 16.4



#41
2,4,6-Trichlorophenol
Concen: 8.530 ng/ul
RT: 12.751 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:196 Resp: 80159
Ion Ratio Lower Upper
196 100
198 97.5 78.8 118.2
200 31.9 26.0 39.0









#42
2,4,5-Trichlorophenol
Concen: 8.240 ng/ul
RT: 12.822 min Scan# 1740
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:196 Resp: 84093
Ion Ratio Lower Upper
196 100
198 95.1 76.1 114.1
200 30.5 25.4 38.2



#43
1,1'-Biphenyl
Concen: 9.864 ng/ul
RT: 13.145 min Scan# 1795
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:154 Resp: 339068
Ion Ratio Lower Upper
154 100
153 40.7 32.3 48.5
76 14.3 10.6 15.8









#44
2-Chloronaphthalene
Concen: 9.380 ng/ul
RT: 13.186 min Scan# 1802
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:162 Resp: 255440
Ion Ratio Lower Upper
162 100
164 33.3 26.2 39.4
127 40.3 29.4 44.0



#45
2-Nitroaniline
Concen: 6.403 ng/ul
RT: 13.386 min Scan# 1836
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion: 65 Resp: 48037
Ion Ratio Lower Upper
65 100
92 71.8 53.9 80.9
138 100.6 83.3 124.9









#47
Dimethylphthalate
Concen: 9.452 ng/ul
RT: 13.763 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:163 Resp: 343176
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 5.658 ng/ul
RT: 13.875 min Scan# 1919
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:165 Resp: 40529
Ion Ratio Lower Upper
165 100
89 67.3 48.6 72.8
121 27.6 18.1 27.1#









#50
Acenaphthylene
Concen: 11.006 ng/ul
RT: 14.039 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:152 Resp: 443468
Ion Ratio Lower Upper
152 100
151 20.6 16.2 24.2
153 12.5 10.2 15.4



#51
3-Nitroaniline
Concen: 6.726 ng/ul
RT: 14.210 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:138 Resp: 45083
Ion Ratio Lower Upper
138 100
108 43.6 7.4 11.2#
92 130.7 96.5 144.7









#52
Acenaphthene
Concen: 9.832 ng/ul
RT: 14.380 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:153 Resp: 287731
Ion Ratio Lower Upper
153 100
152 47.6 37.4 56.2
154 86.4 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 3.583 ng/ul
RT: 14.410 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:184 Resp: 17943
Ion Ratio Lower Upper
184 100
63 60.6 56.9 85.3
154 64.8 49.0 73.6









#55
4-Nitrophenol
Concen: 7.162 ng/ul
RT: 14.522 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:109 Resp: 44252
Ion Ratio Lower Upper
109 100
139 117.6 100.3 150.5
65 120.5 104.6 156.8



#56
Dibenzofuran
Concen: 9.371 ng/ul
RT: 14.716 min Scan# 2062
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:168 Resp: 404420
Ion Ratio Lower Upper
168 100
139 38.8 29.5 44.3









#57
2,4-Dinitrotoluene
Concen: 5.323 ng/ul
RT: 14.663 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:165 Resp: 57184
Ion Ratio Lower Upper
165 100
63 45.3 37.6 56.4
182 4.9 1.8 2.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.042 ng/ul
RT: 14.951 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:232 Resp: 73672
Ion Ratio Lower Upper
232 100
131 45.3 35.0 52.6
130 2.7 1.7 2.5#
166 30.1 24.4 36.6





#59

Diethylphthalate

Concen: 9.708 ng/ul

RT: 15.127 min Scan# 2132

Delta R.T. -0.006 min

Lab File: BM032120.D

Acq: 21 Sep 2021 18:53





Instrument :

BNA_M

ClientSampleId :

SSTD010026

Tgt Ion:149 Resp: 360240
 Ion Ratio Lower Upper
 149 100
 177 21.0 17.0 25.4
 150 12.5 9.9 14.9



#61

Fluorene

Concen: 9.204 ng/ul




RT: 15.363 min Scan# 2172


Delta R.T. -0.000 min

Lab File: BM032120.D

Acq: 21 Sep 2021 18:53

Tgt Ion:166 Resp: 336111
 Ion Ratio Lower Upper
 166 100
 165 99.1 78.4 117.6
 167 13.0 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 8.643 ng/ul
RT: 15.351 min Scan# 2170
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:204 Resp: 161646
Ion Ratio Lower Upper
204 100
206 33.0 26.0 39.0
141 63.4 45.6 68.4



#63
4-Nitroaniline
Concen: 8.913 ng/ul
RT: 15.363 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:138 Resp: 53518
Ion Ratio Lower Upper
138 100
92 59.1 41.5 62.3
108 95.6 61.4 92.2#









#66
4,6-Dinitro-2-methylphenol
Concen: 3.924 ng/ul
RT: 15.427 min Scan# 2183
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:198 Resp: 28058
Ion Ratio Lower Upper
198 100
182 6.1 3.8 5.6#
77 32.2 20.1 30.1#



#67
N-Nitrosodiphenylamine
Concen: 9.609 ng/ul
RT: 15.563 min Scan# 2206
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:169 Resp: 295314
Ion Ratio Lower Upper
169 100
168 67.2 53.8 80.6
167 36.9 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 8.995 ng/ul
RT: 16.239 min Scan# 2321
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:	248	Resp:	98643
Ion Ratio	Lower	Upper	
248	100		
250	102.1	78.6	118.0
141	84.1	56.2	84.2



#69
Hexachlorobenzene
Concen: 8.779 ng/ul
RT: 16.374 min Scan# 2344
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:	284	Resp:	111788
Ion Ratio	Lower	Upper	
284	100		
142	41.0	29.0	43.4
249	32.9	25.0	37.4









#70
Atrazine
Concen: 9.920 ng/ul
RT: 16.504 min Scan# 2366
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:200 Resp: 114866
Ion Ratio Lower Upper
200 100
173 24.9 18.7 28.1
215 47.7 38.9 58.3



#71
Pentachlorophenol
Concen: 6.855 ng/ul
RT: 16.710 min Scan# 2401
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:266 Resp: 57080
Ion Ratio Lower Upper
266 100
264 62.4 48.2 72.4
268 61.4 51.6 77.4









#72
Phenanthrene
Concen: 9.399 ng/ul
RT: 17.086 min Scan# 2465
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:178 Resp: 549399
Ion Ratio Lower Upper
178 100
179 15.3 12.1 18.1
176 19.3 15.2 22.8



#74
Anthracene
Concen: 9.423 ng/ul
RT: 17.174 min Scan# 2480
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:178 Resp: 563679
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 18.5 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.621 ng/uL
RT: 13.122 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:	216	Resp:	133979
Ion Ratio	Lower	Upper	
216	100		
214	77.3	61.9	92.9
218	47.8	39.5	59.3



#76
Pentachlorobenzene
Concen: 9.337 ng/uL
RT: 14.651 min Scan# 2051
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:	250	Resp:	136417
Ion Ratio	Lower	Upper	
250	100		
248	63.1	50.2	75.4
252	62.9	52.2	78.2









#77
 Carbazole
 Concen: 9.633 ng/ul
 RT: 17.439 min Scan# 2525
 Delta R.T. -0.000 min
 Lab File: BM032120.D
 Acq: 21 Sep 2021 18:53




Instrument :
 BNA_M
 ClientSampleId :
 SST010026


Tgt Ion:	167	Resp:	527411
Ion Ratio	Lower	Upper	
167	100		
166	22.4	17.0	25.6
139	13.0	9.9	14.9



#78
 Di-n-butylphthalate
 Concen: 10.181 ng/ul
 RT: 17.998 min Scan# 2620
 Delta R.T. -0.000 min
 Lab File: BM032120.D
 Acq: 21 Sep 2021 18:53

Tgt Ion:	149	Resp:	638702
Ion Ratio	Lower	Upper	
149	100		
150	8.9	7.3	10.9
104	5.6	4.1	6.1









#80
Fluoranthene
Concen: 10.307 ng/ul
RT: 19.092 min Scan# 2806
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:202 Resp: 671133
Ion Ratio Lower Upper
202 100
101 12.0 8.8 13.2
100 9.9 6.8 10.2



#82
Pyrene
Concen: 9.982 ng/ul
RT: 19.457 min Scan# 2868
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:202 Resp: 688131
Ion Ratio Lower Upper
202 100
101 14.0 10.0 15.0
100 11.3 8.5 12.7









#83
Butylbenzylphthalate
Concen: 9.758 ng/ul
RT: 20.345 min Scan# 3019
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:149 Resp: 264155
Ion Ratio Lower Upper
149 100
91 72.0 57.4 86.2
206 20.0 15.9 23.9



#84
3,3'-Dichlorobenzidine
Concen: 9.384 ng/ul
RT: 21.121 min Scan# 3151
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:252 Resp: 241695
Ion Ratio Lower Upper
252 100
254 63.7 49.9 74.9
126 12.0 8.9 13.3









#85
Benzo(a)anthracene
Concen: 9.722 ng/ul
RT: 21.186 min Scan# 3162
Delta R.T. -0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:228 Resp: 691894
Ion Ratio Lower Upper
228 100
229 19.8 16.0 24.0
226 26.9 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 9.804 ng/ul
RT: 21.115 min Scan# 3150
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:149 Resp: 427996
Ion Ratio Lower Upper
149 100
167 26.9 20.7 31.1
279 3.5 3.0 4.4









#87
Chrysene
Concen: 9.957 ng/ul
RT: 21.239 min Scan# 3171
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Instrument :
BNA_M
ClientSampleId :
SSTD010026




Tgt Ion:228 Resp: 681545


Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.6	35.4
229	19.7	15.9	23.9



#89
Di-n-octyl phthalate
Concen: 8.576 ng/ul
RT: 21.968 min Scan# 3295
Delta R.T. 0.006 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:149 Resp: 740156









#90
Benzo(b)fluoranthene
Concen: 9.376 ng/ul
RT: 22.727 min Scan# 3424
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:	252	Resp:	755701
Ion Ratio	Lower	Upper	
252	100		
253	21.7	17.8	26.6
125	10.7	7.7	11.5



#91
Benzo(k)fluoranthene
Concen: 8.673 ng/ul
RT: 22.768 min Scan# 3431
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:	252	Resp:	676719
Ion Ratio	Lower	Upper	
252	100		
253	22.0	17.4	26.2
125	11.2	7.9	11.9









#93
Benzo(a)pyrene
Concen: 9.983 ng/ul
RT: 23.280 min Scan# 3518
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53




Instrument :
BNA_M
ClientSampleId :
SSTD010026


Tgt Ion:252 Resp: 726588
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 11.9 8.6 13.0



#94
Indeno(1,2,3-cd)pyrene
Concen: 8.959 ng/ul
RT: 25.533 min Scan# 3901
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:276 Resp: 834526
Ion Ratio Lower Upper
276 100
138 26.5 19.5 29.3
277 25.3 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 9.091 ng/ul
RT: 25.538 min Scan# 3902
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Instrument :
BNA_M
ClientSampleId :
SSTD010026

Tgt Ion:	278	Resp:	712209
Ion Ratio	Lower	Upper	
278	100		
139	16.7	12.8	19.2
279	23.6	19.1	28.7



#96
Benzo(g,h,i)perylene
Concen: 9.245 ng/ul
RT: 26.197 min Scan# 4014
Delta R.T. -0.000 min
Lab File: BM032120.D
Acq: 21 Sep 2021 18:53

Tgt Ion:	276	Resp:	713108
Ion Ratio	Lower	Upper	
276	100		
138	22.8	16.5	24.7
277	23.2	18.5	27.7

