

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092123\
 Data File : BM041956.D
 Acq On : 21 Sep 2023 14:28
 Operator : MA/JU
 Sample : 04378-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BBKH7

Quant Time: Sep 21 15:48:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 16 03:49:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	5963	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.774	136	14715	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.606	164	7467	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	14085	0.400	ng/ul	-0.02
17) Chrysene-d12	21.527	240	6722	0.400	ng/ul	#-0.01
23) Perylene-d12	23.936	264	7112	0.400	ng/ul	#-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	34604	4.952	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.357	152	6781	0.336	ng/ul	-0.01
18) Fluoranthene-d10	19.375	212	12092	0.487	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.410	88	160	0.022	ng/ul#	26
20) Pyrene	19.747	202	3200	0.061	ng/ul#	1
26) Benzo(a)pyrene	23.784	252	4914	0.144	ng/ul#	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092123\
Data File : BM041956.D
Acq On : 21 Sep 2023 14:28
Operator : MA/JU
Sample : 04378-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BBKH7

Quant Time: Sep 21 15:48:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 16 03:49:33 2023
Response via : Initial Calibration

